


Analyte	Instrument	Result Value	Standard Unit	Z-Score
No Outlier				
Legend @ : z score $\leq$ 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score $\geq$ 3 - Unacceptable # : Not Evaluated ⊙ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: None (All are acceptable)

Corrective Action: Not Applicable.

Reviewed by: 

Dated: 14/02/24

Instrument : ERBA H 360

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Erythrocyte (RBC) Count	mill/cu.mm	2.11	2.13	0.41	0.35
@ Total Leucocytes (WBC) count	/c.mm	3570.00	3544.00	-0.48	-0.40
@ Haemoglobin	g/dL	5.90	6.00	-0.49	-0.31
@ MCV (Mean Corpuscular Volume)	fL	139.50	137.02	0.53	0.21
@ MCH (Mean Corpuscular Hb)	pg	27.90	28.08	-1.10	-1.19
@ MCHC (Mean Corpuscular Hb Concentration)	g/dL	20.00	20.14	-1.59	-0.95
@ Platelet Count	1000/ $\mu$ L	127.00	120.10	0.87	1.08
@ Hematocrit	%	29.40	28.70	0.62	0.49
@ RDW CV%	%	19.50	20.70	-1.47	-0.44

Legend @ : Acceptable

! : 2.0 < z score < 3.0 - Warning

X : z score  $\geq$  3 - Unacceptable

# : Not Evaluated

⊙ : Delayed Result Entry

\* : Not considered for evaluation.

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	0
EQAS Score Haematology - CBC	100.00 %

*Dr Puneet Kumar Nigam*

**Dr Puneet Kumar Nigam**

PT coordinator & Technical Manager, MHL EQAS

Unit No. 409-416.

Commercial Building - 1A

Kohinoor Mall, Kirol Road, Kurla (W),

Mumbai - 400070