



Metropolis Healthcare Ltd., EQAS

SHRI MAHAVIR MEDICAL CENTRE (1077)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 230412 and Sample No 01

Report Date : 13/02/2024

| Analyte | Instrument | Result Value | Standard Unit |
|------------------------|------------|--------------|---------------|
| No Outlier | | | |
| Legend @ : Acceptable | | | |
| X : Unacceptable | | | |
| # : Not Evaluated | | | |
| ⊙ Delayed Result Entry | | | |

Problem Classification: None (All are acceptable.)

Corrective Action: Not Applicable.

Reviewed by: *Jm*

Dated: 14/02/24

Instrument : Manual

| Analyte | Standard Unit | Result Value | Accepted Value |
|---|---------------|------------------|------------------|
| @ HIV Ag + Ab | S/CO | Non-reactive | Non-reactive |
| @ HBsAg | IU/mL | Non-reactive | Non-reactive |
| @ HCV | IU/mL | Non-reactive | Non-reactive |
| @ RPR (Syphilis Serology) | - | Non-reactive | Non-reactive |
| @ Dengue NS 1 | RU/ml | Negative | Negative |
| @ Widal test - Typhi O antigen (TO) | - | No Agglutination | No Agglutination |
| @ Widal test - Typhi H antigen (TH) | - | No Agglutination | No Agglutination |
| @ Widal test - ParaTyphi A-H antigen (AH) | - | No Agglutination | No Agglutination |
| @ Widal test - ParaTyphi B-H antigen (BH) | - | No Agglutination | No Agglutination |
| @ Dengue IgM | RU/ml | Negative | Negative |
| @ Dengue IgG | RU/ml | Negative | Negative |

Legend @ : Acceptable
 @ : Acceptable
 X : Unacceptable
 # : Not Evaluated
 ⊙ : Delayed Result Entry

| | |
|---------------------------------|----------|
| Total Parameters | 11 |
| Not Evaluated Parameters | 0 |
| Evaluated Parameters | 11 |
| Outlier Parameters (X) | 0 |
| EQAS Score Immunoassay Extended | 100.00 % |

Dr Puneet Kumar Nigam

Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A
 Kohinoor Mall, Kirool Road, Kurla (W),
 Mumbai - 400070