

CALIBRATION CERTIFICATE

ULR NO.: CC213324000002074F

Certificate No. : PCLI/T/24/91/001	Date of Calibration. : 15.03.2024
Certificate Issue Date. : 18.03.2024	Suggested date of next calibration. : 14.03.2025

CUSTOMER NAME ADDRESS :	DESCRIPTION OF UNIT UNDER CALIBRATION
M/S: THC Lab (A Unit of Health County Pvt Ltd) 84 2nd Floor Sheetla Mata Road Opposite AVR Invitation Garden Gurugram Haryana-122001	Name Digital Thermohygrometer Range /size 0 to 50 °C/99 %RH Least Count 0.1 °C/1 %RH Make HTC Model NO. HTC-01
CALIBRATION REQUEST DETAILS :	Serial No -----
Date of receipt :15.03.2024	ID No DTHM-01
Service Request form No.:091	Location Lab
Service Request Date :15.03.2024	Accuracy -----
Calibration perform at.:Lab	Visual Inspection Ok

Enviromental	Temperature	25±4 °C	Calibration Reference	ITS-90
Condition	Relative Humidity	50±20 % RH	Calibration Procedure	PCL/CP-06T

Details of reference standards & Major equipments used

Sr.No	Equipment Used	Make	Model/Sr.No.	Certificate No	Calibration Validity	Calibration by
1	Digital Humidity Controller with Sensor	Piyush	PCLI/DHC-01	PMM/020823 23-1	02.08.2024	PMMP
2	Precision Temperature Scanner With SPRT Sensor	Fluke	1586& 5609	CC288823000000 486F	13.05.2024	Sigma Test & Research Center

Calibration Result

Sr.No	Standard Reading In °C	UUC Reading in °C	Deviation °C	Expanded Uncertainty at 95 % C.L (Coverage factor K=2) in °C(±)
1	12.56	12.1	-0.46	0.69
2	18.78	18.1	-0.68	0.69
3	25.91	25.1	-0.81	0.69
4	36.21	35.2	-1.01	0.69
5	46.29	44.8	-1.49	0.69

Remarks : Nil

The report shall not be reproduced except in full, without written permission of the laboratory.
This result of calibration refers to only the particular items submitted for calibration.
The above results are valid at the time of and under the stated conditions measurement.

Calibrated By
Saur
(Calibration Engg/TM)



Approved By
HILAL MAONI
(QM/TM)



PRECISION CALIBRATION & LABORATORY (INDIA)

(NABL ACCREDITED LABORATORY) ISO/IEC 17025:2017
(CALIBRATION TESTING & SERVICES)



CALIBRATION CERTIFICATE

ULR NO.: CC213324000002076F

Certificate No.:	PCLI/T/24/091/003	Date of Calibration.:	15.03.2024
Certificate Issue Date.:	18.03.2024	Suggested date of next calibration.:	14.03.2025

CUSTOMER NAME ADDRESS :	DESCRIPTION OF UNIT UNDER CALIBRATION	
M/S: THC Lab (A Unit of Health County Pvt Ltd) 84 2nd Floor Sheela Mata Road Opposite AVR Invitation Garden Gurgaon Haryana-122001	Name	Digital Thermohygrometer
	Range /size	0 to 50 °C/99 %RH
	Least Count	0.1 °C/1 %RH
	Make	HTC
	Model NO.	HTC-01
CALIBRATION REQUEST DETAILS :	Serial No.	-----
Date of receipt. :15.03.2024	ID No	DTHM-03
Service Request form No. :091	Location	Lab
Service Request Date. :15.03.2024	Accuracy	-----
Calibration perform at. :Lab	Visual Inspection	Ok

Environmental Condition	Temperature	25±4 °C	Calibration Reference	ITS-90
	Relative Humidity	50±20 % RH	Calibration Procedure	PCL/CP-06T

Details of reference standards & Major equipments used

Sr.No	Equipment Used	Make	Model/Sr.No.	Certificate No.	Calibration Validity	Calibration by
1	Digital Humidity Controller with Sensor	Piyush	PCLI/DHC-01	PMM/020823 23-1	02.08.2024	PMMP
2	Precision Temperature Scanner With SPRT Sensor	Fluke	1586& 5609	CC288823000000 486F	13.05.2024	Sigma Test & Research Center

Calibration Result

Sr.No	Standard Reading In °C	UUC Reading in °C	Deviation °C	Expanded Uncertainty at 95 % C.L (Coverage factor K=2) in °C(±)
1	12.58	12.1	-0.48	0.69
2	18.83	18.1	-0.73	0.69
3	25.95	25.2	-0.75	0.69
4	36.14	35.2	-0.94	0.69
5	46.07	44.8	-1.27	0.69

Remarks : Nil

The report shall not be reproduced except in full, without written permission of the laboratory.
This result of calibration refers to only the particular items submitted for calibration.
The above results are valid at the time of and under the stated conditions measurement.

Calibrated By
[Signature]
(Calibration Engg/TM)



Approved By
[Signature]
HILAL MADNI
(QM/TM)

Page 1 of 2

PRECISION CALIBRATION & LABORATORY (INDIA)

(NABL ACCREDITED LABORATORY) ISO/IEC 17025:2017
(CALIBRATION TESTING & SERVICES)

CALIBRATION CERTIFICATE

Certificate No. : PCL/T/24/091/002	Date of Calibration. : 15.03.2024
Certificate Issue Date. : 18.03.2024	Suggested date of next calibration. : 14.03.2025


Calibration Result

Sr.No	Standard Reading Reading In %RH	UUC Reading in %RH	Deviation %RH	Expanded Uncertainty at 95 % C.L (Coverage factor K=2) in %RH(±)
1	35	36	1	3.00
2	51	53	2	3.00
3	65	67	2	3.00
4	75	77	2	3.00
5	85	88	3	3.00

Remarks : Nil

The report shall not be reproduced except in full, without written permission of the laboratory.
This result of calibration refers to only the particular items submitted for calibration.
The above results are valid at the time of and under the stated conditions measurement.


Calibrated By


(Calibration Engg/TM)



END OF THE REPORT

Approved By


HILAL MADNI
(QM/TM)

Page 2 of 2

Head Office : A-554/22, Zakir Nagar, Okhla, New Delhi-110025
Lab : Khasra No. 39, Near Mahadi Chowk, Bhagwanpur, Roorkee, Distt. Haridwar-247661 (U.K.)
Mob. : +91-9953156010, 9990120134, E-mail : precisioncalibrationhrdr.2016@gmail.com

Lab

A-554/22, Zakir Nagar, Okhla, New Delhi-110025
Near Mahadi Chowk, Bhagwanpur, Roorkee, Distt. Haridwar-247661 (U.K.)
+91-9953156010, 9990120134, E-mail : precisioncalibrationhrdr.2016@gmail.com

Page 1 of 2



PRECISION CALIBRATION & LABORATORY (INDIA)

(NABL ACCREDITED LABORATORY) ISO/IEC 17025:2017
(CALIBRATION TESTING & SERVICES)

CALIBRATION CERTIFICATE

Certificate No. : PCLI/T/24/91/001	Date of Calibration. : 15.03.2024
Certificate Issue Date. : 18.03.2024	Suggested date of next calibration. : 14.03.2025

Calibration Result

Sr.No	Standard Reading Reading In %RH	UUC Reading in %RH	Deviation %RH	Expanded Uncertainty at 95 % C.L (Coverage factor K=2) in %RH(±)
1	35	36	1	3.00
2	51	53	2	3.00
3	65	67	2	3.00
4	75	77	2	3.00
5	85	87	2	3.00

Remarks : Nil

The report shall not be reproduced except in full, without written permission of the laboratory.
This result of calibration refers to only the particular items submitted for calibration.
The above results are valid at the time of and under the stated conditions measurement.

Calibrated By
[Signature]
(Calibration Engg/TM)



Approved By
[Signature]
HILAL MADNI
(QM/TM)



PRECISION CALIBRATION & LABORATORY (INDIA)

(NABL ACCREDITED LABORATORY) ISO/IEC 17025:2017
(CALIBRATION TESTING & SERVICES)



CALIBRATION CERTIFICATE

ULR NO:- CC213324000002075F

Certificate No. : PCL/T/24/091/002	Date of Calibration. : 15.03.2024
Certificate Issue Date. : 18.03.2024	Suggested date of next calibration. : 14.03.2025

CUSTOMER NAME ADDRESS :	DESCRIPTION OF UNIT UNDER CALIBRATION	
M/S: THC Lab (A Unit of Health County Pvt Ltd) 84 2nd Floor Sheela Mata Road Opposite AVR Invitation Garden Gurugram Haryana-122001	Name	Digital Thermo hygrometer
	Range /size	0 to 50 °C/99 %RH
	Least Count	0.1 °C/1 %RH
	Make	HTC
	Model NO.	HTC-01
CALIBRATION REQUEST DETAILS :	Serial No	-----
Date of receipt. :15.03.2024	ID No	DTHM-02
Service Request form No. :091	Location	Lab
Service Request Date. :15.03.2024	Accuracy	-----
Calibration perform at. :Lab	Visual Inspection	Ok

Environmental Condition	Temperature	25±4 °C	Calibration Reference	ITS-90
	Relative Humidity	50±20 % RH	Calibration Procedure	PCL/CP-06T

Details of reference standards & Major equipments used

Sr.No	Equipment Used	Make	Model/Sr.No.	Certificate No	Calibration Validity	Calibration by
1	Digital Humidity Controller with Sensor	Piyush	PCLI/DHC-01	PMM/020823 23-1	02.08.2024	PMMP
2	Precision Temperature Scanner With SPRT Sensor	Fluke	1586& 5609	CC288823000000 486F	13.05.2024	Sigma Test & Research Center

Calibration Result

Sr.No	Standard Reading In °C	UUC Reading in °C	Deviation °C	Expanded Uncertainty at 95 % C.L (Coverage factor K=2) in °C(±)
1	12.63	12.1	-0.53	0.69
2	18.78	18.1	-0.68	0.69
3	25.95	25.1	-0.85	0.69
4	36.18	35.2	-0.98	0.69
5	46.29	44.9	-1.39	0.69

Remarks : Nil

The report shall not be reproduced except in full, without written permission of the laboratory.
This result of calibration refers to only the particular items submitted for calibration.
The above results are valid at the time of and under the stated conditions measurement.

Calibrated By

(Calibration Engg/TM)



Approved By

HILAL MADNI
(QM/TM)



PRECISION CALIBRATION & LABORATORY (INDIA)

(NABL ACCREDITED LABORATORY) ISO/IEC 17025:2017
(CALIBRATION TESTING & SERVICES)

CALIBRATION CERTIFICATE


Certificate No. : PCLI/T/24/091/003	Date of Calibration. : 15.03.2024
Certificate Issue Date. : 18.03.2024	Suggested date of next calibration. : 14.03.2025

Calibration Result

Sr.No	Standard Reading Reading In %RH	UUC Reading in %RH	Deviation %RH	Expanded Uncertainty at 95 % C.L (Coverage factor K=2) in %RH(±)
1	35	36	1	3.00
2	51	52	1	3.00
3	65	67	2	3.00
4	75	77	2	3.00
5	85	88	3	3.00

Remarks : Nil

The report shall not be reproduced except in full, without written permission of the laboratory.
This result of calibration refers to only the particular items submitted for calibration.
The above results are valid at the time of and under the stated conditions measurement.

Calibrated By

(Calibration Engg/TM)



Approved By
HILAL MADNI
(QM/TM)

END OF THE REPORT

Page 2 of 2

Head Office : A-554/22, Zakir Nagar, Okhla, New Delhi-110025
Lab : Khasra No. 39, Near Mahadr Chowk, Bhagwanpur, Roorkee, Distt. Haridwar-247661 (U.K.)
Mob. : +91-9953156010, 9990120134, E-mail : precisioncalibrationhrdr.2016@gmail.com

Page No 1 of 1

CALIBRATION CERTIFICATE

U/LR NO. - CC213324000002077F

Certificate No. : PCLI/T/24/091/004	Date of Calibration : 15.03.2024
Certificate issue Date : 18.03.2024	Suggested date of next calibration : 14.03.2025

CUSTOMER NAME ADDRESS :	DESCRIPTION OF UNIT UNDER CALIBRATION
M/S: THC Lab (A Unit of Health County Pvt Ltd) 84 2nd Floor Sheela Mata Road Opposite AVR Invitation Garden Gurgaon Haryana-122001	Name Refrigerator(Digital Thermometer with sensor) Range /size 2 to 8 °C Least Count 0.1 °C Make Imported
CALIBRATION REQUEST DETAILS : Date of receipt : 15.03.2024 Service Request Form No. : 091 Service Request Date : 15.03.2024 Calibration perform at : Lab Calibration perform at : Site	Model NO. _____ Serial No. _____ ID No. RF-01 Location Lab Accuracy _____ Visual inspection OK

Environmental Condition	Temperature	25±4 °C	Calibration Reference	ITS-90
	Relative Humidity	50±20 % RH	Calibration Procedure	PCL/CP-05T

Details of reference standards & Major equipments used

Sr.No	Equipment Used	Make	Model/Sr.No	Certificate No	Calibration Validity	Calibration by
1	Precision Temperature Scanner With SPRT Sensor	Fluke	1586& 5609	CC288823000000 486F	13.05.2024	Sigma Test & Research Center

Calibration Result

Sr.No	Std Reading in °C	UUC Reading in °C	Deviation °C	Expanded Uncertainty at 95 % C.L (Coverage factor K=2) in °C(±)
1	2.08	2.0	-0.08	0.81
2	4.14	4.0	-0.14	0.81
3	6.17	6.0	-0.17	0.81
4	8.24	8.0	-0.24	0.81

Remarks : Nil

The report shall not be reproduced except in full, without written permission of the laboratory.
This result of calibration refers to only the particular items submitted for calibration.
The above results are valid at the time of and under the stated conditions measurement.

Calibrated By
[Signature]
(Calibration Engg/TM)



Approved By
[Signature]
HIJAL MAHANI
(QM/TM)

END OF THE REPORT



PRECISION CALIBRATION & LABORATORY (INDIA)

(NABL ACCREDITED LABORATORY) ISO/IEC 17025:2017
(CALIBRATION TESTING & SERVICES)



CALIBRATION CERTIFICATE

LILR NO.: CC213324000002078F

Certificate No. : PCL/T/24/091/005	Date of Calibration. : 15.03.2024
Certificate Issue Date. : 18.03.2024	Suggested date of next calibration. : 14.03.2025

CUSTOMER NAME ADDRESS :	DESCRIPTION OF UNIT UNDER CALIBRATION	
M/S: THC Lab (A Unit of Health County Pvt Ltd) 84 2nd Floor Sheeta Mata Road Opposite AVR Invitation Garden Gurugram Haryana-122001	Name	Refrigerator(Digital Thermometer with sensor)
CALIBRATION REQUEST DETAILS : Date of receipt :15.03.2024 Service Request form No.:091 Service Request Date. :15.03.2024 Calibration perform at :Lab Calibration perform at :Site	Range /size	2 to 8 °C
	Least Count	0.1 °C
	Make	Imported.
	Model NO.	-----
	Serial No	-----
	ID No	RF-02
	Location	Lab
Accuracy	-----	
	Visual Inspection	Ok

Enviromental Condition	Temperature	25±4 °C	Calibration Reference	ITS-90
	Relative Humidity	50±20 % RH	Calibration Procedure	PCL/CP-05T

Details of reference standards & Major equipments used

Sr.No	Equipment Used	Make	Model/Sr.No.	Certificate No	Calibration Validity	Calibration by
1	Precision Temperature Scanner With SPRT Sensor	Fluke	1586& 5609	CC288823000000 486F	13.05.2024	Sigma Test & Research Center

Calibration Result

Sr.No	Std Reading In °C	UUC Reading in °C	Deviation °C	Expanded Uncertainty at 95 % C.L (Coverage factor K=2) in °C(±)
1	2.05	2.0	-0.05	0.81
2	4.09	4.0	-0.09	0.81
3	6.13	6.0	-0.13	0.81
4	8.17	8.0	-0.17	0.81

Remarks : Nil

The report shall not be reproduced except in full, without written permission of the laboratory.
This result of calibration refers to only the particular items submitted for calibration.
The above results are valid at the time of and under the stated conditions measurement.

Calibrated By
[Signature]
(Calibration Engg/TM)



END OF THE REPORT

Approved By
[Signature]
HIMAL MADNI
(QM/TM)



PRECISION CALIBRATION & LABORATORY (INDIA)

(NABL ACCREDITED LABORATORY) ISO/IEC 17025:2017
(CALIBRATION TESTING & SERVICES)



CALIBRATION CERTIFICATE

ULR NO.: CC213324000002079F

Certificate No. : PCL/T/24/091/006	Date of Calibration : 15.03.2024
Certificate Issue Date : 18.03.2024	Suggested date of next calibration : 14.03.2025

CUSTOMER NAME ADDRESS :	DESCRIPTION OF UNIT UNDER CALIBRATION	
M/S: THC Lab (A Unit of Health County Pvt Ltd) 84 2nd Floor Sheela Mata Road Opposite AVR Invitation Garden Gurugram Haryana-122001	Name	Refrigerator(Digital Thermometer with sensor)
CALIBRATION REQUEST DETAILS : Date of receipt :15.03.2024 Service Request form No :091 Service Request Date :15.03.2024 Calibration perform at :Lab Calibration perform at :Site	Range /size	2 to 8 °C
	Least Count	0.1 °C
	Make	Imported
	Model NO.	-----
	Serial No	-----
	ID No	RF-03
	Location	Lab
Accuracy	-----	
	Visual inspection	Ok

Enviromental Condition	Temperature	25±4 °C	Calibration Reference :	ITS-90
	Relative Humidity	50±20 % RH	Calibration Procedure	PCL/CP-05T

Details of reference standards & Major equipments used

Sr.No	Equipment Used	Make	Model/Sr.No.	Certificate No	Calibration Validity	Calibration by
1	Precision Temperature Scanner With SPRT Sensor	Fluke	1586& 5609	CC288823000000 486F	13.05.2024	Sigma Test & Research Center

Calibration Result

Sr.No	Std Reading in °C	UUC Reading in °C	Deviation °C	Expanded Uncertainty at 95 % C.L (Coverage factor K=2) in °C(±)
1	2.07	2.0	-0.07	0.81
2	4.11	4.0	-0.11	0.81
3	6.15	6.0	-0.15	0.81
4	8.19	8.0	-0.19	0.81

Remarks : Nil

The report shall not be reproduced except in full, without written permission of the laboratory.
This result of calibration refers to only the particular items submitted for calibration.
The above results are valid at the time of and under the stated conditions measurement.

Calibrated By

(Calibration Engg/TM)



Approved By

(QM/TM)



PRECISION CALIBRATION & LABORATORY (INDIA)

(NABL ACCREDITED LABORATORY) ISO/IEC 17025:2017
(CALIBRATION TESTING & SERVICES)



CALIBRATION CERTIFICATE

ULR NO.: CC213324000002080F

Certificate No. : PCL/T/24/091/007	Date of Calibration : 15.03.2024
Certificate Issue Date : 18.03.2024	Suggested date of next calibration : 14.03.2025

CUSTOMER NAME ADDRESS :	DESCRIPTION OF UNIT UNDER CALIBRATION	
M/S: THC Lab (A Unit of Health County Pvt Ltd) 84 2nd Floor Sheetla Mata Road Opposite AVR Invitation Garden Gurugram Haryana-122001	Name	Deep Freezer(Digital Thermometer with sensor)
CALIBRATION REQUEST DETAILS : Date of receipt : 15.03.2024 Service Request form No.:091 Service Request Date : 15.03.2024 Calibration perform at : Lab Calibration perform at : Site	Range /size	Upto -20 °C
	Least Count	0.1 °C
	Make	Haier
	Model NO.	-----
	Serial No	-----
	ID No	DF-01
	Location	Lab
	Accuracy	-----
	Visual Inspection	Ok

Enviromental Condition	Temperature	25±4 °C	Calibration Reference	ITS-90
	Relative Humidity	50±20 % RH	Calibration Procedure	PCL/CP-05T

Details of reference standards & Major equipments used

Sr.No	Equipment Used	Make	Model/Sr.No.	Certificate No	Calibration Validity	Calibration by
1	Precision Temperature Scanner With SPRT Sensor	Fluke	1586& 5609	CC288823000000 486F	13.05.2024	Sigma Test & Research Center

Calibration Result

Sr.No	Std Reading In °C	UUC Reading in °C	Deviation °C	Expanded Uncertainty at 95 % C.L (Coverage factor K=2) in °C(±)
1	-4.09	-4.0	0.09	0.81
2	-8.14	-8.0	0.14	0.81
3	-12.18	-12.0	0.18	0.81
4	-16.62	-16.0	0.62	0.81
5	-20.65	-20.0	0.65	0.81

Remarks : Nil

The report shall not be reproduced except in full, without written permission of the laboratory.
This result of calibration refers to only the particular items submitted for calibration.
The above results are valid at the time of and under the stated conditions measurement.

Calibrated By
[Signature]
(Calibration Engg/TM)



Approved By
[Signature]
(QM/TM)

END OF THE REPORT

Page No 1 of 1

Head Office : A-554/22, Zakir Nagar, Okhla, New Delhi-110025
Lab : Khasra No. 39, Near Mahadi Chowk, Bhagwanpur, Roorkee, Distt. Haridwar-247661 (U.K.)
Mob. : +91-9953156010, 9990120134, E-mail : precisioncalibrationhdr2016@gmail.com



PRECISION CALIBRATION & LABORATORY (INDIA)

(NABL ACCREDITED LABORATORY) ISO/IEC 17025:2017
(CALIBRATION TESTING & SERVICES)



CALIBRATION CERTIFICATE

ULR NO.: CC213324000002081F

Certificate No. : PCLI/T/24/091/008	Date of Calibration. : 15.03.2024
Certificate Issue Date. : 18.03.2024	Suggested date of next calibration : 14.03.2025

CUSTOMER NAME ADDRESS :	DESCRIPTION OF UNIT UNDER CALIBRATION	
M/S: THC Lab (A Unit of Health County Pvt Ltd) 84 2nd Floor Sheetla Mata Road Opposite AVR Invitation Garden Gurugram Haryana-122001	Name	Deep Freezer(Digital Thermometer with sensor)
CALIBRATION REQUEST DETAILS : Date of receipt.:15.03.2024 Service Request form No.:091 Service Request Date.:15.03.2024 Calibration perform at.:Lab Calibration perform at.:Site	Range /size	Upto -20 °C
	Least Count.	0.1 °C
	Make	Haier
	Model NO.	-----
	Serial No	-----
	ID No	DF-02
	Location	Lab
Accuracy	-----	
Visual Inspection	Ok	

Environmental	Temperature	25±4 °C	Calibration Reference	ITS-90
Condition	Relative Humidity	50±20 % RH	Calibration Procedure	PCL/CP-05T

Details of reference standards & Major equipments used

Sr.No	Equipment Used	Make	Model/Sr.No.	Certificate No	Calibration Validity	Calibration by
1	Precision Temperature Scanner With SPRT Sensor	Fluke	1586&5609	CC288823000000486F	13.05.2024	Sigma Test & Research Center

Calibration Result

Sr.No	Std Reading In °C	UUC Reading in °C	Deviation °C	Expanded Uncertainty at 95 % C.L (Coverage factor K=2) in °C(±)
1	-4.12	-4.0	0.12	0.81
2	-8.15	-8.0	0.15	0.81
3	-12.19	-12.0	0.19	0.81
4	-16.46	-16.0	0.46	0.81
5	-20.63	-20.0	0.63	0.81

Remarks : Nil

The report shall not be reproduced except in full, without written permission of the laboratory.
This result of calibration refers to only the particular items submitted for calibration.
The above results are valid at the time of and under the stated conditions measurement.

Calibrated By
[Signature]
(Calibration Engg/TM)



Approved By
[Signature]
(QM/TM)

END OF THE REPORT



PRECISION CALIBRATION & LABORATORY (INDIA)

(NABL ACCREDITED LABORATORY) ISO/IEC 17025:2017
(CALIBRATION TESTING & SERVICES)

CALIBRATION CERTIFICATE

Certificate No. : PCL/E/24/091/010	Date of Calibration : 15.03.2024
Certificate Issue Date : 18.03.2024	Suggested date of next calibration : 14.03.2025

CUSTOMER NAME ADDRESS :	DESCRIPTION OF UNIT UNDER CALIBRATION
M/S. THE Lab (A Unit of Health County Pvt Ltd) B4 2nd Floor Sheela Mata Road Opposite AVR Invitation Garden Gurgaon Haryana-122001	Name : Digital TDS Meter Range /size : 0 to 999 PPM Least Count : 1PPM Make : Imported Model NO. : _____ Serial No : _____ ID No : TDS-01 Location : Lab Accuracy : _____ Visual Inspection : ok
CALIBRATION REQUEST DETAILS : Date of receipt : 15.03.2024 Service Request form No. :091 Service Request Date : 15.03.2024 Calibration perform at : Lab	

Environmental Condition	Temperature : 25±4 °C	Calibration Reference	Comp. method
	Relative Humidity : 50±20 % RH	Calibration Procedure	PCL/CP-27 E

Details of reference standards & Major equipments used

Sr.No	Equipment Used	Make	Model/Sr.No.	Certificate No	Calibration Validity	Calibration by
1	Universal Calibrator	Masibus	3001M/16051	AIPL/100823/1-4	11.08.2024	AIPL
2	TDS Meter	HM	PCLI/TDS-01	AT23000000199	21.08.2024	ASIAN
3	Std Solution					

Calibration Result

Sr.No	Std Reading in (PPM)	UUC Reading in (PPM)	Deviation (PPM)	Expanded Uncertainty at 95.54% C.I. (Coverage factor K=2) in %±
1	100.00	100	0	2.1
2	200.00	200	0	2.1
3	300.00	301	1	2.1
4	500.00	501	1	2.1
5	800.00	802	2	2.1


Remarks - Nil

The report shall not be reproduced except in full, without written permission of the laboratory.
This result of calibration refers to only the particular items submitted for calibration.
The above results are valid at the time of and under the stated conditions measurement.

Calibrated By
[Signature]
(Calibration Engg/TM)



Approved By
[Signature]
HILAL MAJANI
(QM/TM)



CALIBRATION CERTIFICATE

Lab No.: PC/L/24/001/008

Certificate No.:	PC/L/24/001/008	Date of Calibration:	25.03.2024
Certificate Issue Date:	28.03.2024	Suggested date of next Calibration:	14.03.2025

CUSTOMER NAME / ADDRESS	DESCRIPTION OF UNIT / UNDER CALIBRATION
M/s. TNC Lab 1A, Govt. of Health, County Plot (30) SR 2nd Floor, Dharama Mata Road, Opposite Kishor Mahal, Garden Corrigram, Varanasi-221003	Name: Digital Timer Range /Use: 0 to 1 Hr Least Count: 0.1 Hr Make: Imported Model No: --- Serial No: --- ID No: DT-03 Location: Lab Accuracy: --- Visual Inspection: OK

Environmental Condition	Temperature	25±4 °C	Calibration Reference	IS 1248-2003
	Relative Humidity	50±20 % RH	Calibration Procedure	PCL/CP-001

Details of reference standards & Major equipments used

Sr.No	Equipment Used	Make	Model/Sr.No.	Certificate No.	Calibration Validity	Calibration by
1	Digital Timer Calibrator	HC	2407	CBM/CAL/24-01 104	25.03.2025	CBJ/ajayr

Calibration Result

Sr.No	Std Reading In Sec	Std Reading In Hr	Deviation Sec	Expanded Uncertainty at 95 % C.I. (Coverage factor K=2) in % (±) sec
1	725.4	0.2	-1.4	0.15
2	1843	0.4	-3	0.15
3	2345	0.6	-5	0.15
4	2886	0.8	-6	0.15
5	3407	1.0	-7	0.15

Remarks: Nil

The report shall not be reproduced except in full, without written permission of the laboratory.
The result of calibration refers to only the particular items submitted for calibration.
The above results are valid at the time of and under the stated conditions measurement.

Calibrated By
(Signature)
(Calibration Engg/TM)



Approved By
(Signature)
H.M. SAHNI
(DM/TM)

END OF THE REPORT