



FACTUAL CALIBRATION SERVICES (OPC) PVT. LTD.

A House of Complete Solution



Contact Details:- +91-9205523341, +91-9868784023
Office No. 130, 1st Floor Cloud-9 Sec 1 Valsahll Ghaziabad Uttar Pradesh 201010

Calibration Certificate

Format No. : FCS/QF46

ISSUE DATE 01/07/23

Page No. : 01 of 01

SERVICE REQUEST & CALIBRATION REPORT DETAILS (As per agreed with the Customer)

Calibration Request Dated	06.11.2023	Date of Calibration :	07.11.2023
Calibration Request / Job No.	125 / 125-006	Next Date of Calibration	07.11.2024
Certificate No.	FCS/23/125-006	Certificate Issue Date	08.11.2023

CUSTOMER ADDRESS		ENVIRONMENTAL & INSTRUMENT CONDITIONS	
M/s Dipakshi Hospital Udhayan Marg, NTPC Township, C Block, Sector 33, Noida, Uttar Pradesh 201307 INDIA		Temperature	(25 ± 15) °C
		Relative Humidity	(50 ± 20) %RH
		Zerro Error	Nil
		Discipline	Medical
		Instrument Condition	In Working (Good)

CALIBRATION CERTIFICATE OF		DESCRIPTION OF (EQP) UNIT UNDER CALIBRATION	
PIPETTE		Make / Model	YD URISCAN OPTIMA
		Range / Size	As Per Instrument
		Least Count	As Per Range
		Serial No.	UID-00614814
Procedure Followed	FCS/MD/117	Asset Code / LD No.	NA
Reference Standard	NABL 126	Location	LAB
Calibration Performed At	At Site	Accuracy/Acceptance	Not Specified

DETAILS OF REFERENCE STANDARDS AND MAJOR EQUIPMENTS USED FOR CALIBRATION

Sr. No	Instrument Details	Asst Code	Certificate No.	Calibrated By	Valid upto
1	Volumetric Glasswarepe	FCS/MD/VG-01	AT23000007411	ASIAN	02.07.2024

The Standards used for calibration are traceable to National Standards.

CALIBRATION RESULT

1. Electrical Safety		
Visual Test		
Sr. No	TESTS	REMARKS
1	Power Chords, Cable Check	OK
2	Main Socket Check	OK
3	Equipment Type (B/BF/CF)	B
4	Equipment Class (I/II/III)	I

B-BODY TYPE, BF-BODY FLOAT TYPE, CF-CARIDIAC FLOAT TYPE, I-PROPERLY EARTHED, II-DOUBLE INSULATED
III-WITH EXTRA SAFETY LOW VOLTAGE.

2. Electrical Safety Test				
Sr. No	PARAMETERS	Measured Value		REMARKS
1	Voltage Between Line & Neutral (V in)	236.3	V	OK
2	Voltage Between Line & Earth (V ie)	243.6	V	OK
3	Voltage between Neutral & Earth (V en)	2.2	V	OK
4	Load Current (I a)	0.5	A	OK
5	Earth Leakage Current (I L)	7.8	µ A	OK
6	Enclosure Current	0.6	µ A	OK
7	Patient Leakage Current (P L)	0.7	µ A	OK

CALIBRATED BY (Calibration Engineer)	APPROVED BY (Authorized Signatory)
--	--

Note :-	<ol style="list-style-type: none"> UUC or Unit Under Calibration Uncertainty of measurement at approximately 75% confidence level with coverage factor k=2, ±2σ per The Calibration Result reported are valid at the date of Calibration and under the stated conditions of Measurement. The Calibration report is not valid without sign of Calibration Services (OPC) PVT. LTD. This Calibration Certificate only to the particular equipment (Instrument) submitted for Calibration. The Calibration Certificate should not be regenerated without the permission of the Authority of Factual Calibration Services (OPC) PVT. LTD. Calibration Certificate issue for weight & Measure parameters in Metrology Volumetric equipment for scientific Purpose only.
----------------	---

A CALIBRATION LABORATORY

www.factualcalibrationpvtltd.com

factualcalibration23@gmail.com

OUR BRANCHES : GHAZIABAD KATHMANDU LUCKNOW REWARI NOIDA PUNJAB

