



HORIBA India Private Limited

246, Okhla Industrial Estate Phase-III,
New Delhi 110020, India
Tel : +91 (11) 4646 5000 / 4669 5001
Fax : +91 (11) 4646 5020
http : //www.horiba.com
CIN : U73100DL2006PTC153232

HIN/MED/2022-2023/1005101
05 aug 2023

CALIBRATION CERTIFICATE

This is to certify that the Hematology Analyzer **YUMIZEN H500** bearing serial number: **007YOXH03585** installed at **tapdiya diagnostic Aurangabad** as calibrated on 05th aug 2023

Calibrator : ABXMINOCAL
LotNo. : CX482
ExpiryDate : 07th aug 23

The reports of Blank Cycle, Repeatability and Calibration Values were all found in acceptable range.

Next calibration cycle is due on 04th aug 24.



ShrishDixit

(Head- Products &
Marketing) For **Horiba India Pvt
.Ltd.**

HORIBA

Medical

HORIBAIndiaPrivateLimited

246,OkhlaIndustrialEstatePhase-III,New

Delhi110020,India

Tel:+91(11)46465000/46695001

Fax:+91(11)46465020

<http://www.horiba.com>

CIN:U73100DL2006FTC153232