

**CALIBRATION CERTIFICATE**

In accordance with ISO / IEC-17025 : 2017

F10-CC-03

Page 1 of 1

<b>Certificate No.:</b> SCT/100424/16/2	<b>ULR:</b> CC280624200001860F
<b>Customer Name &amp; Address :</b> M/s. TAPADIA DIAGNOSTIC CENTRE., At BANSILAL NAGAR, STATION ROAD,AURANGABAD, Chhatrapati Sambhaji Nagar -431005, Maharashtra	<b>Issue Date</b> : 11-04-2024 <b>Reference Date</b> : 10-04-2024 <b>Calibration Date</b> : 10-04-2024 <b>Calibration Due Date</b> : 09-04-2025

**Details of Unit Under Calibration :**

Description	: Micro Pipette
Make	: TURBODYNE
Model	: SC TM
Range	: 2-20 µl
SI.No.	: 396862
Resolution	: NA
ID.No.	: NA
Calibrated At	: At Lab
Condition on Receipt	: OK

**Environmental Conditions :** Temperature : (23±0.5)°C Relative Humidity : (50±10) % RH Air Pressure: (900-1100)hpa**Calibration SOP/Ref Standard :** SOP-MVL-03, IS/ISO 8655-2&6:2022, ISO/TR 20461:2000**Mechanical-Volume****Master Instruments Detail**

Name of the Master used	Id.No.	Certificate No.	Valid Upto
Micro Balance	SL/PMM/MB/01	TC/23-24/4054 03	23/05/2024

**Calibration Results**

S.No.	UOM	Nominal Volume in UOM	Average Standard Reading in UOM	Systematic Error in UOM	Random Error in UOM	Systematic Error MPE in UOM	Random Error MPE in UOM	Expanded Uncertainty in ± UOM
1	µl	2	1.988	0.012	0.024	0.2	0.1	0.080
2	µl	10	9.958	0.042	0.025	0.2	0.1	0.080
3	µl	20	19.912	0.088	0.025	0.2	0.1	0.080

**Remarks :**

- This certificate pertains only to the item calibrated.
- The calibration results reported in this certificate are valid at the time of and at the stated environmental conditions.
- The calibration interval is determined based on customer's requirements.
- The calibration is traceable to National standards as per traceability details given in the certificate.
- This calibration certificate shall not be reproduced in full, except with prior written approval of Managing Director, SIMCO Calibration & Testing Pvt. Ltd.
- This calibration certificate is meant for scientific and industrial purpose only.
- The NABL Symbol is used as per NABL guidelines in NABL-133.
- The Expanded Uncertainty is reported at 95% confidence level with approximate coverage factor k= 2.
- UOM = Unit of Measurement
- To use this instrument at other temperatures use the formula given below.  
 $V_{27} = V_T (1 - C (t - 27))$   
 where,  $V_T$  = Volume measured at temperature  $t^{\circ}\text{C}$  (ml),  
 $V_{27}$  = Volume measured at  $27^{\circ}\text{C}$  (ml)  
 $C$  = coefficient of cubical expansion of Pipette tips (0.00024 / $^{\circ}\text{C}$ )

Janapati Naga Akhila  
Technical Associate Engineer  
Calibrated By

\*\* End of Certificate \*\*

Digitally signed by  
N.V.Kameswararao  
Date: 2024.04.16 11:25:19 +05:30  
Reason: Calibration CertificateAuthorized Signatory  
Chief Executive Officer



## CALIBRATION CERTIFICATE

In accordance with ISO / IEC-17025 : 2017

F10-CC-03

Page 1 of 1

<b>Certificate No.:</b> SCT/100424/16/3	<b>ULR:</b> CC280624200001861F
<b>Customer Name &amp; Address :</b> M/s. TAPADIA DIAGNOSTIC CENTRE., At BANSILAL NAGAR, STATION ROAD,AURANGABAD, Chhatrapati Sambhaji Nagar -431005, Maharashtra	<b>Issue Date</b> : 11-04-2024 <b>Reference Date</b> : 10-04-2024 <b>Calibration Date</b> : 10-04-2024 <b>Calibration Due Date</b> : 09-04-2025

**Details of Unit Under Calibration :**

Description	: Micro Pipette		
Make	: SUPERHIT	Model	: XL+
Range	: 100-1000 µl	SI.No.	: RB632196
Resolution	: NA	ID.No.	: NA
Calibrated At	: At Lab	Condition on Receipt	: OK

**Environmental Conditions :** Temperature : (23±0.5)°C Relative Humidity : (50±10) % RH Air Pressure: (900-1100)hpa

**Calibration SOP/Ref Standard :** SOP-MVL-03, IS/ISO 8655-2&6:2022, ISO/TR 20461:2000

**Mechanical-Volume**

**Master Instruments Detail**

Name of the Master used	Id.No.	Certificate No.	Valid Upto
Semi Micro Balance	SL/PMM/SMB/02	TC/23-24/4054 02	23/05/2024

**Calibration Results**

S.No.	UOM	Nominal Volume in UOM	Average Standard Reading in UOM	Systematic Error in UOM	Random Error in UOM	Systematic Error MPE in UOM	Random Error MPE in UOM	Expanded Uncertainty in ± UOM
1	µl	100	100.25	-0.25	0.02	8	3	0.060
2	µl	500	500.39	-0.39	0.02	8	3	0.200
3	µl	1000	1000.26	-0.26	0.02	8	3	0.200

**Remarks :**

- This certificate pertains only to the item calibrated.
- The calibration results reported in this certificate are valid at the time of and at the stated environmental conditions.
- The calibration interval is determined based on customer's requirements.
- The calibration is traceable to National standards as per traceability details given in the certificate.
- This calibration certificate shall not be reproduced in full, except with prior written approval of Managing Director, SIMCO Calibration & Testing Pvt. Ltd.
- This calibration certificate is meant for scientific and industrial purpose only.
- The NABL Symbol is used as per NABL guidelines in NABL-133.
- The Expanded Uncertainty is reported at 95% confidence level with approximate coverage factor k= 2.
- UOM = Unit of Measurement
- To use this instrument at other temperatures use the formula given below.  
 $V_{27} = V_T (1 - C (t - 27))$   
 where,  $V_T$  = Volume measured at temperature  $t^\circ\text{C}$  (ml),  
 $V_{27}$  = Volume measured at  $27^\circ\text{C}$  (ml)  
 $C$  = coefficient of cubical expansion of Pipette tips (0.00024 /°C)



Janapati Naga Akhila  
Technical Associate Engineer  
Calibrated By

\*\* End of Certificate \*\*

Digitally signed by  
N.V.Kameswararao  
Date: 2024.04.16 11:25:27 +05:30  
Reason: Calibration Certificate

Authorized Signatory  
Chief Executive Officer



## CALIBRATION CERTIFICATE

In accordance with ISO / IEC-17025 : 2017

F10-CC-03

Page 1 of 1

<b>Certificate No.:</b> SCT/100424/16/1	<b>ULR :</b> CC280624200001859F
<b>Customer Name &amp; Address :</b> M/s. TAPADIA DIAGNOSTIC CENTRE., At BANSILAL NAGAR, STATION ROAD,AURANGABAD, Chhatrapati Sambhaji Nagar -431005, Maharashtra	<b>Issue Date</b> : 11-04-2024 <b>Reference Date</b> : 10-04-2024 <b>Calibration Date</b> : 10-04-2024 <b>Calibration Due Date</b> : 09-04-2025

### Details of Unit Under Calibration :

Description : Micro Pipette	
Make : SUPERHIT	Model : XL+
Range : 5 - 50 µl	SI.No. : QJ616915
Resolution : NA	ID.No. : NA
Calibrated At : At Lab	Condition on Receipt : OK

**Environmental Conditions :** Temperature : (23±0.5)°C Relative Humidity : (50±10) % RH Air Pressure: (900-1100)hpa

**Calibration SOP/Ref Standard :** SOP-MVL-03, IS/ISO 8655-2&6:2022, ISO/TR 20461:2000

**Mechanical-Volume**

### Master Instruments Detail

Name of the Master used	Id.No.	Certificate No.	Valid Upto
Micro Balance	SL/PMM/MB/01	TC/23-24/4054 03	23/05/2024

### Calibration Results

S.No.	UOM	Nominal Volume in UOM	Average Standard Reading in UOM	Systematic Error in UOM	Random Error in UOM	Systematic Error MPE in UOM	Random Error MPE in UOM	Expanded Uncertainty in ± UOM
1	µl	5	498.530	-493.530	0.002	0.5	0.25	0.080
2	µl	25	482.492	-457.492	0.002	0.5	0.25	0.080
3	µl	50	491.933	-441.933	0.002	0.5	0.25	0.080

### Remarks :

- This certificate pertains only to the item calibrated.
- The calibration results reported in this certificate are valid at the time of and at the stated environmental conditions.
- The calibration interval is determined based on customer's requirements.
- The calibration is traceable to National standards as per traceability details given in the certificate.
- This calibration certificate shall not be reproduced in full, except with prior written approval of Managing Director, SIMCO Calibration & Testing Pvt. Ltd.
- This calibration certificate is meant for scientific and industrial purpose only.
- The NABL Symbol is used as per NABL guidelines in NABL-133.
- The Expanded Uncertainty is reported at 95% confidence level with approximate coverage factor k= 2.
- UOM = Unit of Measurement
- To use this instrument at other temperatures use the formula given below.  
 $V_{27} = V_T (1 - C (t - 27))$   
 where,  $V_T$  = Volume measured at temperature  $t^\circ\text{C}$  (ml),  
 $V_{27}$  = Volume measured at  $27^\circ\text{C}$  (ml)  
 $C$  = coefficient of cubical expansion of Pipette tips (0.00024 /°C)



Janapati Naga Akhila  
Technical Associate Engineer  
Calibrated By

\*\* End of Certificate \*\*

Digitally signed by  
N.V.Kameswararao  
Date: 2024.04.16 11:25:10 +05:30  
Reason: Calibration Certificate

Authorized Signatory  
Chief Executive Officer



## CALIBRATION CERTIFICATE

In accordance with ISO / IEC-17025 : 2017

F10-CC-03

Page 1 of 1

<b>Certificate No.:</b> SCT/100424/17/1	<b>ULR:</b> CC280624100003570F
<b>Customer Name &amp; Address :</b> M/s. TAPADIA DIAGNOSTIC CENTRE., At BANSILAL NAGAR, STATION ROAD,AURANGABAD, Chhatrapati Sambhaji Nagar -431005, Maharashtra	<b>Issue Date</b> : 13-04-2024 <b>Reference Date</b> : 10-04-2024 <b>Calibration Date</b> : 10-04-2024 <b>Calibration Due Date</b> : 09-04-2025

**Details of Unit Under Calibration :**

Description : Centrifuge	
Make : ELEKTROCRAFT INDIA PVT LTD	Model : AAM 9056
Range : Upto 4500 RPM	SI.No. : EAC 40
Resolution : 10 RPM	ID.No. : TDC/AB/INST/01
Calibrated At : At Site	Condition on Receipt : OK

**Environmental Conditions :** Temperature : (24±4)°C Relative Humidity : (50±20) % RH

**Calibration SOP/Ref Standard :** SOP-PL-08, SANAS TR 45-01

**Mechanical Calibration-Speed**

**Master Instruments Detail**

Name of the Master used	Id.No.	Certificate No.	Valid Upto
Digital Tachometer	SL/PMP/DTM/01	SL2308PL0286-001	18/08/2024

**NON CONTACT TYPE**

S.No.	Average Standard Reading (RPM)	Average UUC Reading (RPM)	Error (RPM)	Expanded Uncertainty in RPM (k =2)
1	499.8	500	0.2	0.70
2	999.6	1000	0.4	0.70
3	1999.5	2000	0.5	1.64
4	3499.3	3500	0.7	1.64
5	4499.1	4500	0.9	1.64

**Remarks :**

- This certificate pertains only to the item calibrated.
- The calibration results reported in this certificate are valid at the time of and at the stated environmental conditions.
- The calibration interval is determined based on customer's requirements.
- The calibration is traceable to National standards as per traceability details given in the certificate.
- This calibration certificate shall not be reproduced in full, except with prior written approval of Managing Director, SIMCO Calibration & Testing Pvt. Ltd.
- This calibration certificate is meant for scientific and industrial purpose only.
- The NABL Symbol is used as per NABL guidelines in NABL-133.
- The Expanded Uncertainty is reported at 95% confidence level with approximate coverage factor k= 2.
- UOM = Unit of Measurement



Battu Nagarjuna  
Technical Executive Engineer  
Calibrated By

\*\* End of Certificate \*\*

Digitally signed by  
N.V.Kameswararao  
Date: 2024.04.16 12:41:57 +05:30  
Reason: Calibration Certificate

Authorized Signatory  
Chief Executive Officer

**CALIBRATION CERTIFICATE**

In accordance with ISO / IEC-17025 : 2017

F10-CC-03

Page 1 of 1

<b>Certificate No.:</b> SCT/100424/17/2	<b>ULR:</b> CC280624000004708F
<b>Customer Name &amp; Address:</b> M/s. TAPADIA DIAGNOSTIC CENTRE., At BANSILAL NAGAR, STATION ROAD, AURANGABAD, Chhatrapati Sambhaji Nagar -431005, Maharashtra	<b>Issue Date:</b> 13-04-2024 <b>Reference Date:</b> 10-04-2024 <b>Calibration Date:</b> 10-04-2024 <b>Calibration Due Date:</b> 09-04-2025

**Details of Unit Under Calibration :**

<b>Description :</b> Refrigerator	
<b>Make :</b> KELVINATOR	<b>Model :</b> KCP314-NUTRICOOL PLUS
<b>Range :</b> 2 to 8 °C	<b>SI.No. :</b> NA
<b>Resolution :</b> NA	<b>ID.No. :</b> NA
<b>Calibrated At :</b> At Site	<b>Condition on Receipt :</b> OK

**Environmental Conditions :** Temperature : (25±4)°C Relative Humidity : (50±20) % RH**Calibration SOP/Ref Standard :** SOP-TL-01, DKD-R5-1**Thermal Calibration-Temperature****Master Instruments Detail**

Name of the Master used	Id.No.	Certificate No.	Valid Upto
Universal Data Logger With Sensors (8-Channel)	SL/SMT/DL/02	SL2307TL0642-001	27/07/2024

**Calibration Results**

S.No.	Average Standard Reading in (°C)	Average UUC Reading in (°C)	Error in (°C)	Expanded Uncertainty in ± °C
1	4.2	5	0.8	0.703

**Remarks :**

- This certificate pertains only to the item calibrated.
- The calibration results reported in this certificate are valid at the time of and at the stated environmental conditions.
- The calibration interval is determined based on customer's requirements.
- The calibration is traceable to National standards as per traceability details given in the certificate.
- This calibration certificate shall not be reproduced in full, except with prior written approval of Managing Director, SIMCO Calibration & Testing Pvt. Ltd.
- This calibration certificate is meant for scientific and industrial purpose only.
- The NABL Symbol is used as per NABL guidelines in NABL-133.
- The Expanded Uncertainty is reported at 95% confidence level with approximate coverage factor k= 2.
- UOM = Unit of Measurement

Battu Nagarjuna  
Technical Executive Engineer  
Calibrated By



\*\* End of Certificate \*\*

Digitally signed by  
N.V.Kameswararao  
Date: 2024.04.16 12:41:33 +05:30  
Reason: Calibration Certificate

Authorized Signatory  
Chief Executive Officer

**CALIBRATION CERTIFICATE**

In accordance with ISO / IEC-17025 : 2017

F10-CC-03

Page 1 of 1

<b>Certificate No.:</b> SCT/100424/17/3	<b>ULR :</b> CC280624000004709F
<b>Customer Name &amp; Address :</b> M/s. TAPADIA DIAGNOSTIC CENTRE., At BANSILAL NAGAR, STATION ROAD, AURANGABAD, Chhatrapati Sambhaji Nagar -431005, Maharashtra	<b>Issue Date</b> : 13-04-2024 <b>Reference Date</b> : 10-04-2024 <b>Calibration Date</b> : 10-04-2024 <b>Calibration Due Date</b> : 09-04-2025

**Details of Unit Under Calibration :**

<b>Description</b> : Refrigerator	
<b>Make</b> : SAMSUNG	<b>Model</b> : RT25M-FRESHTECH
<b>Range</b> : 2 to 8 °C	<b>Sl.No.</b> : NA
<b>Resolution</b> : NA	<b>ID.No.</b> : NA
<b>Calibrated At</b> : At Site	<b>Condition on Receipt</b> : OK

**Environmental Conditions :** Temperature : (25±4)°C Relative Humidity : (50±20) % RH**Calibration SOP/Ref Standard :** SOP-TL-01, DKD-R5-1**Thermal Calibration-Temperature****Master Instruments Detail**

Name of the Master used	Id.No.	Certificate No.	Valid Upto
Universal Data Logger With Sensors (8-Channel)	SL/SMT/DL/02	SL2307TL0642-001	27/07/2024

**Calibration Results**

S.No.	Average Standard Reading in (°C)	Average UUC Reading in (°C)	Error in (°C)	Expanded Uncertainty in ± °C
1	4.6	5	0.4	0.703

**Remarks :**

- This certificate pertains only to the item calibrated.
- The calibration results reported in this certificate are valid at the time of and at the stated environmental conditions.
- The calibration interval is determined based on customer's requirements.
- The calibration is traceable to National standards as per traceability details given in the certificate.
- This calibration certificate shall not be reproduced in full, except with prior written approval of Managing Director, SIMCO Calibration & Testing Pvt. Ltd.
- This calibration certificate is meant for scientific and industrial purpose only.
- The NABL Symbol is used as per NABL guidelines in NABL-133.
- The Expanded Uncertainty is reported at 95% confidence level with approximate coverage factor k= 2.
- UOM = Unit of Measurement

Battu Nagarjuna  
Technical Executive Engineer  
Calibrated By

**\*\* End of Certificate \*\***

Digitally signed by  
N.V.Kameswararao  
Date: 2024.04.16 12:41:45 +05:30  
Reason: Calibration Certificate

Authorized Signatory  
Chief Executive Officer