

## 115<sup>th</sup> IAMM EQAS Microbiology: Bacteriology/ Serology CMC MICRO EQAS

Department of Clinical Microbiology, Christian Medical College, Vellore-632004, Tamil Nadu Email: <a href="mailto:eqas@cmcvellore.ac.in">eqas@cmcvellore.ac.in</a>, Twitter: @microeqas, Phone: +91-416-2282588



OCTOBER 2023

### 115<sup>th</sup> EQAS – EVALUATION REPORT

MEMBER ID:

M 0 0 6 0

Marks Obtained: 4/4 (100%)

#### OCTOBER 2023 / SEROLOGY

Test method employed for detection C-reactive protein (CRP) at your lab: Turbidimetry

Peer group (n) = 502

Please refer the attached evaluation format/answer template for details on the criteria for evaluation.

	Parameter	Your Result	Your Value (mg/L)	Intended Result	Robust Mean	Robust SD	Range (mg/L)	Uncertainty of Assigned value	Z & Z' score	Max Marks	Your Score
SE1	CRP	Positive	59.5	Positive	48.01	7.94	0 - 480	0.44	1.4	2	2
SE2	CRP	Negative	1.8	Negative	1.58	0.70	0 - 53.4	0.03	0.3	2	2

#### OCTOBER 2023 / SEROLOGY

Test method employed for detection Antistreptolysin O (ASO) your lab: NOT DONE

Please refer the attached evaluation format/answer template for details on the criteria for evaluation.

,	Parameter	Your Result	Your Value (IU/mL)	Intended Result	Robust Mean	Robust SD	Range (IU/mL)	Uncertainty of Assigned value	Z & Z' score	Max Marks	Your Score
SE3	ASO			NOT DONE							

#### Disclaimer:

This is a confidential document and subject to the rules of confidentiality as described by the ISO 17043:2010 standard.

MEMBER ID:

M	0	0	6	0
---	---	---	---	---

SE1	SE2	SE3	Marks	obtained
2	2	NOT	4	100%
2	2	DONE	Maximu	m marks = 4

Dr. Rani Diana Sahni Scientific Co-ordinator

Jan Maria

Dr. John A Jude Prakash Quality Manager Dr. V. Balaji PT Co-ordinator

Report Dispatch Date: 26.02.24

\*\*\*\*\*\*\*\*\*\*\* End of Report \*\*\*\*\*\*\*\*\*\*\*



# The Department of Clinical Microbiology, Christian Medical College, Vellore - 632 004.



This is to certify that

### **Pathology Laboratory**

## Vinamra Swaraj Hospital, Navi Mumbai

LAB CODE NO: M0060

Participated in the Microbiology External Quality Assessment Scheme

Basic Bacteriology and Serology: Tier 1 under the aegis of

Indian Association of Medical Microbiologists (IAMM) for the year 2023.

Balaji Veeraraghavan, MD., Ph.D., FRCP.,

Microbiology EQAS Coordinator