



RELIABLE TECHNICAL SERVICES

(Division of Reliable Technocare Pvt. Ltd.)

"RELIABLE HOUSE" 497/2834-35, Sant Tukaram Nagar,
Pimpri, Pune - 411018. MH, India. Telefax : 020-27421170
Cell : 7774055755, 7774055855, 7774058855, 7774022900
Email : reliable1010@gmail.com/reliabletechnocare@gmail.com
Web. : www.reliable.world



CC-2927

CALIBRATION CERTIFICATE

1. CUSTOMER	:-	Page No.	:- 1 of 1
Vinamra Swaraj Hospital		NABL Accreditation No.	:- CC-2927
Plot 108, 108/1, 109,		Certificate No.	:- 24.03.09.001
Sector 11, Vashi, Navi Mumbai-400 703		Date of Received	:- 09.03.2024
Temperature. (°C)	:- 22.5	Date of Calibration	:- 09.03.2024
Relative Humidity (% RH)	:- 51	Next Calibration Due On	:- 08.03.2025
Location of calibration	:- In Lab/On site	Calibration method No.	:- RTS-WI-06
Condition of Item	:- OK	Date of Issue	:- 11.03.2024
		ULR No.	:- CC292724000002041F

2. Description of Item			
Name	:- Centrifuge Machine	Range	:- 0 To 5250 RPM
Id No	:- CF-01	Least Count	:- 10 RPM
Make	:- REMI/R 8C		
Type	:- Digital	Location	:- Pathology

3. Details of Equipment used for calibration	
Name	:- DIG. TACHOMETER
Certificate No.	:- HT/CC/230922-12/001
Certified By	:- HI TECH
ID/Sr. No.	:- RTS-RPM-03
Range	:- 1 to 99999 RPM
Uncertainty	:- ± 3.43 RPM
Calibration Validity	:- 26.09.2024

Discipline :- Mechanical Calibration **Group** :- Acceleration, Speed & Acoustics-Speed

4. Calibration Results				
Calibration Points	Standard Reading	UUC Reading	Error in	Expanded Uncertainty in±
RPM	RPM	RPM	RPM	%
1000	1001	1000	-1	0.14
2000	2004	2000	-4	0.14
3000	3006	3000	-6	0.14
4000	4009	4000	-9	0.14
5250	5259	5250	-9	0.14

- Note:**
- 1) The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor $k=2$, which corresponds to a coverage probability of approximately 95% for normal distribution
 - 2) This certificate refers only to the particular item submitted for calibration. UUC stands for Unit Under Calibration.
 - 3) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.
 - 4) Calibration point were selected as per customer specifications.
 - 5) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Reliable Technical Services, Pune".
 - 6) The Standard used are traceable to National / International Standard

Calibrated By

S. Kore
Calibration Engineer
Sagar Kore
RF-21, R0



Approved By

R. Sable
Technical Manager
RAHUL SABLE



RELIABLE TECHNICAL SERVICES

(Division of Reliable Technocare Pvt. Ltd.)

"RELIABLE HOUSE" 497/2834-35, Sant Tukaram Nagar,
Pimpri, Pune - 411018. MH, India. Telefax : 020-27421170
Cell : 7774055755, 7774055855, 7774058855, 7774022900
Email : reliable1010@gmail.com/reliabletechnocare@gmail.com
Web. : www.reliable.world



CC-2927

CALIBRATION CERTIFICATE

1. CUSTOMER	:-	Page No.	:- 1 of 1
Vinamra Swaraj Hospital		NABL Accreditation No.	:- CC-2927
Plot 108, 108/1, 109,		Certificate No.	:- 24.03.09.002
Sector 11, Vashi, Navi Mumbai-400 703		Date of Received	:- 09.03.2024
Temperature. (°C)	:- 22.5	Date of Calibration	:- 09.03.2024
Relative Humidity (% RH)	:- 51	Next Calibration Due On	:- 08.03.2025
Location of calibration	:- In Lab/On site	Calibration method No.	:- RTS-WI-06
Condition of Item	:- OK	Date of Issue	:- 11.03.2024
		ULR No.	:- CC292724000002042F

2. Description of Item			
Name	:- Centrifuge Machine	Range	:- 0 To 6000 RPM
Id No	:- CF-02	Least Count	:- 10 RPM
Make	:- REMI/R 8C		
Type	:- Digital	Location	:- Pathology

3. Details of Equipment used for calibration	
Name	:- DIG. TACHOMETER
Certificate No.	:- HT/CC/230922-12/001
Certified By	:- HI TECH
ID/Sr. No.	:- RTS-RPM-03
Range	:- 1 to 99999 RPM
Uncertainty	:- ± 3.43 RPM
Calibration Validity	:- 26.09.2024

Discipline :- Mechanical Calibration **Group** :- Acceleration, Speed & Acoustics-Speed

4. Calibration Results				
Calibration Points	Standard Reading	UUC Reading	Error in	Expanded Uncertainty in±
RPM	RPM	RPM	RPM	%
500	500.5	500	-0.5	0.12
1000	1002	1000	-2	0.12
2000	2005	2000	-5	0.12
4000	4008	4000	-8	0.12
6000	6010	6000	-10	0.12

- Note:**
- 1) The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor k=2, which corresponds to a coverage probability of approximately 95% for normal distribution
 - 2) This certificate refers only to the particular item submitted for calibration. UUC stands for Unit Under Calibration.
 - 3) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.
 - 4) Calibration point were selected as per customer specifications.
 - 5) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Reliable Technical Services, Pune".
 - 6) The Standard used are traceable to National / International Standard.

Calibrated By
S. Kore
Calibration Engineer
Sagar Kore



Approved By
R. Sable
Technical Manager
RAHUL SABLE



RELIABLE TECHNICAL SERVICES

(Division of Reliable Technocare Pvt. Ltd.)

"RELIABLE HOUSE" 497/2834-35, Sant Tukaram Nagar,
Pimpri, Pune - 411018. MH, India. Telefax : 020-27421170
Cell : 7774055755, 7774055855, 7774058855, 7774022900
Email : reliable1010@gmail.com/reliabletechnocare@gmail.com
Web. : www.reliable.world



CC-2927

CALIBRATION CERTIFICATE

1. CUSTOMER	:-	Page No.	:- 1 of 1
Vinamra Swaraj Hospital		NABL Accreditation No.	:- CC-2927
Plot 108,108/1,109,		Certificate No.	:- 24.03.09.003
Sector 11,Vashi ,Navi Mumvai-400 703		Date of Received	:- 09.03.2024
Temperature. (°C)	:- 22.9	Date of Calibration	:- 09.03.2024
Relative Humidity (% RH)	:- 55	Next Calibration Due On	:- 08.03.2025
Location of calibration	:- In Lab/On site	Calibration method No.	:- RTS-WI-06
Condition of Item	:- OK	Date of Issue	:- 11.03.2024
		ULR No.	:- CC292724000002043F

2. Description of Item			
Name	:- Centrifuge Machine	Range	:- 0 To 5250 RPM
Id No	:- CF-03	Least Count	:- 10 RPM
Make	:- REMI/R 8C		
Type	:- Digital	Location	:- Pathology

3. Details of Equipment used for calibration	
Name	:- DIG. TACHOMETER
Certificate No.	:- HT/CC/230922-12/001
Certified By	:- HI TECH
ID/Sr. No.	:- RTS-RPM-03
Range	:- 1 to 99999 RPM
Uncertainty	:- ± 3.43 RPM
Calibration Validity	:- 26.09.2024

Discipline :- Mechanical Calibration **Group** :- Acceleration, Speed & Acoustics-Speed

4. Calibration Results				
Calibration Points	Standard Reading	UUC Reading	Error in	Expanded Uncertainty in±
RPM	RPM	RPM	RPM	%
1000	1002	1000	-2	0.14
2000	2005	2000	-5	0.14
3000	3007	3000	-7	0.14
4000	4010	4000	-10	0.14
5250	5263	5250	-13	0.14

- Note:**
- 1) The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor k=2, which corresponds to a coverage probability of approximately 95% for normal distribution
 - 2) This certificate refers only to the particular item submitted for calibration. UUC stands for Unit Under Calibration.
 - 3) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.
 - 4) Calibration point were selected as per customer specifications.
 - 5) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Reliable Technical Services, Pune".
 - 6) The Standard used are traceable to National / International Standard

Calibrated By

S. Kore

Calibration Engineer
Sagar Kore



Approved By

R. Sable

Technical Manager
RAHUL SABLE

CALIBRATION CERTIFICATE			
Certificate No.	HT/CC/230922-12/001	ULR No.	ULR - CC215123000037486F
CUSTOMER DETAILS		DEVICE UNDER CALIBRATION DETAILS	
Name	M/s. Reliable Technical Services,	Description	Digital Tachometer
Address	(Division of Reliable Technocare Pvt. Ltd.), Reliable House, 497/2834-35, Sant Tukaram Nagar, Pimpri, Pune - 411 018.	Make	Fluke Model 931
		Mode	Non - Contact Contact
		Range	1 to 99999 RPM 1 to 19999 RPM
Job No.	230922-12	Least Count	0.01/0.1/1 RPM
Job Receipt Date	22-Sep-2023	Accuracy	± (0.02 % rdg + 1 digit)
Issue Date	30-Sep-2023	Serial No.	1736017
Received Condition of DUC	Satisfactory	Identification	RTS-RPM-03
ENVIRONMENTAL CONDITION		Calibration On	27-Sep-2023
Temperature	25 °C ± 2 °C	Due On	26-Sep-2024
Humidity	30 % RH to 75 % RH	Calibrated By	Mohana. V
Discipline & Group	Mechanical Calibration & Acceleration, Speed		
Calibration Procedure	HT/WI/M/AS (Section 01) & SANAS TR 45-02 by Comparison Method		
STANDARD DETAILS			
Description	Digital Tachometer		
Identification No.	M/AS/TACHO/08 & M/AS/TACHO/05		
Certificate No.	SALPR/0284/1/23 & HT/CC/230306-17/001		
Valid Till	03-Apr-2024 & 05-Mar-2024		

Refer Next Page for Result

CALIBRATION CERTIFICATE			
Certificate No.	HT/CC/230922-12/001	ULR No.	ULR - CC215123000037486F

CALIBRATION RESULT

Mode : Non - Contact

DUC Reading in RPM	Standard Reading in RPM	Error in RPM	Expanded Uncertainty in ± RPM
10.04	10.033	0.007	0.26
50.07	50.067	0.003	0.26
90.07	90.067	0.003	0.27
100.1	100.1	0.0	0.36
500.5	500.6	-0.1	0.42
900.9	900.9	0.0	0.43
1001	1001.0	0.0	0.73
10022	10022.2	-0.2	1.82
50004	50004.1	-0.1	3.06
90007	90007.0	0.0	3.43

Mode : Contact

DUC Reading in RPM	Standard Reading in RPM	Error in RPM	Expanded Uncertainty in ± RPM
10.04	10.035	0.005	0.26
50.22	50.225	-0.005	0.27
90.15	90.143	0.007	0.27
100.3	100.1	0.2	0.32
500.5	500.5	0.0	0.42
900.5	900.4	0.1	0.47
1000	1000.4	-0.4	0.76
5002	5002.0	0.0	0.82
10005	10005.4	-0.4	1.79
15000	15000.4	-0.4	1.79

Remarks:

- The Calibration is performed at Permanent Facility.
- The Calibration Certificate is valid for scientific & industrial purpose only.
- The Expanded uncertainty is expressed at 95.45 % confidence level with coverage factor k=2.

Checked By
M. DEEPAK RAM - Sr. CALIBRATION ENGINEER
Ambarore G - 7B - FI



Approved By
R.R. HANUMAN - TECHNICAL MANAGER
Page 1 of 2



Checked By
M. DEEPAK RAM - Sr. CALIBRATION ENGINEER
Ambarore G - 7B - FI



Approved By
R.R. HANUMAN - TECHNICAL MANAGER
Page 2 of 2 | End of Certificate



TAX INVOICE

(ORIGINAL FOR RECIPIENT)

SCAN & PAY



UPI ID: REMI.RSEL@SBI

REMI Remi Sales & Engineering Limited
 Building "B", Survey No. 65/1, Waliv, Vasai (E)
 Dist. Palghar - 401 208, Maharashtra.
 CIN - U31100MH1980PLC022314 -NEVV
 Drug Licence-MH-T22-222024 Validity 12-07-2022 (Discontinued w.e.f 31.12.2021)
 Drug Licence-MH-T22-222025 Validity 12-07-2022 (Discontinued w.e.f 31.12.2021)
 R.O.: Remi House, Plot No.11, Cama Industrial Estate,
 Walibhat Road, Goregaon (East), Mumbai-400063
 Tel.No. 022- 40589888
 GSTIN/UIN: 27AAACR0421M1ZJ
 State Name : Maharashtra, Code : 27
 CIN: L31100MH1980PLC022314
 E-Mail : customercare@remilabworld.Com

Invoice No. RSLG/0379/24	Dated 29-Feb-24
Delivery Note	Mode/Terms of Payment CHQ NO : 035407 DT : 17.02.2024
Reference No. & Date. AMC/0061/24 dt. 29-Feb-24	Other References PROP/RSL/23-24/0246
Buyer's Order No. AMC/0061/24	Dated 19-Feb-24
Dispatch Doc No. DIRECT	Delivery Note Date
Dispatched through BY HAND	Destination NAVI MUMBAI
Terms of Delivery DOOR DELIVERY	

Consignee (Ship to)
DR.S.TECKCHANDANI
 VINAMRA Swaraj Hospital, PLOT 108,
 109, SECTOR 11, OPP. SHVE, SENA
 OFFICE, NAVI MUMBAI-400703
 State Name : Maharashtra, Code : 27
 Contact person : Dr S Teckchandani
 Contact : 9820384505
 E-Mail : pritesh10@hotmail.com

Buyer (Bill to)
DR.S.TECKCHANDANI
 VINAMRA Swaraj Hospital, PLOT 108,
 109, SECTOR 11, OPP. SHVE, SENA
 OFFICE, NAVI MUMBAI-400703
 State Name : Maharashtra, Code : 27
 Contact person : Dr S Teckchandani
 Contact : 9820384505
 E-Mail : pritesh10@hotmail.com

Sl No.	Description of Services	HSN/SAC	GST Rate	Quantity	Rate	per	Disc. %	Amount
1	Annual Maintenance Contract R-8C, ZBHN-23404 R-8C PLUS, ZDFN-15216 PERIOD - 16.02.2024 TO 15.02.2025	998719	18 %	2 Visit	5,500.00	Visit		11,000.00
								990.00
								990.00
								12,980.00
				2 Visit				12,980.00

Amount Chargeable (in words) **Rupees Twelve Thousand Nine Hundred Eighty Only** E. & O.E

HSN/SAC	Taxable Value	Central Tax		State Tax		Total Tax Amount
		Rate	Amount	Rate	Amount	
998719	11,000.00	9%	990.00	9%	990.00	1,980.00
Total	11,000.00		990.00		990.00	1,980.00

Tax Amount (in words) : **Rupees One Thousand Nine Hundred Eighty Only**

Company's PAN : **AAACR0421M**

Declaration
 We declare that this invoice shows the actual price of the goods described and that all particulars are true and correct.

Company's Bank Details
 A/c Holder's Name: **Remi Sales & Engineering Limited**
 Bank Name : **S.B.I. CURRENT A/C. (MALAD)**
 A/c No. : **39976505913**
 Branch & IFS Code: **Industrial Finance Branch & SBIN0004760**
 SWIFT Code :

for Remi Sales & Engineering Limited

 Authorised Signatory