



BIO-RAD LABORATORIES (INDIA) PRIVATE LIMITED

Site name:	VINAMRA SWARAJ HOSPITAL
Address:	PLOT# 108, 108A & 109, Sector 11, Vashi
Province-Postal code	400703
Contact Name	Ms. Shriti

Calibration Certificate

Date	26/09/2023 <small>(dd mmmm, yyyy)</small>	Instrument Model	D-10 Hb Analyzer	Instrument Serial Number	DJ7B056501
-------------	--	-------------------------	------------------	---------------------------------	------------

We certify that the Bio-Rad instrument specified above and installed in your lab is calibrated and the performance was found satisfactory.

High Pressure Pump Flow Rate Reference: Service Manual: Section 7.7.3

Measured with <small>(Device manufacturer and model)</small>	Device Serial Number	Device Calibration date <small>(dd mmmm, yyyy)</small>
GlasscoMeasurung Cylinder Hex Base: 10ml	11.20/2118	21/11/2020
Specified Range@ 1.5ml/Minute	Measured Value	
1.425 – 1.575 ml/minute	1.5	

Column Holder temperature Reference: Service Manual: Section 7.9

Measured with <small>(Device manufacturer and model)</small>	Device Serial Number	Device Calibration date <small>(dd mmmm, yyyy)</small>
Fluke 179	36030142	24/02/2023
Specified Range @ 35°C	Measured Value	
33.4°C – 36.4°C	35.1	

Detector log value Reference: Service Manual: Section 7.2 or Field Bulletin D-10-081

Measured with	Specified Range	Measured Value
Instrument Service Software	0.018 – 0.022 AU	0.020

The next calibration is due on	26/09/2024	<small>(dd mmmm, yyyy)</small>
--------------------------------	------------	--------------------------------

For Bio-Rad Laboratories India Pvt Limited

Authorized Signatory:



Acumen Measurements & Consultancy Pvt. Ltd.

(Calibration Division)



D-10/6, First Floor, Okhla Industrial Area, Phase-1, New Delhi - 110020
 Ph.: 011-42701221, 42701222, 42701223 Mobile: 8585965526, 9350380535
 E: info@acumendelhi.com, calibration@acumendelhi.com W: www.acumendelhi.com

CC-2364

CALIBRATION CERTIFICATE

Page 1 of 1

Format No.: 7.8F-01

ULR No.: CC236423000000263F	Certificate No.: AMC/00000263
CUSTOMER NAME & ADDRESS:- M/s. Bio-Rad Laboratories (India) Pvt Ltd Emaar Digital Greens, 9th Floor, Tower A, Golf Course Extension Road, Sec. - 61, Gurgaon - 122102	Service Request No.&Date: 230050 & 22-02-2023
	Date of Calibration: 24-02-2023
	Sug. Cal. Due Date: 23-02-2024
	Issue Date: 27-02-2023

ITEM'S DESCRIPTION & IDENTIFICATION

Instrument Name : DIGITAL MULTIMETER	Range / Size : Refer to Result
Make : Fluke	Least Count : As per Range
Model No. : 179	Visual Inspection : OK
Serial No. : 36030142	Location : N/A
Party ID Mark : N/A	
ENVIRONMENTAL CONDITIONS	
Temperature : (25 ± 2)°C	Calibration Procedure : ECP-02,03,04,13,14 & TCP-03
Relative Humidity : (60 ± 15)%	Reference Standard : IS: 13875 (Part 1 & 2)
	Calibration Performed at : Laboratory

MAJOR EQUIPMENT USED	Ident. No.	Traceability Certificate No.	Valid upto
Nomenclature			
Multi Products Calibrator (Fluke 5502A)	4451802 (MEI/26)	TSC/22-23/2023-2	25-04-23
Univarsal Calibrator (Fluke 9100)	958756002(MEI/01)	TSC/22-23/2023-1	25-04-23
SSPRT Sensor With DAQ	MTI/19	TSC/22-23/4380-1	15-06-23

CALIBRATION RESULT

Standard Reading	*DUC Reading	Error (±)	Uncertainty of Measurement (±)(%)
DC Voltage Measurement			
1.000 mV	1.0 mV	0.0 mV	5.80
10.000 mV	10.0 mV	0.0 mV	0.60
100.000 mV	100.0 mV	0.0 mV	0.06
1.00000 V	1.000 V	0.000 V	0.06
10.0000 V	10.00 V	0.00 V	0.06
100.000 V	100.0 V	0.0 V	0.06
600.00 V	600.1 V	0.1 V	0.06
1000.00 V	1000 V	0 V	0.06
AC Voltage Measurement (at 50 Hz.)			
10.000 mV	10.1 mV	0.1 mV	1.15
100.000 mV	100.3 mV	0.3 mV	0.12
1.00000 V	1.003 V	0.003 V	0.07
10.0000 V	10.03 V	0.03 V	0.07
100.000 V	100.2 V	0.2 V	0.08
600.00 V	600.0 V	0.0 V	0.10
1000.00 V	999 V	-1 V	0.10

Issued & Authorised By

Munesh Jadaun
(Calibrated By)



U C Sutar
(TM / QM)



Acumen Measurements & Consultancy Pvt. Ltd.

(Calibration Division)



D-10/6, First Floor, Okhla Industrial Area, Phase-1, New Delhi - 110020
 Ph: 011-42701221, 42701222, 42701223 Mobile: 8585965526, 9350380535
 E: info@acumendelhi.com, calibration@acumendelhi.com W: www.acumendelhi.com **CC-2364**

CALIBRATION CERTIFICATE

Page 2 of :

Format No.:7.8F-01

ULR No. : CC236423000000263F	Certificate No.	: AMC/00000263
CUSTOMER NAME & ADDRESS:- M/s. Bio-Rad Laboratories (India) Pvt Ltd Emaar Digital Greens, 9th Floor, Tower A, Golf Course Extension Road, Sec. - 61, Gurgaon - 122102	Service Request No.&Date	: 230050 & 22-02-2023
	Date of Calibration	: 24-02-2023
	Sug. Cal. Due Date	: 23-02-2024
	Issue Date	: 27-02-2023

Standard Reading	*DUC Reading	Error (±)	Uncertainty of Measurement (±)(%)
DC Current Measurement			
1.00000 mA	1.01 mA	0.01 mA	0.58
10.0000 mA	10.01 mA	0.01 mA	0.06
100.000 mA	100.0 mA	0.00 mA	0.06
1.00000 A	1.000 A	0.000 A	0.08
10.0000 A	10.00 A	0.00 A	0.10
AC Current Measurement (at 50 Hz)			
1.00000 mA	1.01 mA	0.01 mA	0.60
10.0000 mA	10.02 mA	0.02 mA	0.30
100.000 mA	100.3 mA	0.3 mA	0.30
1.00000 A	1.000 A	0.000 A	0.50
10.0000 A	10.00 A	0.00 A	3.50
Resistance Measurement			
1.0000 Ω	1.1 Ω	0.1 Ω	5.80
10.0000 Ω	10.1 Ω	0.1 Ω	0.60
100.000 Ω	100.0 Ω	0.0 Ω	0.06
1.00000 kΩ	1.000 kΩ	0.000 kΩ	0.06
10.0000 kΩ	10.00 kΩ	0.00 kΩ	0.06
100.000 kΩ	100.0 kΩ	0.0 kΩ	0.06
1.00000 MΩ	1.000 MΩ	0.000 MΩ	0.06
10.0000 MΩ	10.00 MΩ	0.00 MΩ	0.06
30.0000 MΩ	29.98 MΩ	-0.02 MΩ	0.07

Issued & Authorised By

Munesh Jadaun
 (Calibrated By)



U C Sutar
 (TM / QM)



Acumen Measurements & Consultancy Pvt. Ltd.

(Calibration Division)



D-10/6, First Floor, Okhla Industrial Area, Phase-1, New Delhi - 110020
 Ph.: 011-42701221, 42701222, 42701223 Mobile: 8585965526, 9350380535

E: info@acumendelhi.com, calibration@acumendelhi.com W: www.acumendelhi.com

CC-2364

CALIBRATION CERTIFICATE

Page 3 of 3

Format No.:7.8F-01

ULR No. : CC236423000000263F	Certificate No.	: AMC/00000263
CUSTOMER NAME & ADDRESS:- M/s. Bio-Rad Laboratories (India) Pvt Ltd Emaar Digital Greens, 9th Floor, Tower A, Golf Course Extension Road, Sec. - 61, Gurgaon - 122102	Service Request No.&Date	: 230050 & 22-02-2023
	Date of Calibration	: 24-02-2023
	Sug. Cal. Due Date	: 23-02-2024
	Issue Date	: 27-02-2023

Standard Reading	*DUC Reading	Error (±)	Uncertainty of Measurement
Temperature Probe Measurement			
(°C)	(°C)	(°C)	(°C)
0.009	0.0	0.009	0.49
49.926	49.9	0.026	0.49
99.632	99.4	0.232	0.75
151.025	150.6	0.425	0.75
198.531	197.4	1.131	0.75




REMARKS :

1. DUC = Device Under Calibration, UUC=Unit Under Calibration i.e.
2. Expanded uncertainty of measurement at approx 95% confidence level for coverage factor(k)=2
3. This result of Calibration refer to only to the particular item submitted for calibration.
4. Suggested next due date of calibration is given as requested by customer.
5. DUC has been calibrated against lab standards are traceable to national & international standards.
6. The result reported in this certificate are valid at the time of and under the stated conditions of Measurement.
7. This certificate shall not be reproduced, except in full, unless a written permission for the publication of an approved abstract has been obtained from the CEO, Acumen Measurements & Consultancy Pvt. Ltd., New Delhi.

Issued & Authorised By

Munesh Jadaun
 Munesh Jadaun
 (Calibrated By)



U C Sutar
 U C Sutar
 (TM / QM)



Glassco Laboratory Equipments Pvt. Ltd.

CALIBRATION CERTIFICATE

(Glassco Calibration Laboratory)

(ISO 17025:2017 Certified laboratory)

CERTIFICATE NO	GLASSCO/20/11-293/2118		
CRF	29320112020	CAL. REPORT NO	20/11-293

CUSTOMER	LCP-STOCK
TEST METHOD	GRAVIMETRIC METHOD
ITEM CALIBRATED	MEASURING CYLINDER HEX BASE : 10 ml
CATALOGUE NUMBER	139.523.01A
BATCH NO. & SERIAL NO.	11.20/2118
ITEM CALIBRATED ON	21/11/2020

CALIBRATION RESULT:

The product was found to contain the following volume at 27°C

VOLUME CONTAINED	9.990 ml
TOLERANCE	± 0.100 ml
CORRECTION	0.010 ml

Expanded uncertainty (at 95% C.L.k=2) is ± 0.001 ml

The product has been calibrated with the following devices:

Standard weights	10 g
Weighing Balance	ABAL-30,1020g/0.001 g(METTLER)
Thermometer	DM-35:-50°C to 300°C/0.1°C

All the above devices have valid calibrations traceable to NPL India; traceable to international standards.

The above product has been examined for compliance to ISO 4788:2005 Class-A. It has been found to satisfy the specified requirements. The above item has been calibrated at (25±3)°C at RH (60±25)%.

The above calibration result is valid at the time of calibration under lab conditions. This report shall not be reproduced, except in full, unless with written permission from Glassco Laboratory Equipments Pvt. Ltd.

Date of issue: 21/11/2020

This is an electronically generated certificate.

Authorised Signatory

Manager Technical
(Yogendra Panwar)

Page 1 of 1



CALIBRATION CERTIFICATE

CMC-SALE

(ORIGINAL FOR RECIPIENT)

Mass Healthcare

Shop No. 17, A Wing, Ground Floor, "COSMOS ARCADE"
 Bramhand, Kolshet, Thane (West) - 400607
 Maharashtra, INDIA
 DRUG LICENCE NO. -
 20B - MH 172-487914, 21B - MH-TZ2-487915
 Valid From 13.08.2022 to 12.08.2027
 GSTIN/UIN: 27ARJPG2233M1ZM
 State Name : Maharashtra, Code : 27
 E-Mail : mass.alroll@gmail.com
 Consignee (Ship to)

DR. S.D. TECKCHANDANI

VINAMRA SWARAJ HOSPITAL, PLOT NO. 108
 108A & 109, OPPO. SHIV SENA SHAKHA,
 SECTOR 11, VASHI, NAVI MUMBAI - 400 703
 State Name : Maharashtra, Code : 27

Buyer (Bill to)

DR. S.D. TECKCHANDANI

VINAMRA SWARAJ HOSPITAL, PLOT NO. 108
 108A & 109, OPPO. SHIV SENA SHAKHA,
 SECTOR 11, VASHI, NAVI MUMBAI - 400 703
 State Name : Maharashtra, Code : 27
 Place of Supply : Maharashtra

Invoice No. **CMC-23-24-0037** Dated **12-Jan-24**
 Delivery Note Mode/Terms of Payment
30 Days
 Dated
 Buyer's Order No. Dated
 Dispatch Doc No. Delivery Note Date
 Dispatched through Destination
 Terms of Delivery
 Customer Order No. :
 Customer Order Dt. :

Sl No.	Description of Services	HSN/SAC	Part No.	Quantity		Rate	per	Disc. %	Amount
				Shipped	Billed				
1	CMC - D-10 HEMOGLOBIN TESTER GOLD Batch : SC-04113422 Expiry : 15-Nov-24 S.NO. - DJ7B056501 VALID FROM 16.11.2023 TO 15.11.2024	698 - 2253		1 CMC	1 CMC	85,000.00	CMC		85,000.00
				1 CMC	1 CMC				
	SERVICE OUTPUT CGST 9%						9 %		7,650.00
	SERVICE OUTPUT SGST 9%						9 %		7,650.00

Total 1 CMC 1 CMC ₹ 1,00,300.00

Amount Chargeable (in words)

INR One Lakh Three Hundred Only


E & O E

HSN/SAC	Taxable Value	Rate	CGST Amount	SGST/UTGST Rate	SGST/UTGST Amount	Total Tax Amount
	85,000.00	9%	7,650.00	9%	7,650.00	15,300.00
Total	85,000.00		7,650.00		7,650.00	15,300.00

Tax Amount (in words) : **INR Fifteen Thousand Three Hundred Only**

Company's Bank Details

Bank Name : **THE FEDERAL BANK LTD.- 673**
 A/c No. : **13835500050673**
 Branch & IFS Code: **TEEN PETROL PUMP, THANE (W) & FDRL0001383**



SUBJECT TO THANE JURISDICTION

This is a Computer Generated Invoice

MASS HEALTHCARE

Sales Office: Shop no: - 17, Cosmos Arcade, A-Wing Ground Floor, Bramhand, Kolshet, Thane (West),
Thane – 400610 Mob: 8655076200 E-Mail: mass.airoli@gmail.com

12.01.2024

TO WHOMSOEVER IT MAY CONCERN

Dear Sir/Madam,

This is to inform you that, We, Mass Healthcare have placed a 'Bio-Rad D-10 system at Vinamra Swaraj Hospital, on Reagent Rental Basis.

This is to confirm that the instrument is currently covered under CMC till 15.11.24

The CMC will be renewed by us every year, till the completion of the Reagent Rental Contract.



Milind Gharpure
(Mass Healthcare)