



Metropolis Healthcare Ltd., EQAS

KRSNAA DIAGNOSTICS LTD, PATHANKOT (1052)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 230412 and Sample No 01

Report Date : 13/02/2024

Analyte	Instrument	Result Value	Standard Unit	Z-Score
! Albumin	Diasys 200 Pro	4.10	g/dL	2.54
Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : Diasys 200 Pro

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Urea	mg/dl	17.50	23.05	-1.62	-0.72
@ Creatinine	mg/dl	0.86	0.79	0.97	0.73
# Bilirubin Total	mg/dl	0.80	0.43		0.68
@ Total Protein	g/dL	5.84	5.55	0.60	0.76
! Albumin	g/dL	4.10	3.62	2.54	2.15
@ Uric Acid	mg/dl	3.82	3.62	-0.08	-0.93
@ Calcium	mg/dl	7.16	6.84	0.59	-0.16
@ Cholesterol	mg/dl	134.00	130.74	0.46	0.26
@ Triglycerides	mg/dl	133.00	122.35	0.65	0.66
@ HDL	mg/dl	30.00	31.40	0.50	0.24
@ LDL	mg/dl	64.50	72.09	-1.27	-1.22
@ ALT	U/L	14.30	11.71	1.26	0.91
@ AST	U/L	21.00	16.30	1.70	0.72
@ ALP	U/L	40.21	42.48	-0.57	-0.94
@ Potassium	mmol/L	3.82	--	0.01	-0.03

Instrument : ACCULYTE-3P

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Chloride	mmol/L	79.55	83.97	-0.44	-0.41
@ Sodium	mmol/L	112.77	106.81	0.20	0.21

Legend @ : Acceptable

! : 2.0 < z score < 3.0 - Warning


X : z score ≥ 3 - Unacceptable

: Not Evaluated

⊙ : Delayed Result Entry

* : Not considered for evaluation.

Total Parameters	17
Not Evaluated Parameters	1
Evaluated Parameters	16
Outlier Parameters (X)	0
EQAS Score Biochemistry	100.00 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS

Unit No. 409-416.

Commercial Building - 1A

Kohinoor Mall, Kirol Road, Kurla (W),

Mumbai - 400070