



Metropolis Healthcare Ltd., EQAS

KRSNAA DIAGNOSTICS LTD, PATHANKOT (1052)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 230412 and Sample No 02

Report Date : 13/02/2024

Analyte	Instrument	Result Value	Standard Unit	Z-Score
! Bilirubin Total	Diasys 200 Pro	1.10	mg/dl	2.45
! Albumin	Diasys 200 Pro	4.20	g/dL	2.35
Legend @ : z score \leq 2.0 - Acceptable ! : $2.0 < z \text{ score} < 3.0$ - Warning X : z score ≥ 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : Diasys 200 Pro

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Glucose	mg/dl	72.80	72.36	0.38	0.28
@ Urea	mg/dl	18.60	22.82	-1.25	-0.72
@ Creatinine	mg/dl	0.92	0.84	0.88	0.73
! Bilirubin Total	mg/dl	1.10	0.61	2.45	0.68
@ Bilirubin Direct	mg/dl	0.10	0.14	-0.82	-0.41
@ Total Protein	g/dL	6.22	6.04	0.32	0.76
! Albumin	g/dL	4.20	3.66	2.35	2.15
@ Uric Acid	mg/dl	4.05	3.85	-0.12	-0.93
@ Calcium	mg/dl	6.79	6.18	0.73	-0.16
@ Cholesterol	mg/dl	131.00	126.61	0.14	0.26
@ Triglycerides	mg/dl	148.00	144.28	0.00	0.66
@ LDL	mg/dl	66.97	77.25	-1.33	-1.22
@ ALP	U/L	38.38	41.55	-0.68	-0.94
@ Potassium	mmol/L	3.98	--	-0.10	-0.03

Instrument : ACCULYTE-3P

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Chloride	mmol/L	81.95	87.53	-0.62	-0.41
@ Sodium	mmol/L	113.80	108.73	0.03	0.21

Legend @ : Acceptable

! : 2.0 < z score < 3.0 - Warning


X : z score ≥ 3 - Unacceptable

: Not Evaluated

⊙ : Delayed Result Entry

* : Not considered for evaluation.

Total Parameters	16
Not Evaluated Parameters	0
Evaluated Parameters	16
Outlier Parameters (X)	0
EQAS Score Biochemistry	100.00 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS

Unit No. 409-416.

Commercial Building - 1A

Kohinoor Mall, Kirol Road, Kurla (W),

Mumbai - 400070