



# Metropolis Healthcare Ltd., EQAS

KRSNAA DIAGNOSTICS LTD, PATHANKOT (1052)

## Outlier And Analyte Summary Report

Outlier Details For Cycle No 230412 and Sample No 03

Report Date : 13/02/2024

Analyte	Instrument	Result Value	Standard Unit	Z-Score
No Outlier				
Legend @ : z score $\leq$ 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score $\geq$ 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

**Problem Classification:** \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

**Corrective Action:** \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

**Reviewed by:** \_\_\_\_\_

**Dated:** \_\_\_\_\_

Instrument : Horiba Yumizen H550

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Erythrocyte (RBC) Count	mill/cu.mm	2.07	2.03	0.42	0.00
@ Total Leucocytes (WBC) count	/c.mm	2140.00	2047.65	0.13	0.27
@ Haemoglobin	g/dL	5.60	5.47	0.47	0.13
@ MCV (Mean Corpuscular Volume)	fL	116.00	117.12	-0.18	-0.02
@ MCH (Mean Corpuscular Hb)	pg	27.00	26.96	0.01	0.06
@ MCHC (Mean Corpuscular Hb Concentration)	g/dL	23.20	23.00	0.06	0.02
@ Platelet Count	1000/ $\mu$ L	125.00	130.07	0.12	0.02
@ Hematocrit	%	24.00	23.76	0.08	-0.03
@ RDW CV%	%	19.20	19.25	0.06	-0.11

Legend @ : Acceptable

! : 2.0 < z score < 3.0 - Warning


X : z score  $\geq$ 3 - Unacceptable

# : Not Evaluated

⌚ : Delayed Result Entry

\* : Not considered for evaluation.

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	0
EQAS Score Haematology - CBC	100.00 %



**Dr Puneet Kumar Nigam**  
PT coordinator & Technical Manager, MHL EQAS  
Unit No. 409-416.  
Commercial Building - 1A  
Kohinoor Mall, Kirol Road, Kurla (W),  
Mumbai - 400070