



Metropolis Healthcare Ltd., EQAS

KRSNAA DIAGNOSTICS LTD, PATHANKOT (1052)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 230412 and Sample No 04

Report Date : 13/02/2024

Analyte	Instrument	Result Value	Standard Unit	Z-Score
No Outlier				
Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : Horiba Yumizen H550

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Erythrocyte (RBC) Count	mill/cu.mm	2.11	2.14	-0.39	0.00
@ Total Leucocytes (WBC) count	/c.mm	6060.00	5806.22	0.04	0.27
@ Haemoglobin	g/dL	6.20	6.22	-0.06	0.13
@ MCV (Mean Corpuscular Volume)	fL	121.80	122.79	-0.23	-0.02
@ MCH (Mean Corpuscular Hb)	pg	29.10	29.02	0.16	0.06
@ MCHC (Mean Corpuscular Hb Concentration)	g/dL	23.90	23.71	0.25	0.02
@ Platelet Count	1000/ μ L	234.00	227.36	0.43	0.02
@ Hematocrit	%	25.70	26.07	-0.40	-0.03
@ RDW CV%	%	18.90	18.24	0.28	-0.11

Legend @ : Acceptable

! : $2.0 < z \text{ score} < 3.0$ - Warning


X : $z \text{ score} \geq 3$ - Unacceptable

: Not Evaluated

⌚ : Delayed Result Entry

* : Not considered for evaluation.

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	0
EQAS Score Haematology - CBC	100.00 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A
Kohinoor Mall, Kirol Road, Kurla (W),
Mumbai - 400070