

## TESTNCAL LABORATORY

(A DIVISION OF TESTING & CALIBRATION INDIA (OPC) PVT. LTD)



Plot No-113, Pocket-A2, Sector-17, Dwarka, New Delhi-110075 (India) Mob. +91-8882056269, 9990730550, E-mail-testncal@gmail.com; Website - www.tnclab.com

## **Calibration Certificate**

P-17 F03-00

Valid upto

Certificate/ULR No.

Certificate/ULR No.:	CC205523000003320F		Page 1 of 1
	Customer Details	Instrument Receipt Date	24-Nov-2023
M/S. GALAXY IMAGIN	G & PATH LABS	SRF No.	TNC/JC/23-24/325
		Date of Calibration	24-Nov-2023
Sector - 7, Dwarka, New	v Delhi - 110075	Next Calibration Date	23-Nov-2024
		Certificate Issue Date	25-Nov-2023
A) Identification On Ul	JC (Unit Under Calibration)	and the second s	
Instrument Name	Centrifuge	Make / Model No.	Remi/ Remi-R-8C
Range	0 to 4000 rpm	Instrument Condition	OK
Least Count	10 rpm	Location of Calibration	At Site
Serial No.	ZDBN-02914	Location	Lab
Instrument ID.	GIPL/CF01-9	Parameter	Speed
ENVIRONMENTAL	Temperature: 25±2°C, Relative Humidity: 50±10%	Calibration Reference Standard	IS: 12508
CONDITION		Calibration Procedure	TNC -02-RPM 2.0

**Details of Test Equipments Used** 

I.D. NO./ SR NO

		2 9 9		AP 70	00/00/000
	Dig. Tachometer	l condo	TNC/RPM/C/002/19	CC233923000000317F	08/02/2024
_		The same of the sa	Calibration Results		
Г	SI. No.	UUC Value (rpm)	Std. Value (rpm)	Error (rpm)	Uncertainty at k=2 ,± %
$\vdash$	1	1000	1001	-1	0.40
$\vdash$	2	2000	2003	-3	0.40

SI. No.	UUC Value (rpm)	Std. Value (rpm)	Error (rpin)	Officertainty at K-2,176
1	1000	1001	1 1	0.40
2	2000	2003	-3	0.40
3	3000	3005	-5	0.40
4	4000	4009	-9	0.40
Mate				

- 1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.

  2. This report should not be reproduced except in full without our prior permission in writing.

  3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
- 4. All our certificates are Traceable to national/International standard.

Name

- 5. This Certificate relates to the item Calibrated total liability of Testnoal Laboratory will be limited to the invoiced amount only
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2

brated By



(Amit Kumar)

Dwarka, New Delhi-110075

Dwarka, New Delhi-110075

