



UNIQUE
CALIBRATION & MACHINES

UNIQUE CALIBRATION SOLUTIONS LLP

An ISO/IEC-17025:2017 / ACCREDITED BY NABL

Tel: 0484-4858563, 0484 - 7966984, 9971328563, 9061728563, 9778197603

E-mail: info@uniquecalibration.co.in, website: www.uniquecalibration.co.in



CC-3102

CALIBRATION CERTIFICATE				
UNIQUE LAB REPORT NO (ULR NO) :		CC310224000001066F		
SRF NO. :	091	CERTIFICATE NO	UCSL/04-24/091_06	
Customer Details :	M/s. MICRO HEALTH LABORATORIES MP TOWER, NEXT TO LAVANYA PLAZA, NEAR THALUK HOSPITAL, THAMARASSERY, KERALA - 673 573			
Device Under Calibration	CENTRIFUGE (RPM)		Equipment & Master Used For Calibration	
Make / Model No.	ELECTRA BL		01. DIGITAL CONTACT & NON CONTACT TACHOMETER	
Range	0 to 6000 rpm		Make : LUTRON	
Least Count	1 rpm		Serial No. : Q693180	
Instrument Id. No.	MHL-TSY-PQP-08		Certificate No. : C-230620-8-21	
Instrument Sr. No.	EPBL20160046		Next Due Date : 22.06.2024	
Location	LAB		Calibrated By : Godrej & Boyce	
Visual Inspection	OK		Accreditation No : CC-2559	
Material Receipt Date	---			
Calibration performed At	ON SITE			
Date of Calibration	23.04.2024			
Next calibration due (As per customer request)	23.04.2025			
Calibration Certificate Issue Date	23.04.2024			
Discipline	Mechanical			
Method Used	Comparison			
ENVIRONMENTAL CONDITION				
Temperature	27.1°C			
Relative Humidity	57%RH			
Condition Of Equipment :	Un-Loaded			
OBSERVATION RESULTS				
SL. NO.	Nominal Value on DUC (rpm)	Observed Mean Value on MASTER (rpm)	% Error of Reading (±)	± Expanded Uncertainty At Approx.95% Confidence Level (k=2)
1.	500	500.5	0.10	0.60 %
2.	1000	1006	0.56	0.60 %
3.	3000	3006	0.21	0.60 %
4.	4000	4005	0.11	0.60 %
5.	6000	6005	0.08	0.60 %
Calibration Procedure Based On		: UCSL/SOP/RPM		
Decision Rule is Applied or Not			Yes	No
				✓
NOTE :-				
1. The calibration results reported in this certificate are valid at the time of and the stated condition of measurement.				
2. This report should not be reproduced except in full without our prior permission in writing.				
3. Calibration certificate without signature are not valid.				
4. All the measurements are traceable to SI units through unbroken chain of calibration from the competent laboratories as per ISO/IEC/17025 standard requirements.				
5. DUC : Device under calibration				
6. This calibration certificate relates only to the above DUC.				
Calibrated By	Checked By		 Issued / Approved By (APPU K MANI) Technical Director	

--- End of Certificate ---
Page 1 of 1

5/D3, EK Appartments, Model engineering college road Thrikkakara, Edappally, Cochin, Kerala-682021

SALES ★ SPARES ★ SERVICE ★ TESTING ★ CALIBRATION ★ MODERNIZATION

