




CALIBRATION CERTIFICATE											
ULR NO :		CC31022400001061F				CERTIFICATE No. :		UCSL/04-24/091_01			
Customer Name & Address:						Instrument receipt Date		16.04.2024			
M/s. MICRO HEALTH LABORATORIES MP TOWER, NEXT TO LAVANYA PLAZA, NEAR THALUK HOSPITAL, THAMARASSERY, KERALA - 673 573						SRF No.		091			
						Calibration Date		17.04.2024			
						Next calibration date (As per customer request)		17.04.2025			
						Certificate Issue date		23.04.2024			
Identification on DUC (Device Under Calibration)											
Instrument Name		MICROPIPETTE				Make		DRAGON LAB			
Range		100 - 1000 µl				Model		---			
Least count		5 µl				Instrument Condition		OK			
Serial No.		YEA17AD0058373				Calibration Performed at		Mass & Volume Lab			
ID No.		MHL/TSY/13				Location		---			
Environmental Condition				Avg. Temperature (°C)		Avg. RH (%)		Avg. Atmospheric Pressure (hpa)			
				23.5		55		1010			
Equipment & Master Used For calibration											
Instrument Used		Calibrated By		ID No:		Certificate No		Valid Upto		Sr No.	
Weighing Balance		UCSL		UCSL-WB-01		UCSL/10-23/311_01		14.10.2024		0037905909	
Method Used :						Gravimetric method			Discipline & Category : Mechanical - Volume		
Calibration Reference Standard						Calibration Procedure					
ISO-8655-6 & ISO /TR 20461						UCSL/SOP/01-MPT					
CALIBRATION RESULTS @27°C:-											
1. Lower Volume 100 µl											
1	2	3	4	5	6	7	8	9	10	Mean Volume V	
101.15	100.17	100.58	100.63	99.96	101.49	101.57	101.63	99.67	100.20	100.71	
Systematic Error es µl :			-0.71			Error Limits (± 8 µl)		Random Error in sr µl :		0.72	
Systematic Error es in % :			-0.71			(± 8.0%)		Random Error in Cv in % :		0.71	
2. Middle Volume 500 µl											
1	2	3	4	5	6	7	8	9	10	Mean Volume V	
504.05	502.84	502.23	502.69	502.78	504.42	502.71	502.70	504.91	504.59	503.39	
Systematic Error es µl :			-3.39			Error Limits (± 8 µl)		Random Error in sr µl :		0.98	
Systematic Error es in % :			-0.68			(±1.6%)		Random Error in Cv in % :		0.20	
3. Nominal Volume 1000 µl											
1	2	3	4	5	6	7	8	9	10	Mean Volume V	
1003.58	1004.66	1004.60	1005.18	1003.88	1003.30	1004.81	1004.87	1003.51	1005.05	1004.34	
Systematic Error es µl :			-4.34			Error Limits (± 8.0 µl)		Random Error in sr µl :		0.70	
Systematic Error es in % :			-0.43			(± 0.80 %)		Random Error in Cv in % :		0.07	
Decision Rule is Applied or Not										Yes	
										No	
										✓	
Conclusion /Remarks:											
The Reported Uncertainty from 100 to 1000 µl is 0.60 µl at Coverage Factor K=2, which corresponds to a confidence level at approximately 95 %											
1. The calibration results reported in this certificate are valid at the time of and the stated condition of measurement.											
2. This report is valid for Scientific & Industrial Purpose Only											
3. This report should not be reproduced except in full without our prior permission in writing.											
4. Calibration certificate without signature are not valid.											
5. All the measurements are traceable to SI units through unbroken chain of calibration from the competent laboratories as per ISO/IEC/17025 standard requirements.											
6. This Calibration Certificate relates only to the above DUC											
7. DUC : Device under calibration											
Calibrated by						Checked By					
						 <p>Issued / Approved By (APPU K MANI) Technical Director</p>					

--- End of Certificate ---

Page 1 of 1





UNIQUE CALIBRATION SOLUTIONS LLP

An ISO/IEC-17025:2017 / ACCREDITED BY NABL

Tel: 0484-4858563, 0484 - 7966984, 9971328563, 9061728563, 9778197603

E-mail: info@uniquecalibration.co.in, website: www.uniquecalibration.co.in



CALIBRATION CERTIFICATE											
ULR NO :		CC31022400001062F				CERTIFICATE No. :		UCSL/04-24/091_02			
Customer Name & Address:						Instrument receipt Date		16.04.2024			
M/s. MICRO HEALTH LABORATORIES MP TOWER, NEXT TO LAVANYA PLAZA, NEAR THALUK HOSPITAL, THAMARASSERY, KERALA - 673 573						SRF No.		091			
						Calibration Date		17.04.2024			
						Next calibration date (As per customer request)		17.04.2025			
						Certificate Issue date		23.04.2024			
Identification on DUC (Device Under Calibration)											
Equipment Name		MICROPIPETTE				Make		MICROLIT			
Range		10 - 100 µl				Model		---			
Least count		1 µl				Instrument Condition		OK			
Serial No.		19111929				Calibration Performed at		Mass & Volume Lab			
Instrument ID.		MHL/TSY/21				Location		---			
Environmental Condition			Avg. Temperature (°C)			Avg. RH (%)			Avg. Atmospheric Pressure (hpa)		
			23.7			56			1009		
Equipment & Master Used For calibration											
Instrument Used		Calibrated By		ID No:		Certificate No		Valid Upto		Sr No.	
Weighing Balance		EIE		---		EIE/M/231046		30.07.2024		T1004465	
Method Used :						Gravimetric method					
Calibration Reference Standard						Discipline & Category : Mechanical - Volume					
ISO-8655 6 & ISO /TR 20461						Calibration Procedure					
						UCSL/SOP/01-MPT					
CALIBRATION RESULTS @27°C :-											
1. Nominal Volume 10 µl											
1	2	3	4	5	6	7	8	9	10	Mean Volume V	
9.92	9.80	9.82	9.52	9.92	9.91	9.91	9.70	9.85	9.88	9.82	
Systematic Error es µl :				0.18		Error Limits (± 0.8 µl)		Random Error in sr µl :		0.13	
Systematic Error es in % :				1.77		(± 8.0 %)		Random Error in Cv in % :		1.28	
Error Limits (± 0.3 µl)		Error Limits (± 0.6 %)									
2. Nominal Volume 50 µl											
1	2	3	4	5	6	7	8	9	10	Mean Volume V	
49.00	49.37	49.40	49.04	49.44	49.45	49.43	49.41	49.38	49.35	49.33	
Systematic Error es µl :				0.67		Error Limits (± 0.8 µl)		Random Error in sr µl :		0.17	
Systematic Error es in % :				1.34		(± 1.6 %)		Random Error in Cv in % :		0.34	
Error Limits (± 0.3 µl)		Error Limits (± 0.6 %)									
3. Nominal Volume 100 µl											
1	2	3	4	5	6	7	8	9	10	Mean Volume V	
100.52	100.44	100.34	100.50	100.18	100.58	100.58	100.23	100.23	100.29	100.39	
Systematic Error es µl :				-0.39		Error Limits (± 0.8 µl)		Random Error in sr µl :		0.15	
Systematic Error es in % :				-0.39		(± 0.8 %)		Random Error in Cv in % :		0.15	
Error Limits (± 0.3 µl)		Error Limits (± 0.3 %)									
Decision Rule is Applied or Not										Yes	
										No	
Conclusion /Remarks:											
The Reported Uncertainty upto 10 µl is 0.06 µl and from 10 µl to 100 µl is 0.12µl at Coverage Factor K=2, which corresponds to a confidence level at approximately 95 %											
The observed values and the measurement uncertainty are found to be within the maximum permissible errors as per ISO 8655, Hence the instrument is pass.											
1. The calibration results reported in this certificate are valid at the time of and the stated condition of measurement.											
2. This report is valid for Scientific & Industrial Purpose Only											
3. This report should not be reproduced except in full without our prior permission in writing.											
4. Calibration certificate without signature are not valid.											
5. This Calibration Certificate relates only to the above DUC											
6. All the measurements are traceable to SI units through unbroken chain of calibration from the competent laboratories as per ISO/IEC/17025 standard requirements.											
7. DUC : Device under calibration											
Calibrated by		Checked by				Issued / Approved By (APPU K MANI) Technical Director					





UNIQUE
CALIBRATION & MACHINES

UNIQUE CALIBRATION SOLUTIONS LLP

An ISO/IEC-17025:2017 / ACCREDITED BY NABL

Tel: 0484-4858563, 0484 - 7966984, 9971328563, 9061728563, 9778197603

E-mail: info@uniquecalibration.co.in, website: www.uniquecalibration.co.in



CC-3102

CALIBRATION CERTIFICATE											
ULR NO :		CC310224000001063F				CERTIFICATE No. :		UCSL/04-24/091_03			
Customer Name & Address:						Instrument receipt Date		16.04.2024			
M/s. MICRO HEALTH LABORATORIES						SRF No.		091			
MP TOWER, NEXT TO LAVANYA PLAZA, NEAR THALUK HOSPITAL, THAMARASSERY, KERALA - 673 573						Calibration Date		17.04.2024			
						Next calibration date (As per customer request)		17.04.2025			
						Certificate Issue date		23.04.2024			
Identification on DUC (Device Under Calibration)											
Instrument Name		MICROPIPETTE				Make		THERMO SCIENTIFIC			
Range		20 - 200 µl				Model		FINN PIPETTE			
Least count		1 µl				Instrument Condition		OK			
Serial No.		MW19904				Calibration Performed at		Mass & Volume Lab			
Instrument ID		MHL/TSY/14				Location		---			
Environmental Condition			Avg. Temperature (°C)			Avg. RH (%)			Avg. Atmospheric Pressure (hpa)		
			23.1			55			1010		
Equipment & Master Used For calibration											
Instrument Used		Calibrated By		ID No:		Certificate No		Valid Upto		Sr No.	
Weighing Balance		EIE		---		EIE/M/231046		30.07.2024		T1004465	
Weighing Balance		UCSL		UCSL-WB 01		UCSL/10-23/311_01		14.10.2024		0037905909	
										CC-2222	
										CC-3102	
Method Used :						Gravimetric method			Discipline & Category : Mechanical - Volume		
Calibration Reference Standard						Calibration Procedure					
ISO-8655-6 & ISO /TR 20461						UCSL/SOP/01-MPT					
CALIBRATION RESULTS :-											
1. Lower Volume		20 µl									
1	2	3	4	5	6	7	8	9	10	Mean Volume V	
20.17	20.17	20.14	20.17	20.15	20.18	20.18	20.18	20.20	20.15	20.17	
Systematic Error es µl :			-0.17			Error Limits (± 1.6 µl)			Random Error in sr µl :		
Systematic Error es in % :			-0.85			(± 8.0 %)			0.02		
									0.08 (± 3.0 %)		
2. Middle Volume		100 µl									
1	2	3	4	5	6	7	8	9	10	Mean Volume V	
101.27	101.38	100.45	101.38	101.37	101.35	101.38	101.42	101.24	101.42	101.26	
Systematic Error es µl :			-1.26			Error Limits (± 1.6 µl)			Random Error in sr µl :		
Systematic Error es in % :			-1.26			(± 1.6 %)			0.29		
									0.29 (± 0.6 %)		
3. Nominal Volume		200 µl									
1	2	3	4	5	6	7	8	9	10	Mean Volume V	
201.17	201.08	201.13	200.82	201.34	201.24	201.54	201.23	201.18	200.88	201.16	
Systematic Error es µl :			-1.16			Error Limits (± 1.6 µl)			Random Error in sr µl :		
Systematic Error es in % :			-0.58			(± 0.8 %)			0.21		
									0.10 (± 0.3 %)		
Decision Rule is Applied or Not										Yes	
										No	
Conclusion /Remarks:											
The Reported Uncertainty from 20 µl to 100 µl is 0.12µl and at 200 µl is 0.60 µl at , Coverage Factor K=2, which corresponds to a confidence level at approximately 95 %											
1. The calibration results reported in this certificate are valid at the time of and the stated condition of measurement.											
2. This report is valid for Scientific & Industrial Purpose Only											
3. This report should not be reproduced except in full without our prior permission in writing.											
4. Calibration certificate without signature are not valid.											
5. This Calibration Certificate relates only to the above DUC											
6. DUC : Device under calibration											
7. All the measurements are traceable to SI units through unbroken chain of calibration from the competent laboratories as per ISO/IEC 17025 standard requirements.											
Calibrated by		Checked By				Issued / Approved By (APPU K MANI) Technical Director					

--- End of Certificate ---

Page 1 of 1

5/D3, EK Appartments, Model engineering college road Thrikkakara, Edappally, Cochin, Kerala-682021

SALES ★ SPARES ★ SERVICE ★ TESTING ★ CALIBRATION ★ MODERNIZATION






CALIBRATION CERTIFICATE												
ULR NO :		CC310224000001064F				CERTIFICATE No. :		UCSL/04-24/091_04				
Customer Name & Address:						Instrument receipt Date		16.04.2024				
M/s. MICRO HEALTH LABORATORIES MP TOWER, NEXT TO LAVANYA PLAZA, NEAR THALUK HOSPITAL, THAMARASSERY, KERALA - 673 573						SRF No.		091				
						Calibration Date		17.04.2024				
						Next calibration date (As per customer request)		17.04.2025				
						Certificate Issue date		23.04.2024				
Identification on DUC (Device Under Calibration)												
Instrument Name		MICROPIPETTE				Make		MISPA				
Range		20 - 200 µl				Model		---				
Least count		0.2 µl				Instrument Condition		OK				
Serial No.		NW0085J				Calibration Performed at		Mass & Volume Lab				
Instrument ID		MHL/TSY/22				Location		---				
Environmental Condition			Avg. Temperature (°C)			Avg. RH (%)			Avg. Atmospheric Pressure (hpa)			
			24.2			54			1009			
Equipment & Master Used For calibration												
Instrument Used		Calibrated By		ID No:		Certificate No		Valid Upto		Sr No.		Accreditation No
Weighing Balance		EIE		---		EIE/M/231046		30.07.2024		T1004465		CC-2222
Weighing Balance		UCSL		UCSL-WB-01		UCSL/10-23/311_01		14.10.2024		0037905909		CC-3102
Method Used :						Gravimetric method			Discipline & Category :			Mechanical - Volume
Calibration Reference Standard						Calibration Procedure						
ISO-8655-6 & ISO /TR 20461						UCSL/SOP/01-MPT						
CALIBRATION RESULTS :-												
1. Lower Volume		20 µl										
1	2	3	4	5	6	7	8	9	10	Mean Volume V		
20.09	20.09	20.07	20.08	20.07	20.10	20.10	20.10	20.11	20.07	20.09		
Systematic Error es µl :			-0.09			Error Limits (± 1.6 µl)			Random Error in sr µl :			0.02
Systematic Error es in % :			-0.43			(± 8.0 %)			Random Error in Cv in % :			0.08
Error Limits (± 0.6 µl)						Error Limits (± 3.0 %)						
2. Middle Volume		100 µl										
1	2	3	4	5	6	7	8	9	10	Mean Volume V		
100.13	100.10	100.10	100.13	100.10	100.08	100.11	100.14	100.11	100.14	100.11		
Systematic Error es µl :			-0.11			Error Limits (± 1.6 µl)			Random Error in sr µl :			0.02
Systematic Error es in % :			-0.11			(± 1.6 %)			Random Error in Cv in % :			0.02
Error Limits (± 0.6 µl)						Error Limits (± 0.6 %)						
3. Nominal Volume		200 µl										
1	2	3	4	5	6	7	8	9	10	Mean Volume V		
200.86	200.91	200.60	200.65	200.81	200.71	201.01	201.06	200.88	200.71	200.82		
Systematic Error es µl :			-0.82			Error Limits (± 1.6 µl)			Random Error in sr µl :			0.15
Systematic Error es in % :			-0.41			(± 0.8 %)			Random Error in Cv in % :			0.08
Error Limits (± 0.6 µl)						Error Limits (± 0.3 %)						
Decision Rule is Applied or Not										Yes		No
												✓
Conclusion /Remarks:												
The Reported Uncertainty <u>from 20 µl to 100 µl is 0.12µl and at 200 µl is 0.60 µl</u> at ,Coverage Factor K=2,which corresponds to a confidence level at approximately 95 %												
1. The calibration results reported in this certificate are valid at the time of and the stated condition of measurement.												
2. This report is valid for Scientific & Industrial Purpose Only												
3. This report should not be reproduced except in full without our prior permission in writing.												
4. Calibration certificate without signature are not valid.												
5. This Calibration Certificate relates only to the above DUC												
6. DUC : Device under calibration												
7. All the measurements are traceable to SI units through unbroken chain of calibration from the competent laboratories as per ISO/IEC 17025 standard requirements.												
Calibrated by		Checked By				Issued / Approved By (APPU K MANI) Technical Director						

--- End of Certificate ---

Page 1 of 1



CALIBRATION CERTIFICATE											
ULR NO :		CC31022400001065F				CERTIFICATE No. :		UCSL/04-24/091_05			
Customer Name & Address:					Instrument receipt Date			16.04.2024			
M/s. MICRO HEALTH LABORATORIES					SRF No.			091			
MP TOWER, NEXT TO LAVANYA PLAZA, NEAR THALUK HOSPITAL, THAMARASSERY, KERALA - 673 573					Calibration Date			17.04.2024			
					Next calibration date (As per customer request)			17.04.2025			
					Certificate Issue date			23.04.2024			
Identification on DUC (Device Under Calibration)											
Instrument Name		MICROPIPETTE			Make		THERMO SCIENTIFIC				
Range		2 - 20 µl			Model		FINN PIPETTE F3				
Least count		0.02 µl			Instrument Condition		OK				
Serial No.		KW03951			Calibration Performed at		Mass & Volume Lab				
ID No.		MHL/TSY/14			Location		---				
Environmental Condition		Avg. Temperature (°C)			Avg. RH (%)		Avg. Atmospheric Pressure (hpa)				
		24.1			55		1010				
Equipment & Master Used For calibration											
Instrument Used		Calibrated By		ID No:		Certificate No		Valid Upto		Sr No.	Accreditation No
Weighing Balance		EIE		---		EIE/M/231046		30.07.2024		T1004465	CC-2222
Method Used :					Gravimetric method			Discipline & Category : Mechanical - Volume			
Calibration Reference Standard					Calibration Procedure						
ISO-8655-6 & ISO /TR 20461					UCSL/SOP/01-MPT						
CALIBRATION RESULTS @ 27 °C :-											
1. Lower Volume 2 µl											
1	2	3	4	5	6	7	8	9	10	Mean Volume V	
2.009	2.024	2.001	2.009	2.032	2.017	2.015	2.005	2.050	2.016	2.018	
Systematic Error es µl :				-0.02		Error Limits (± 0.20 µl)		Random Error in sr µl :		0.01	
Systematic Error es in % :				-0.88		(±10.0%)		Random Error in Cv in % :		0.72	
2. Middle Volume 10 µl											
1	2	3	4	5	6	7	8	9	10	Mean Volume V	
10.043	10.028	10.016	10.042	10.069	10.051	10.045	10.054	10.019	10.035	10.041	
Systematic Error es µl :				-0.04		Error Limits (± 0.20 µl)		Random Error in sr µl :		0.02	
Systematic Error es in % :				-0.41		(± 2.0 %)		Random Error in Cv in % :		0.16	
3. Nominal Volume 20 µl											
1	2	3	4	5	6	7	8	9	10	Mean Volume V	
20.187	20.168	20.186	20.106	20.171	20.176	20.101	20.187	20.178	20.194	20.165	
Systematic Error es µl :				-0.17		Error Limits (± 0.20 µl)		Random Error in sr µl :		0.03	
Systematic Error es in % :				-0.83		(± 1.0 %)		Random Error in Cv in % :		0.17	
Decision Rule is Applied or Not										Yes	No
											✓
Conclusion /Remarks:											
The Reported Uncertainty from 0.5 to 10 µl is 0.06 µl and 20 µl is 0.12 µl at ,Coverage Factor K=2,which corresponds to a confidence level at approximately 95 %											
1. The calibration results reported in this certificate are valid at the time of and the stated condition of measurement.											
2. This report is valid for Scientific & Industrial Purpose Only											
3. This report should not be reproduced except in full without our prior permission in writing.											
4. Calibration certificate without signature are not valid.											
5. This Calibration Certificate relates only to the above DUC											
6. All the measurements are traceable to SI units through unbroken chain of calibration from the competent laboratories as per ISO/IEC 17025 standard requirements.											
7. DUC : Device under calibration											
Calibrated by		Checked by				Issued / Approved By					
											
						(APPU K MANI) (Technical Director)					

--- End of Certificate ---

Page 1 of 1





UNIQUE
CALIBRATION & MACHINES

UNIQUE CALIBRATION SOLUTIONS LLP


An ISO/IEC-17025:2017 / ACCREDITED BY NABL

Tel: 0484-4858563, 0484 - 7966984, 9971328563, 9061728563, 9778197603

E-mail: info@uniquecalibration.co.in, website: www.uniquecalibration.co.in




CC-3102

CALIBRATION CERTIFICATE				
UNIQUE LAB REPORT NO (ULR NO) :		CC310224000001066F		
SRF NO. :	091	CERTIFICATE NO	UCSL/04-24/091_06	
Customer Details :	M/s. MICRO HEALTH LABORATORIES MP TOWER, NEXT TO LAVANYA PLAZA, NEAR THALUK HOSPITAL, THAMARASSERY, KERALA - 673 573			
Device Under Calibration	CENTRIFUGE (RPM)		Equipment & Master Used For Calibration	
Make / Model No.	ELECTRA BL		01. DIGITAL CONTACT & NON CONTACT TACHOMETER	
Range	0 to 6000 rpm		Make : LUTRON	
Least Count	1 rpm		Serial No. : Q693180	
Instrument Id. No.	MHL-TSY-PQP-08		Certificate No. : C-230620-8-21	
Instrument Sr. No.	EPBL20160046		Next Due Date : 22.06.2024	
Location	LAB		Calibrated By : Godrej & Boyce	
Visual Inspection	OK		Accreditation No : CC-2559	
Material Receipt Date	---			
Calibration performed At	ON SITE			
Date of Calibration	23.04.2024			
Next calibration due (As per customer request)	23.04.2025			
Calibration Certificate Issue Date	23.04.2024			
Discipline	Mechanical			
Method Used	Comparison			
ENVIRONMENTAL CONDITION				
Temperature	27.1°C			
Relative Humidity	57%RH			
Condition Of Equipment :	Un-Loaded			
OBSERVATION RESULTS				
SL. NO.	Nominal Value on DUC (rpm)	Observed Mean Value on MASTER (rpm)	% Error of Reading (±)	± Expanded Uncertainty At Approx.95% Confidence Level (k=2)
1.	500	500.5	0.10	0.60 %
2.	1000	1006	0.56	0.60 %
3.	3000	3006	0.21	0.60 %
4.	4000	4005	0.11	0.60 %
5.	6000	6005	0.08	0.60 %
Calibration Procedure Based On		: UCSL/SOP/RPM		
Decision Rule is Applied or Not			Yes	No
				✓
NOTE :-				
1. The calibration results reported in this certificate are valid at the time of and the stated condition of measurement.				
2. This report should not be reproduced except in full without our prior permission in writing.				
3. Calibration certificate without signature are not valid.				
4. All the measurements are traceable to SI units through unbroken chain of calibration from the competent laboratories as per ISO/IEC/17025 standard requirements.				
5. DUC : Device under calibration				
6. This calibration certificate relates only to the above DUC.				
Calibrated By	Checked By		 Issued / Approved By (APPU K MANI) Technical Director	

--- End of Certificate ---
Page 1 of 1



CALIBRATION CERTIFICATE			
SRF NO :	091	Certificate No :	UCSL/04-24/091_07
Customer Details :	M/s. MICRO HEALTH LABORATORIES MP TOWER, NEXT TO LAVANYA PLAZA, NEAR THALUK HOSPITAL, THAMARASSERY, KERALA - 673 573		
Device Under Calibration	REFRIGERATOR	Equipment & Master Used For Calibration	
Make / Model	SAMSUNG	01. 4 WIRE RTD SENSOR WITH INDICATOR	
Range	-20 to 8 °C	Make	: TEMPESENS (Sensor) / YUDIAN (Indicator)
Least Count	---	Sr. No.	: 2113028448 / 20L737937
Instrument Id. No.	MHRL-BM-REC-82	Certificate No.	: TL/024/60.1.1
Instrument Sr. No.	ROC140BH100B3ZZ	Next Due Date	: 15.01.2025
Visual Inspection	OK	Calibrated By	: TEMPESENS
Location	LAB		
Material Receipt Date			
Date of Calibration		---	
Next Calibration Due (As per Customer Request)		23.04.2024	
Certificate Issue Date		23.04.2025	
Calibration performed At		23.04.2024	
Discipline		ON SITE	
Method used		THERMAL COMPARISON	
ENVIRONMENTAL CONDITION			
Temperature	25.3°C		
Relative Humidity	56%RH		
OBSERVATION RESULTS			
SL.NO	SET VALUE (°C)	MASTER READING (°C)	
01	Maintaining Temperature	8.729	
Calibration Procedure Based On : UCSL/SOP/01-T (ITS-90)			
NOTE :-			
1. The calibration results reported in this certificate are valid at the time of and the stated condition of measurement.			
2. This report should not be reproduced except in full without our prior permission in writing.			
3. Calibration certificate without signature are not valid.			
4. All the measurements are traceable to SI units through unbroken chain of calibration from the competent laboratories as per ISO/IEC/17025 standard requirements.			
5. DUC : Device under calibration.			
6. Reported readings of MASTER are average of 5 measurements.			
7. Measurement Uncertainty reported is at approximately 95% confidence level with K=2.			
8. This Calibration Certificate relates only to the above DUC.			
Calibrated By	Checked By		 Issued / Approved By (APPU K MANI) Technical Director
