



UNIQUE
CALIBRATION & MACHINES

UNIQUE CALIBRATION SOLUTIONS LLP

An ISO/IEC-17025:2017 / ACCREDITED BY NABL

Tel: 0484-4858563, 0484 - 7966984, 9971328563, 9061728563, 9778197603

E-mail: info@uniquecalibration.co.in, website: www.uniquecalibration.co.in



CC-3102

CALIBRATION CERTIFICATE													
ULR NO :		CC310224000001063F				CERTIFICATE No. :		UCSL/04-24/091_03					
Customer Name & Address:						Instrument receipt Date		16.04.2024					
M/s. MICRO HEALTH LABORATORIES						SRF No.		091					
MP TOWER, NEXT TO LAVANYA PLAZA, NEAR THALUK HOSPITAL, THAMARASSERY, KERALA - 673 573						Calibration Date		17.04.2024					
						Next calibration date (As per customer request)		17.04.2025					
						Certificate Issue date		23.04.2024					
Identification on DUC (Device Under Calibration)													
Instrument Name		MICROPIPETTE				Make		THERMO SCIENTIFIC					
Range		20 - 200 µl				Model		FINN PIPETTE					
Least count		1 µl				Instrument Condition		OK					
Serial No.		MW19904				Calibration Performed at		Mass & Volume Lab					
Instrument ID		MHL/TSY/14				Location		---					
Environmental Condition			Avg. Temperature (°C)			Avg. RH (%)			Avg. Atmospheric Pressure (hpa)				
			23.1			55			1010				
Equipment & Master Used For calibration													
Instrument Used		Calibrated By		ID No:		Certificate No		Valid Upto		Sr No.			
Weighing Balance		EIE		---		EIE/M/231046		30.07.2024		T1004465			
Weighing Balance		UCSL		UCSL-WB 01		UCSL/10-23/311_01		14.10.2024		0037905909			
										CC-2222			
										CC-3102			
Method Used :						Gravimetric method			Discipline & Category : Mechanical - Volume				
Calibration Reference Standard						Calibration Procedure							
ISO-8655-6 & ISO /TR 20461						UCSL/SOP/01-MPT							
CALIBRATION RESULTS :-													
1. Lower Volume		20 µl											
1	2	3	4	5	6	7	8	9	10	Mean Volume V			
20.17	20.17	20.14	20.17	20.15	20.18	20.18	20.18	20.20	20.15	20.17			
Systematic Error es µl :				-0.17				Error Limits (± 1.6 µl)				Random Error in sr µl : 0.02	
Systematic Error es in % :				-0.85				Error Limits (± 8.0 %)				Random Error in Cv in % : 0.08	
2. Middle Volume		100 µl											
1	2	3	4	5	6	7	8	9	10	Mean Volume V			
101.27	101.38	100.45	101.38	101.37	101.35	101.38	101.42	101.24	101.42	101.26			
Systematic Error es µl :				-1.26				Error Limits (± 1.6 µl)				Random Error in sr µl : 0.29	
Systematic Error es in % :				-1.26				Error Limits (± 1.6 %)				Random Error in Cv in % : 0.29	
3. Nominal Volume		200 µl											
1	2	3	4	5	6	7	8	9	10	Mean Volume V			
201.17	201.08	201.13	200.82	201.34	201.24	201.54	201.23	201.18	200.88	201.16			
Systematic Error es µl :				-1.16				Error Limits (± 1.6 µl)				Random Error in sr µl : 0.21	
Systematic Error es in % :				-0.58				Error Limits (± 0.8 %)				Random Error in Cv in % : 0.10	
Decision Rule is Applied or Not										Yes		No	
												✓	
Conclusion /Remarks:													
The Reported Uncertainty from 20 µl to 100 µl is 0.12µl and at 200 µl is 0.60 µl at , Coverage Factor K=2, which corresponds to a confidence level at approximately 95 %													
1. The calibration results reported in this certificate are valid at the time of and the stated condition of measurement.													
2. This report is valid for Scientific & Industrial Purpose Only													
3. This report should not be reproduced except in full without our prior permission in writing.													
4. Calibration certificate without signature are not valid.													
5. This Calibration Certificate relates only to the above DUC													
6. DUC : Device under calibration													
7. All the measurements are traceable to SI units through unbroken chain of calibration from the competent laboratories as per ISO/IEC 17025 standard requirements.													
Calibrated by		Checked By				Issued / Approved By (APPU K MANI) Technical Director							

--- End of Certificate ---

Page 1 of 1

5/D3, EK Appartments, Model engineering college road Thrikkakara, Edappally, Cochin, Kerala-682021

SALES ★ SPARES ★ SERVICE ★ TESTING ★ CALIBRATION ★ MODERNIZATION

