

CALIBRATION CERTIFICATE														
ULR NO :		CC310224000001064F				CERTIFICATE No. :		UCSL/04-24/091_04						
Customer Name & Address:						Instrument receipt Date		16.04.2024						
M/s. MICRO HEALTH LABORATORIES MP TOWER, NEXT TO LAVANYA PLAZA, NEAR THALUK HOSPITAL, THAMARASSERY, KERALA - 673 573						SRF No.		091						
						Calibration Date		17.04.2024						
						Next calibration date (As per customer request)		17.04.2025						
						Certificate Issue date		23.04.2024						
Identification on DUC (Device Under Calibration)														
Instrument Name		MICROPIPETTE				Make		MISPA						
Range		20 - 200 µl				Model		---						
Least count		0.2 µl				Instrument Condition		OK						
Serial No.		NW0085J				Calibration Performed at		Mass & Volume Lab						
Instrument ID		MHL/TSY/22				Location		---						
Environmental Condition			Avg. Temperature (°C)			Avg. RH (%)			Avg. Atmospheric Pressure (hpa)					
			24.2			54			1009					
Equipment & Master Used For calibration														
Instrument Used		Calibrated By		ID No:		Certificate No		Valid Upto		Sr No.		Accreditation No		
Weighing Balance		EIE		---		EIE/M/231046		30.07.2024		T1004465		CC-2222		
Weighing Balance		UCSL		UCSL-WB-01		UCSL/10-23/311_01		14.10.2024		0037905909		CC-3102		
Method Used :						Gravimetric method			Discipline & Category :			Mechanical - Volume		
Calibration Reference Standard						Calibration Procedure								
ISO-8655-6 & ISO /TR 20461						UCSL/SOP/01-MPT								
CALIBRATION RESULTS :-														
1. Lower Volume		20 µl												
1	2	3	4	5	6	7	8	9	10	Mean Volume V				
20.09	20.09	20.07	20.08	20.07	20.10	20.10	20.10	20.11	20.07	20.09				
Systematic Error es µl :			-0.09			Error Limits (± 1.6 µl)			Random Error in sr µl :			0.02		
Systematic Error es in % :			-0.43			(± 8.0 %)			Random Error in Cv in % :			0.08		
Error Limits (± 0.6 µl)						Error Limits (± 3.0 %)								
2. Middle Volume		100 µl												
1	2	3	4	5	6	7	8	9	10	Mean Volume V				
100.13	100.10	100.10	100.13	100.10	100.08	100.11	100.14	100.11	100.14	100.11				
Systematic Error es µl :			-0.11			Error Limits (± 1.6 µl)			Random Error in sr µl :			0.02		
Systematic Error es in % :			-0.11			(± 1.6 %)			Random Error in Cv in % :			0.02		
Error Limits (± 0.6 µl)						Error Limits (± 0.6 %)								
3. Nominal Volume		200 µl												
1	2	3	4	5	6	7	8	9	10	Mean Volume V				
200.86	200.91	200.60	200.65	200.81	200.71	201.01	201.06	200.88	200.71	200.82				
Systematic Error es µl :			-0.82			Error Limits (± 1.6 µl)			Random Error in sr µl :			0.15		
Systematic Error es in % :			-0.41			(± 0.8 %)			Random Error in Cv in % :			0.08		
Error Limits (± 0.6 µl)						Error Limits (± 0.3 %)								
Decision Rule is Applied or Not										Yes		No		
												✓		
Conclusion /Remarks:														
The Reported Uncertainty <u>from 20 µl to 100 µl is 0.12µl and at 200 µl is 0.60 µl</u> at ,Coverage Factor K=2,which corresponds to a confidence level at approximately 95 %														
1. The calibration results reported in this certificate are valid at the time of and the stated condition of measurement.														
2. This report is valid for Scientific & Industrial Purpose Only														
3. This report should not be reproduced except in full without our prior permission in writing.														
4. Calibration certificate without signature are not valid.														
5. This Calibration Certificate relates only to the above DUC														
6. DUC : Device under calibration														
7. All the measurements are traceable to SI units through unbroken chain of calibration from the competent laboratories as per ISO/IEC 17025 standard requirements.														
Calibrated by		Checked By				Issued / Approved By (APPU K MANI) Technical Director								

--- End of Certificate ---

Page 1 of 1

