



INNOWAVE DIAGNOSTICS SERVICES

Service Report No.: 1226	Date: 20/04/2024	
Customer Name & Address: Unipath Specialty Gujarat LLP.	STATUS: <input type="checkbox"/> Warranty <input type="checkbox"/> AMC <input type="checkbox"/> CMC <input type="checkbox"/> Ad-hoc	CALL TYPE: <input type="checkbox"/> Installation <input type="checkbox"/> P.M. visit <input type="checkbox"/> Application <input type="checkbox"/> Breakdown

Instrument Model: ST000 A00A (Alakhtika)	Serial No.: ST000-2614
Received Date:	Attended Date: 20/04/2024

Problem Description:
- Preventive maintenance

Action Summary:
- The preventive maintenance has been provided as per the standard service
- All the lubings, electrodes with seals, probe, reference housing, bubble detector, probe seal etc. have been cleaned.

Service Engineer Remarks: Machine works properly.

Customer Remarks:

Feedback:	<input type="checkbox"/> Excellent	<input type="checkbox"/> Very Good	<input type="checkbox"/> Good	<input type="checkbox"/> Fair	<input type="checkbox"/> Poor
-----------	------------------------------------	------------------------------------	-------------------------------	-------------------------------	-------------------------------

Service Engineer Name & Signature Vijay Khat Khat	Customer Name & Signature 20/4/24
---	--