

Analyte	Instrument	Result Value	Standard Unit
No Outlier			
Legend <input checked="" type="radio"/> Acceptable			
X Unacceptable			
# Not Evaluated			
○ Delayed Result Entry			

Problem Classification: No outliers

Corrective Action:

Reviewed by: _____ Dated: _____

W. G. G...

20/2/24

Instrument : Manual

Analyte	Standard Unit	Result Value	Accepted Value
@ pH	-	6	5.5-8.5
@ Specific Gravity	-	1.025	1.025-1.030
@ Glucose (urine)	-	4+	-2+3+4
@ Protein (Urine)	-	2+	+1+2+3
@ Ketone (urine)	-	Negative	Neg
@ Blood (urine)	-	2+	+1-2+3
@ Bilirubin (urine)	-	Negative	Neg
@ Urobilinogen (urine)	-	Normal	Nor
@ Nitrite (urine)	-	Present	Pre

Instrument : -

Analyte	Standard Unit	Result Value	Accepted Value
@ Urine Pregnancy Test	-	Positive	Pos
Legend <input checked="" type="radio"/> Acceptable		Total Parameters 10	
<input type="radio"/> Acceptable		Not Evaluated Parameters 0	
X Unacceptable		Evaluated Parameters 10	
# Not Evaluated		Outlier Parameters (X) 0	
○ Delayed Result Entry		EQAS Score Clinical Path - Urine Routine Chemistry 100.00 %	

DK Nigam

Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416
Commercial Building - 1A
Kohinor Mall, Kirti Road, Kurla (W)
Mumbai - 400070