

Date: 21/02/2024

PM & CALIBRATION CERTIFICATE

This is to certify that the Hematology Analyzer of **HORIBA Medical:YUMIZEN H 500**, bearing serial no:**208YODH04845** installed at **SPARSHA DIAGNOSTICS CENTER, MANGLORE** was calibrated with Calibrator on **21/02/2024**.

Calibrator Used: **ABX MINOCAL**

Lot No.: **CX 492**

Expiry Date: **05/04/2024**

The reports of Blank Cycle, Repeatability and Calibration And QC Values were all found in acceptable range.

Next PM & Calibration cycle is due on 21/02/2025.

SUDHEESH K V
(Engineer Lab Tech Support)



For Horiba India Pvt.Ltd.

Checked By:

Designation:

*Lab Director &
Consultant Pathologist*