



CALIBRATION CERTIFICATE

Certificate No.	MES/23-24/SDUBM/CEN/D01	Date of Issue : 22/03/2024
		Calibration Done On : 13/03/2024
		Calibration Due On : 12/03/2025

Customer Name	SPARSHA DIAGNOSTICS, UPPER BENDOOR, MANGALORE.
---------------	---

DETAILS OF DEVICE UNDER CALIBRATION

Equipment Name	Make / Model	Serial No.	Asset ID	Location
Centrifuge	REMI	-	SD/CL/CEN/01	Central Laboratory

DETAILS OF MASTER EQUIPMENT USED

SI No.	Master Equipment	Make / Model	Serial No.	Reference ID	Cal. Due
1	Digital Tachometer	Metravi / NCTM-2000	210532216	TVCSPL23/09/313-02	15/09/2024
2	Digital Thermo Hygrometer	Equinox & EQ - 308 CTH	20191602	TVCSPL 23/06/372 - 01	13/06/2024

The instruments used for calibration are under valid calibration and are traceable to National Standards.

Environmental Conditions	Temperature	27°C	Relative Humidity	55 %
---------------------------------	-------------	------	-------------------	------

PERFORMANCE DATA

Set Reading	Display Reading	Measured Reading	Tolerance
1200 rpm	1200 rpm	1200 rpm	± 5 rpm
1500 rpm	1500 rpm	1503 rpm	± 5 rpm
2000 rpm	2000 rpm	2003 rpm	± 5 rpm
2500 rpm	2500 rpm	2503 rpm	± 5 rpm
3000 rpm	3000 rpm	3000 rpm	± 5 rpm

Calibration Status	Pass	✓	Fail
--------------------	------	---	------

Calibrated By

Bisto T.P.

Bisto T.P (Service Engineer)

Approved By



Binoj Thomas (Quality Manager)