





Cert No. CC-2543



**Medical Engineering & Services**  
**Things Beyond Thoughts**

Abhcon Crown, X1/411/21, Ambadi Lane  
Kokkalai, Thrissur-21. Ph: 0487 2420367  
Email: mescalibration.tsr@gmail.com  
Visit us @www.mescalibration.com

**CALIBRATION CERTIFICATE**

Certificate Number	: MES24/03/058-03	ULR Number	: CC254324000000355F
Customer Details	: <b>SPARSHA DIAGNOSTICS, UPPER BENDOOOR, MANGALORE.</b>		
Date of Calibration	: <b>19/03/2024</b>	SRF Number	: <b>058</b>
Next Calibration Due	: <b>18/03/2025</b>	Date of Issue	: <b>19/03/2024</b>

**DETAILS OF DEVICE UNDER CALIBRATION(DUC)**

Nomenclature	: Micropipette	Serial No.	: YN21BAG0013028
Make / Model	: Mispa i 2	ID No.	: -
Equipment Condition	: Good	Resolution	: 1 µl
Range	: 20 - 200 µL	Location	: Central Laboratory

**DETAILS OF MASTER EQUIPMENT USED**

Nomenclature	: Electronic Precision Balance	Range	: 0 – 210 gm
Make	: Boeco Germany	Readability	: 0.01/0.1 mg
Model	: BBC-22	Serial No.	: 27104178
Certificate No.	: TVCSPL 23/10/1135-01	Calibration Due Date	: 02.10.2024

**ENVIRONMENTAL CONDITIONS**

Temperature(°C)	Relative Humidity (%)	Atmospheric Pressure (mbar)	Water Temperature(°C)
26.9	57	1007	26.7

**CALIBRATION RESULTS**

Discipline		Mechanical		Parameter		Volume	
Sl. No.	Set Volume	Standard Value	Systematic Error	Acceptable Systematic Error	Random Error	Acceptable Random Error	Expanded Uncertainty
	µl	µl	µl	± µl	µl	± µl	± µl
1	20	19.97	0.03	0.40	0.03	0.20	0.819
2	100	100.27	-0.27	1.40	0.04	0.60	0.821
3	200	200.46	-0.46	12.00	0.32	4.00	1.091

**Notes: -**

- The results in this certificate are only related to DUC submitted calibration and valid at the time of calibration under stated condition of measurements
- The measurement uncertainty is estimated at a confidence level of 95.45% with coverage factor k=2.
- Results in this certificate shall not be reproduced in partial or full without the written approval of the laboratory.
- Next calibration due date is mentioned as per customer requirement.
- Calibration certificate issued for scientific or industrial purpose only.
- Acceptable error limit is given as per ISO 8655-2:2002(E) Guidelines
- Calibration is carried out as per ISO 8655-6 Guidelines;MES work procedure: MES/WP/VOLUME.

Calibrated by

*Prabin KP*  
Prabin KP

(Calibration Engineer)



Authorized Signatory

*Binoj Thomas*

Binoj Thomas  
(Technical Manager)



Cert No. CC-2543



## Medical Engineering & Services

**Things Beyond Thoughts**

Abhcon Crown, X1/411/21, Ambadi Lane  
Kokkalai, Thrissur-21. Ph. 0487 2420367  
Email: mescalibration.tsr@gmail.com  
Visit us @www.mescalibration.com

### CALIBRATION CERTIFICATE

Certificate Number	: MES24/03/058-02	ULR Number	: CC254324000000354F
Customer Details	: <b>SPARSHA DIAGNOSTICS, UPPER BENDLOOR, MANGALORE.</b>		
Date of Calibration	: <b>19/03/2024</b>	SRF Number	: <b>058</b>
Next Calibration Due	: <b>18/03/2025</b>	Date of Issue	: <b>19/03/2024</b>

#### DETAILS OF DEVICE UNDER CALIBRATION(DUC)

Nomenclature	: Micropipette	Serial No.	: YE20BAS0077285
Make / Model	: Dragon Lab	ID No.	: -
Equipment Condition	: Good	Resolution	: 0 µl
Range	: 50 µL	Location	: Central Laboratory

#### DETAILS OF MASTER EQUIPMENT USED

Nomenclature	: Electronic Precision Balance	Range	: 0 – 210 gm
Make	: Boeco Germany	Readability	: 0.01/0.1 mg
Model	: BBC-22	Serial No.	: 27104178
Certificate No.	: TVCSPL 23/10/1135-01	Calibration Due Date	: 02.10.2024

#### ENVIRONMENTAL CONDITIONS

Temperature(°C)	Relative Humidity (%)	Atmospheric Pressure (mbar)	Water Temperature(°C)
26.7	58	1007	26.5

#### CALIBRATION RESULTS

Discipline		Mechanical		Parameter		Volume	
Sl. No.	Set Volume	Standard Value	Systematic Error	Acceptable Systematic Error	Random Error	Acceptable Random Error	Expanded Uncertainty
	µl	µl	µl	± µl	µl	± µl	± µl
1	50	50.04	-0.04	0.50	0.06	0.20	0.183

Notes: -

- The results in this certificate are only related to DUC submitted calibration and valid at the time of calibration under stated condition of measurements
- The measurement uncertainty is estimated at a confidence level of 95.45% with coverage factor k=2.
- Results in this certificate shall not be reproduced in partial or full without the written approval of the laboratory.
- Next calibration due date is mentioned as per customer requirement.
- Calibration certificate issued for scientific or industrial purpose only.
- Acceptable error limit is given as per ISO 8655-2:2002(E) Guidelines
- Calibration is carried out as per ISO 8655-6 Guidelines;MES work procedure: MES/WP/VOLUME.

Calibrated by

*Prabin KP*  
Prabin KP

(Calibration Engineer)



Authorized Signatory

*Binoj Thomas*  
Binoj Thomas  
(Technical Manager)



Cert No. CC-2543



# Medical Engineering & Services

**Things Beyond Thoughts**

Abhcon Crown, X1/411/21, Ambadi Lane  
Kokkalai, Thrissur-21. Ph: 0487 2420367  
Email: mescalibration.tsr@gmail.com  
Visit us @www.mescalibration.com

## CALIBRATION CERTIFICATE

Certificate Number	: MES24/03/053-03	ULR Number	: CC254324000000316F
Customer Details	: SPARSHA DIAGNOSTICS, UPPER BENDOOR, MANGALORE.		
Date of Calibration	: 14/03/2024	SRF Number	: 053
Next Calibration Due	: 13/03/2025	Date of Issue	: 14/03/2024

### DETAILS OF DEVICE UNDER CALIBRATION(DUC)

Nomenclature	: Micropipette	Serial No.	: YE202AM0110224
Make / Model	: Dragon Lab	ID No.	: -
Equipment Condition	: Good	Resolution	: 5 µl
Range	: 100 - 1000 µl	Location	: Central Laboratory

### DETAILS OF MASTER EQUIPMENT USED

Nomenclature	: Electronic Precision Balance	Range	: 0 – 210 gm
Make	: Boeco Germany	Readability	: 0.01/0.1 mg
Model	: BBC-22	Serial No.	: 27104178
Certificate No.	: TVCSPL 23/10/1135-01	Calibration Due Date	: 02.10.2024

### ENVIRONMENTAL CONDITIONS

Temperature(°C)	Relative Humidity (%)	Atmospheric Pressure (mbar)	Water Temperature(°C)
26.7	57	1007	26.5

### CALIBRATION RESULTS

Sl. No.	Discipline	Mechanical			Parameter		Volume	
		Set Volume	Standard Value	Systematic Error	Acceptable Systematic Error	Random Error	Acceptable Random Error	Expanded Uncertainty
		µl	µl	µl	± µl	µl	± µl	± µl
1		100	100.32	-0.32	1.40	0.02	0.60	4.083
2		500	502.05	-2.05	12.00	0.06	4.00	4.085
3		1000	1004.20	-4.20	12.00	0.12	4.00	4.093

#### Notes: -

- The results in this certificate are only related to DUC submitted calibration and valid at the time of calibration under stated condition of measurements
- The measurement uncertainty is estimated at a confidence level of 95.45% with coverage factor k=2.
- Results in this certificate shall not be reproduced in partial or full without the written approval of the laboratory.
- Next calibration due date is mentioned as per customer requirement.
- Calibration certificate issued for scientific or industrial purpose only.
- Acceptable error limit is given as per ISO 8655-2:2002(E) Guidelines
- Calibration is carried out as per ISO 8655-6 Guidelines;MES work procedure: MES/WP/VOLUME.

Calibrated by

  
Prabin KP

(Calibration Engineer)



Authorized Signatory

  
Binoj Thomas

(Technical Manager)

**CALIBRATION CERTIFICATE**

Certificate Number : MES24/03/053-04      ULR Number : CC254324000000317F  
Customer Details : **SPARSHA DIAGNOSTICS, UPPER BENDLOOR, MANGALORE.**  
Date of Calibration : **14/03/2024**      SRF Number : **053**  
Next Calibration Due : **13/03/2025**      Date of Issue : **14/03/2024**

**DETAILS OF DEVICE UNDER CALIBRATION(DUC)**

Nomenclature : Micropipette      Serial No. : YE175AB0105015  
Make / Model : Erba Mannheim      ID No. : -  
Equipment Condition : Good      Resolution : 1 µl  
Range : 10 - 100 µL      Location : Central Laboratory

**DETAILS OF MASTER EQUIPMENT USED**

Nomenclature : Electronic Precision Balance      Range : 0 – 210 gm  
Make : Boeco Germany      Readability : 0.01/0.1 mg  
Model : BBC-22      Serial No. : 27104178  
Certificate No. : TVCSPL 23/10/1135-01      Calibration Due Date : 02.10.2024

**ENVIRONMENTAL CONDITIONS**

Temperature(°C)	Relative Humidity (%)	Atmospheric Pressure (mbar)	Water Temperature(°C)
26.9	55	1007	26.7

**CALIBRATION RESULTS**

Sl. No.	Discipline : Mechanical			Parameter : Volume			
	Set Volume	Standard Value	Systematic Error	Acceptable Systematic Error	Random Error	Acceptable Random Error	Expanded Uncertainty
	µl	µl	µl	± µl	µl	± µl	± µl
1	10	9.96	0.04	0.12	0.02	0.08	0.817
2	50	50.07	-0.07	0.50	0.04	0.20	0.820
3	100	100.25	-0.25	1.40	0.05	0.60	0.823

**Notes: -**

- The results in this certificate are only related to DUC submitted calibration and valid at the time of calibration under stated condition of measurements
- The measurement uncertainty is estimated at a confidence level of 95.45% with coverage factor k=2.
- Results in this certificate shall not be reproduced in partial or full without the written approval of the laboratory.
- Next calibration due date is mentioned as per customer requirement.
- Calibration certificate issued for scientific or industrial purpose only.
- Acceptable error limit is given as per ISO 8655-2:2002(E) Guidelines
- Calibration is carried out as per ISO 8655-6 Guidelines;MES work procedure: MES/WP/VOLUME.


Calibrated by

  
Prabin KP

(Calibration Engineer)



Authorized Signatory

  
Binoj Thomas  
(Technical Manager)





Cert No. CC-2543



# Medical Engineering & Services

Things Beyond Thoughts

Abhcon Crown, X1/411/21, Ambadi Lane  
Kokkalai, Thrissur-21. Ph: 0487 2420367  
Email: mescalibration.tsr@gmail.com  
Visit us @www.mescalibration.com

## CALIBRATION CERTIFICATE

Certificate Number	: MES24/03/053-06	ULR Number	: CC254324000000319F
Customer Details	: SPARSHA DIAGNOSTICS, UPPER BENDOOOR, MANGALORE.		
Date of Calibration	: 14/03/2024	SRF Number	: 053
Next Calibration Due	: 13/03/2025	Date of Issue	: 14/03/2024

### DETAILS OF DEVICE UNDER CALIBRATION(DUC)

Nomenclature	: Micropipette	Serial No.	: PW13793
Make / Model	: Themo Scientific/ Finnpiette F3	ID No.	: -
Equipment Condition	: Good	Resolution	: 10 µl
Range	: 500 – 5000 µL	Location	: Central Laboratory

### DETAILS OF MASTER EQUIPMENT USED

Nomenclature	: Electronic Precision Balance	Range	: 0 – 210 gm
Make	: Boeco Germany	Readability	: 0.01/0.1 mg
Model	: BBC-22	Serial No.	: 27104178
Certificate No.	: TVCSPL 23/10/1135-01	Calibration Due Date	: 02.10.2024

### ENVIRONMENTAL CONDITIONS

Temperature(°C)	Relative Humidity (%)	Atmospheric Pressure (mbar)	Water Temperature(°C)
27.5	56	1007	27.3

### CALIBRATION RESULTS

Sl. No.	Discipline : Mechanical			Parameter :			Volume
	Set Volume	Standard Value	Systematic Error	Acceptable Systematic Error	Random Error	Acceptable Random Error	Expanded Uncertainty
	µl	µl	µl	± µl	µl	± µl	± µl
1	500	502.21	-2.21	12.00	0.01	4.00	8.165
2	2500	2511.53	-11.53	40.00	0.09	15.00	8.176
3	5000	5022.62	-22.62	40.00	0.33	15.00	8.226

#### Notes: -

- The results in this certificate are only related to DUC submitted calibration and valid at the time of calibration under stated condition of measurements
- The measurement uncertainty is estimated at a confidence level of 95.45% with coverage factor k=2.
- Results in this certificate shall not be reproduced in partial or full without the written approval of the laboratory.
- Next calibration due date is mentioned as per customer requirement.
- Calibration certificate issued for scientific or industrial purpose only.
- Acceptable error limit is given as per ISO 8655-2:2002(E) Guidelines
- Calibration is carried out as per ISO 8655-6 Guidelines;MES work procedure: MES/WP/VOLUME.

Calibrated by

*Prabin*  
Prabin KP

(Calibration Engineer)



Authorized Signatory

*Binoj Thomas*  
Binoj Thomas

(Technical Manager)



Cert No. CC-2543



# Medical Engineering & Services

**Things Beyond Thoughts**

Abhcon Crown, X1/411/21, Ambadi Lane  
Kokkalai, Thrissur-21. Ph: 0487 2420367  
Email: mescalibration.tsr@gmail.com  
Visit us @www.mescalibration.com

## CALIBRATION CERTIFICATE

Certificate Number	: MES24/03/053-02	ULR Number	: CC254324000000315F
Customer Details	: SPARSHA DIAGNOSTICS, UPPER BENDOOR, MANGALORE.		
Date of Calibration	: 14/03/2024	SRF Number	: 053
Next Calibration Due	: 13/03/2025	Date of Issue	: 14/03/2024

### DETAILS OF DEVICE UNDER CALIBRATION(DUC)

Nomenclature	: Micropipette	Serial No.	: YN4A009301
Make / Model	: Mispa i 2	ID No.	: -
Equipment Condition	: Good	Resolution	: 1 µl
Range	: 20 - 200 µl	Location	: Central Laboratory

### DETAILS OF MASTER EQUIPMENT USED

Nomenclature	: Electronic Precision Balance	Range	: 0 – 210 gm
Make	: Boeco Germany	Readability	: 0.01/0.1 mg
Model	: BBC-22	Serial No.	: 27104178
Certificate No.	: TVCSPL 23/10/1135-01	Calibration Due Date	: 02.10.2024

### ENVIRONMENTAL CONDITIONS

Temperature(°C)	Relative Humidity (%)	Atmospheric Pressure (mbar)	Water Temperature(°C)
26.5	56	1007	26.3

### CALIBRATION RESULTS

Sl. No.	Discipline : Mechanical			Parameter :		Volume	
	Set Volume	Standard Value	Systematic Error	Acceptable Systematic Error	Random Error	Acceptable Random Error	Expanded Uncertainty
	µl	µl	µl	± µl	µl	± µl	± µl
1	20	19.95	0.05	0.40	0.01	0.20	0.817
2	100	100.20	-0.20	1.40	0.09	0.60	0.838
3	200	200.42	-0.42	12.00	0.32	4.00	1.095

#### Notes: -

- The results in this certificate are only related to DUC submitted calibration and valid at the time of calibration under stated condition of measurements
- The measurement uncertainty is estimated at a confidence level of 95.45% with coverage factor k=2.
- Results in this certificate shall not be reproduced in partial or full without the written approval of the laboratory.
- Next calibration due date is mentioned as per customer requirement.
- Calibration certificate issued for scientific or industrial purpose only.
- Acceptable error limit is given as per ISO 8655-2:2002(E) Guidelines
- Calibration is carried out as per ISO 8655-6 Guidelines;MES work procedure: MES/WP/VOLUME.

Calibrated by

*Prabin KP*  
Prabin KP

(Calibration Engineer)



Authorized Signatory

*Binoj Thomas*  
Binoj Thomas  
(Technical Manager)





Cert No. CC-2543



# Medical Engineering & Services

**Things Beyond Thoughts**

Abhcon Crown, X1/411/21, Ambadi Lane  
Kokkai, Thrissur-21 Ph: 0487 2420367  
Email: mescalibration.tsr@gmail.com  
Visit us @www.mescalibration.com

## CALIBRATION CERTIFICATE

Certificate Number	: MES24/03/053-01	ULR Number	: CC254324000000314F
Customer Details	: SPARSHA DIAGNOSTICS, UPPER BENDOOR, MANGALORE.		
Date of Calibration	: 14/03/2024	SRF Number	: 053
Next Calibration Due	: 13/03/2025	Date of Issue	: 14/03/2024

### DETAILS OF DEVICE UNDER CALIBRATION(DUC)

Nomenclature	: Micropipette	Serial No.	: NK479663
Make / Model	: Matrix Gel System	ID No.	: -
Equipment Condition	: Good	Resolution	: 0 µl
Range	: 50 µL	Location	: Central Laboratory

### DETAILS OF MASTER EQUIPMENT USED

Nomenclature	: Electronic Precision Balance	Range	: 0 – 210 gm
Make	: Boeco Germany	Readability	: 0.01/0.1 mg
Model	: BBC-22	Serial No.	: 27104178
Certificate No.	: TVCSPL 23/10/1135-01	Calibration Due Date	: 02.10.2024

### ENVIRONMENTAL CONDITIONS

Temperature(°C)	Relative Humidity (%)	Atmospheric Pressure (mbar)	Water Temperature(°C)
26.3	58	1007	26.1

### CALIBRATION RESULTS

Sl. No.	Discipline : Mechanical			Parameter		Volume	
	Set Volume	Standard Value	Systematic Error	Acceptable Systematic Error	Random Error	Acceptable Random Error	Expanded Uncertainty
	µl	µl	µl	± µl	µl	± µl	± µl
1	50	50.06	-0.06	0.50	0.05	0.20	0.131

#### Notes: -

- The results in this certificate are only related to DUC submitted calibration and valid at the time of calibration under stated condition of measurements
- The measurement uncertainty is estimated at a confidence level of 95.45% with coverage factor k=2.
- Results in this certificate shall not be reproduced in partial or full without the written approval of the laboratory.
- Next calibration due date is mentioned as per customer requirement.
- Calibration certificate issued for scientific or industrial purpose only.
- Acceptable error limit is given as per ISO 8655-2:2002(E) Guidelines
- Calibration is carried out as per ISO 8655-6 Guidelines;MES work procedure: MES/WP/VOLUME.

Calibrated by  
*Prabin*  
Prabin KP  
(Calibration Engineer)



Authorized Signatory  
*Binoj Thomas*  
Binoj Thomas  
(Technical Manager)



ULR No. :	CC21442400001742F	Page 2 of 2
Certificate No. :	TVCSPL 24/03/0362-01	
<b>MECHANICAL CALIBRATION</b>		
(Volume)		
<b>Calibration Results</b>		
<b>2. Middle Volume :</b>	<b>10      μl</b>	<b>No. of Measurements :      10</b>
10.02	10.04	9.96
9.88	10.02	9.97
10.00	9.97	10.03
10.03	9.92	
<b>Mean Value :</b>	9.99    μl	
<b>Error Limits(±)</b>		
<b>Systematic Error :</b>	-0.01    μl	0.12      μl
<b>Systematic Error :</b>	-0.14    %	1.20      %
<b>Random Error :</b>	0.05    μl	0.08      μl
<b>Random Error :</b>	0.52    %	0.80      %
<b>Measurement Uncertainty :</b>	±      0.26    μl	
<b>3. Nominal Volume :</b>	<b>20      μl</b>	<b>No. of Measurements :      10</b>
19.97	19.91	20.09
19.98	20.00	19.97
19.93	20.06	19.95
20.04		
<b>Mean Value :</b>	19.99    μl	
<b>Error Limits(±)</b>		
<b>Systematic Error :</b>	-0.01    μl	0.20      μl
<b>Systematic Error :</b>	-0.03    %	1.00      %
<b>Random Error :</b>	0.06    μl	0.10      μl
<b>Random Error :</b>	0.29    %	0.50      %
<b>Measurement Uncertainty :</b>	±      0.26    μl	

**Remarks**

1. The reported Expanded Uncertainty is calculated at 95.45 % C.L. with coverage factor  $k=2$
2. The above Micropipette was within the error limits
3. Masters are Traceable to SI Units.

Calibrated by :  
**M Jeevitha**  
 Ms. Jeevitha  
 (Calibration Engineer)



Authorised by:  
*DVS*  
 D. Vētri Selvi  
 (QM & TM)

*...redefining the true value*





CC-2144

CERTIFICATE OF CALIBRATION

FT-Q-25		Page 1 of 2	
ULR No. :	CC214424000001742F	Date of Issue :	17-03-2024
Certificate No. :	TVCSPL 24/03/0362-01	Recom. Due Date :	14-03-2025
Date of Calibration :	15-03-2024	Customer Details :	M/s. Sparsha Diagnostics, Mangalore - 575 002.
Customer Details :		SRF No. :	0362
		Calibrated at :	Lab
		Date of Receipt :	15-03-2024
		Cond. On Receipt :	Satisfactory
Details of Test Instrument:			
Description :	Micropipette	Model No. :	Mispa-i-series
Range :	2-20µl	Serial No. :	OW12967
Least Count :	0.02µl	Id. No. :	--
Make :	Agappe	Accuracy :	As per Manual
Location :	Central Laboratory		
Details of Standard Used :			
Name	Certificate No.	Valid upto	Calibrated at
Weighing Machine	TVCSPL 24/03/0260-01	04-03-2025	CC-2144
Weighing Machine	TVCSPL 24/02/0219-03	23-02-2025	CC-2144
Work Instruction :	WI-M-03		
Environmental Details :	Temperature : 25±2°C	Relative Humidity : 50±10 % RH	
<b>MECHANICAL CALIBRATION</b>			
(Volume)			
<b>Calibration Results</b>			
1. Lower Volume :	2 µl	No. of Measurements :	10
1.951	1.952	1.972	1.991
1.954	1.992	1.966	1.961
1.976	2.003		
Mean Value :	1.972 µl		
Error Limits(±)			
Systematic Error :	-0.03 µl	0.08 µl	
Systematic Error :	-1.42 %	4.00 %	
Random Error :	0.02 µl	0.04 µl	
Random Error :	0.93 %	2.00 %	
Measurement Uncertainty :	± 0.094 µl		

Calibrated by :  
*M. Jeevitha*  
 Ms. Jeevitha  
 (Calibration Engineer)



Authorised by:  
*DV*  
 D. Vetri Selvi  
 (QM & TM)

...redefining the true value

