

CERTIFICATE OF CALIBRATION

Customer Name	CURA MULTISPECIALITY HOSPITAL	Certificate No.	CMHC 137
Address	Kammanahalli, Bengaluru, Karnataka 560043	Date of Calibration	07-12-2023
		Valid Upto	06-12-2024
		Date of Issue	08-12-2023

DUC* Details					
DUC* Name	Make/Model No	S No.	ASSET ID	LOCATION	Visual Inspection
HEMATOLOGY ANALYSER	ASPEN/AD 3200	3200PET-00385	CH/ME/LAB/333	LABORATORY	OK

Details of Reference Standards & Major Equipments Used		
Equipment Name	ELECTRICAL SAFETY ANALYZER	THERMO HYGROMETER
Make	FLUKE	HTC-2
Model/SR No.	ESA 612/2959012	...
Equipment Id	CHPL/MT-KL/ESA/002	
Calibration certificate No	ET/230117A/01	UCSL/07-23/202_03
Calibration Validity	23/01/2024	20/07/2024
Calibration By	STIC	UNIQUE CALIBRATION SOLUTIONS

**STANDARD USED IS IEC 60601-1

ELECTRICAL SAFETY TEST VALUES

TEST NO.	NAME OF THE TEST	READING	HIGH LIMIT	LOW LIMIT	UNIT	STATUS
1	MAIN VOLTAGE -LIVE TO NEUTRAL	229.4	250	225	V	PASS
2	MAIN VOLTAGE -LIVE TO EARTH	228.4	240	220	V	PASS
3	MAIN VOLTAGE-NEUTRAL TO EARTH	0.2	3	0	V	PASS
4	EQUIPMENT CURRENT	0.3	16	0.1	mA	PASS
5	PROTECTIVE EARTH RESISTANCE	0.3	2	0	Ω	PASS
6	EARTH LEAKAGE CURRENT - NC	0.6	1	0.5	mA	PASS
7	EARTH LEAKAGE CURRENT - OC	0.7	1	0.5	mA	PASS
8	ENCLOSURE LEAKAGE CURRENT -NC	0.4	0.5	0.1	mA	PASS
9	ENCLOSURE LEAKAGE CURRENT -OC	0.3	0.5	0.1	mA	PASS
10	INSULATION RESITANCE					
	1.MAINS-PROTECTIVE EARTH	OL	100	0	MΩ	PASS
	2.APPLIED PART TO PROTECTIVE EARTH	NA	100	0	MΩ	PASS
	3.MAINS TO APPLIED PART	NA	100	0	MΩ	PASS
	4.MAINS TO NEUTRAL	OL	100	0	MΩ	PASS
	5.APPLIED PART TO NEUTRAL	NA	100	0	MΩ	PASS

*NOTE (NC-NORMAL CONDITION, OC-OPEN CONDITON). REMARKS- *DUC- DEVICE UNDER CALIBRATION

CALIBRATED BY
ACHUTH P K



ISSUED AND APPROVED BY

PREJITH B
TECHNICAL MANAGER