

CERTIFICATE OF CALIBRATION

Customer Name	CURA MULTISPECIALITY HOSPITAL	Certificate No.	CMHC 132
Customer Address	Kammanahalli, Bengaluru, Karnataka 560043	Date of Calibration	07-12-2023
		Valid Upto	06-12-2024
		Date of Issue	08-12-2023

DUC* Details			
DUC* Name	CENTRIFUGE	S No.	ZHCN-09336
Make/Model No.	REMI/R-8C	Location	LABORATORY
ID No.	CH/ME/LAB/319	Visual Inspection	OK

Details of Reference Standards & Major Equipments Used		
Equipment Name	DIGITAL TACHO METER	THERMO HYGROMETER
Make	METRAVI	HTC-2
Model/SR No.	NCTM 1000/190518478	...
Equipment Id	CHPL/MT-KL/TAM/002	
Calibration certificate No	UCSL/07-23/202_05	UCSL/07-23/202_03
Calibration Validity	20/07/2024	20/07/2024
Calibration By	UNIQUE CALIBRATIONS	UNIQUE CALIBRATION SOLUTIONS

Environmental Condition			
Temperature	27°C	Relative Humidity	54%

Calibration Results				
SL NO	SET READING (RPM)	MEASURED READING (RPM)	TOLERANCE	STATUS
1	500	506	±5%	PASS
2	1000	1005	±5%	PASS
	2500	2512	±5%	PASS

CALIBRATION STATUS	PASS	✓	FAIL	
---------------------------	------	---	------	--

REMARKS:- *DUC- DEVICE UNDER CALIBRATION

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not used for commercial.

Calibrated By

ACHUTH P K



Issued and Approved By

PREJITH B
TECHNICAL MANAGER