

ASIAN TECHNOLOGY

(A HOUSE OF CALIBRATION)



CALIBRATION CERTIFICATE

Customer name	M/S: Ankur Pathology	Service request No. & date	P/00/19/03/24&19/03/2024
And	KL Tower Kadma Market Main	ULR no.	CC223924000002146F
Address	Road Opposite Kadma Market	Cert. No.	AT24000002146
	Auto Stand Kadma Jamshedpur	Date of Receipt of DUC	19/03/2024
	831005 Dist-East Singhbhum	Date of calibration	19/03/2024
	Jharkhand.	Date of issue	20/03/2024
	The second secon	Suggested due date	18/03/2025

, P	Instrume	nt Details	
Instrument name	Multi Thermometer	Location	Refrigerator Up
Make/Model No.	/ST-9283B	Accuracy	
Range / Size	-50 to 300 °C	Visual Inspection	OK
Least Count	0.1 °C	Calibration Performed at	Lab
I.D. No. / Sr. no.	MT-1 /		10

Detail of reference standards & Major equipments used			
Equipment Name	SPRT With Indicator		
Make	Tempsans & Fluke		
Model / SR No.	PT-100X1 & 1586A / 05640		
Certificate No.	CC35082300000098F		
Calibration Validity	25/10/2024		
Calibration by	S.B International		

Environmental	Temperature	(25±4)°C	Calibration Reference	IS:6274 (1971)
Condition	Relative Humidity	(50±20) %	Work Instruction	AT/WI-01TH

Serial No.	Std. Reading in (°C)	Measured Reading in (°C)	Uncertainty At 95% C.L. (coverage factor k=2)
01.	2.011	2.0	
02.	3.017	3.1	
03.	5.021	5.3	±0.3 °C
04.	6.039	6.2	
05.	7.874	8.0	

Remarks:

- (1) Standard equipment use for calibration are traceable to national/international standards.
- (2)The reported expended uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by coverage factork = 2 such that the coverage probability corresponds to approximately 95%.(3)The above results are valid at the time of and under the stated conditions measurement.(4)This certificate is refers only to the particular item submitted for calibration. (5) Next calibration due date given as requested by the customer.

Calibrated By (Calibration Engg./TM) (D.B SAXENA) Form No.- QF-47



Approved By (QM/IM) (NEERA) TYAGI) Page No. 1 of 1

---End of Report---