



## **CERTIFICATE OF CALIBRATION**

Certificate No : CC-2	4-25-S-138/02		Page. No: 1 of	
Customer Name & Address M/s. Vishnu Clinical Lab, #Rasipuram, Main Road, Vaiyappamalai.		ULR No.	CC2725240000001978F	
		SRF No.	VCI-24-25-S-138 10.05.2024 10.05.2024	
		Date of Receipt		
		Date of Calibration		
		Due Date of Calibration	09.05.2025	
		Date of Issue	11.05.2024	
Details of Unit Unde	r Calibration			
Description	Refrigerator-Reagent	Make	Samsung	
Range	-10 to 8°C	Model		
Resolution	1°C	Condition of UUC	Good	
Serial Number		Instrument Location		
ID Number	VL/LAB/RFR/001	Cal. At	On Site	

Environmental Condition		Calibration Method Used	
Temperature (°C)	30±5	National / International Standard	DKD-R 5-7
Humidity (%RH)	30 - 75	Cal Procedure No	VCI-L03-WI-THER-005

## Standard Used

Sl. No.	Description	ID.No.	Certificate No.	Make/Model	Traceability	Valid till
1	RTD Sensor with Indicator	VCI-T-RTD-002& VCI-ET-DT-013	CC-24-25-L-005/03	Aparajit Instruments/ (Pt100) &Yudian/AI-5500	Venus Calibration CC-2725	05.01.2025

Result of Calibration				
Sl. No.	Standard Reading (°C)	UUC Reading (°C)	Error (°C)	Measurement Uncertainty ± (°C)
01	2.04	2	-0.04	0.64
02	4.14	4	-0.14	0.64
03	6.18	6	-0.18	0.64
04	8.22	8	-0.22	0.64

## Remarks:

- 1. UUC is Defined as the Unit Under Calibration.
- This Calibration Certificate Results relates only to the above UUC & This report shall not be reproduced except in full/ part without approval of the Venus Calibration and Instruments.
- 3. The reported results are valid at the time of under stated condition of measurement.
- 4. Calibration of the UUC are traceable to National / International Standards.
- 5. The recalibration interval shall be decided by the user.
- 6. Nabl-133 guidelines are adopted for use of NABL symbol.
- 7. The Measurement uncertainty is expressed at 95.45% confidence level with coverage factor k=2.
- 8. Temperature Chart: ITS-90

Checked & Issued by:

R.Bhuvaneshwari (Engineer - Calibration)

\*\*\*\*\*\*END OF THE CERTIFICATE\*\*\*\*\*\*

CHENNAI CHENNA

Authorised by:

C.Dhinakaran

(Technical / Quality Manager)