



CC-2725

**CERTIFICATE OF CALIBRATION**

Certificate No : CC-24-25-S-138/01

Page No : 1 of 2

Customer Name & Address M/s. Vishnu Clinical Lab, #Rasipuram, Main Road, Vaiyappamalai.	ULR No.	CC272524000001977F
	SRF No.	VCI-24-25-S-138
	Date of Receipt	10.05.2024
	Date of Calibration	10.05.2024
	Due Date of Calibration	09.05.2025
	Date of Issue	11.05.2024

Details of Unit Under Calibration

Description	Centrifuge	Make/Model	Monisto/8-BKT
Range	0 to 3500 rpm, 10 to 60 min	Operating Range	0 to 3500 rpm, 10 to 60 min
Resolution	1 rpm, 10 min	Condition of UUC	Good
Serial Number	--	Instrument Location	--
ID Number	VL/LAB/CEN/001	Cal. At	On Site

Environmental Condition		Calibration Method Used	
Temperature (°C)	23±5	National / International Standard	SANAS TR 45-02, NIST 960-12
Humidity (%RH)	40-70	Cal Procedure No	VCI-L03-WI-SPE-001, VCI-L03-WI-ET-005

Standard Used

Sl. No.	Description	ID.No.	Certificate No.	Make/Model	Traceability	Valid till
1	Digital Tachometer	VCI-M-TM-068	2023-24/EQN/2307136-01	Line Seiki/TM-5000	Equation Calibration CC-2760	24.07.2024
2	Digital Timer	VCI-ET-DT-005	CC-24-25-L-005/02	M.TEC /CT7	Venus Calibration CC- 2725	05.01.2025

Remarks:

- UUC is Defined as the Unit Under Calibration.
- This Calibration Certificate Results relates only to the above UUC & This report shall not be reproduced except in full/ part without approval of the Venus Calibration and Instruments.
- The reported results are valid at the time of under stated condition of measurement.
- Calibration of the UUC are traceable to National / International Standards.
- The recalibration interval shall be decided by the user.
- Nabl-133 guidelines are adopted for use of NABL symbol.
- The Measurement uncertainty is expressed at 95.45% confidence level with coverage factor $k=2$.

Checked & Issued by:

R.B.
R.Bhuvaneshwari
(Engineer - Calibration)



Authorised by:

C.D.
C.Dhinakaran
(Technical / Quality Manager)