

Instrument : Roche cobas c 311

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Urea	mg/dl	17.80	20.23	-1.90	-0.94
@ Creatinine	mg/dl	0.70	0.67	0.24	-0.29
@ Bilirubin Total	mg/dl	0.09	0.12	-0.50	-0.12
@ Total Protein	g/dL	5.73	5.68	0.31	0.15
@ Albumin	g/dL	3.20	3.35	-0.85	-0.62
@ Uric Acid	mg/dl	3.60	3.63	-0.22	-0.07
@ Calcium	mg/dl	6.80	6.79	0.12	0.28
@ Phosphorous Inorganic	mg/dl	5.08	4.83	0.80	0.32
@ Cholesterol	mg/dl	136.00	129.47	1.59	0.66
● Triglycerides	mg/dl	128.00	129.41	-0.11	0.08
@ HDL	mg/dl	22.00	21.83	0.25	-0.01
@ LDL	mg/dl	82.00	80.88	0.22	0.18
@ ALT	U/L	7.00	10.61	-1.62	-0.64
@ AST	U/L	15.00	15.78	-0.24	-0.43
@ ALP	U/L	46.00	46.68	0.15	0.21
@ GGT	U/L	20.00	21.18	-0.33	0.07

Instrument : Erba EC 90

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Chloride	mmol/L	81.40	82.25	0.72	0.80
@ Sodium	mmol/L	105.40	109.24	-0.72	0.07
@ Potassium	mmol/L	3.74	3.86	-0.65	0.14

Legend @ : Acceptable
 : 2.0 < z score < 3.0 - Warning
 X : z score ≥ 3 - Unacceptable
 # : Not Evaluated
 ⊙ : Delayed Result Entry
 * : Not considered for evaluation.

Total Parameters	19
Not Evaluated Parameters	0
Evaluated Parameters	19
Outlier Parameters (X)	0
EQAS Score Biochemistry	100.00 %

Dr Puneet Nigam

Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A
 Kohinoor Mall, Kiroli Road, Kurla (W),
 Mumbai - 400070

Satish

Instrument : Roche cobas c 311

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Glucose	mg/dl	68.00	67.06	0.11	-0.09
Urea	mg/dl	17.50	20.78	-2.32	-0.94
@ Creatinine	mg/dl	0.70	0.76	-0.92	-0.29
@ Bilirubin Total	mg/dl	0.27	0.27	-0.04	-0.12
@ Bilirubin Direct	mg/dl	0.12	0.12	0.03	-0.27
@ Total Protein	g/dL	5.90	5.86	0.28	0.15
@ Albumin	g/dL	3.40	3.50	-0.52	-0.62
@ Uric Acid	mg/dl	3.80	3.76	0.08	-0.07
@ Calcium	mg/dl	6.50	6.32	0.57	0.28
● Phosphorous Inorganic	mg/dl	4.74	4.47	1.13	0.32
@ Cholesterol	mg/dl	141.00	131.99	0.89	0.66
@ Triglycerides	mg/dl	161.00	159.07	0.55	0.08
@ LDL	mg/dl	86.00	84.13	0.50	0.18
@ ALP	U/L	48.00	47.34	0.47	0.21
@ GGT	U/L	20.00	22.59	-0.65	0.07

Instrument : Erba EC 90

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Chloride	mmol/L	84.70	84.32	0.71	0.80
@ Sodium	mmol/L	109.80	111.29	-0.69	0.07
@ Potassium	mmol/L	4.01	4.06	-0.31	0.14

Legend @ : Acceptable

○ : 2.0 < z score < 3.0 - Warning

X : z score ≥ 3 - Unacceptable

: Not Evaluated

⊙ : Delayed Result Entry

* : Not considered for evaluation.

Total Parameters	18
Not Evaluated Parameters	0
Evaluated Parameters	18
Outlier Parameters (X)	0
EQAS Score Biochemistry	100.00 %

Dr. Nigam

Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-418
 Commercial Building - 1A
 Kohinoor Mall, Kirod Road, Kurla (W),
 Mumbai - 400070

S. Suresh