

Instrument : -

Analyte	Standard Unit	Result Value	Accepted Value
@ Neutrophil.	-	<40%	<40%
# Lymphocyte.	-	>40%	
@ Eosinophil.	-	<6%	<6%

  

Legend @ : Acceptable	Total Parameters	3
# : Acceptable	Not Evaluated Parameters	1
X : Unacceptable	Evaluated Parameters	2
# : Not Evaluated	Outlier Parameters (X)	0
⊙ : Delayed Result Entry	EQAS Score Haematology-PS Slide Count Qualitative	<u>100.00 %</u>

*S. S. S. S. S.*

*✓*

*Dr. Nigam*

**Dr Puneet Kumar Nigam**  
 PT coordinator & Technical Manager, MHL EQAS  
 Unit No. 409-416.  
 Commercial Building - 1A  
 Kohinoor Mall, Kiroli Road, Kurla (W),  
 Mumbai - 400070

Instrument : Erba ELite 580

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Erythrocyte (RBC) Count	mill/cu.mm	2.00	2.04	-0.85	-0.79
@ Total Leucocytes (WBC) count	/c.mm	3870.00	3810.00	0.42	0.33
@ Haemoglobin	g/dL	5.90	5.95	-0.49	-0.46
@ MCV (Mean Corpuscular Volume)	fL	136.90	136.92	-0.01	-0.31
@ MCH (Mean Corpuscular Hb)	pg	29.40	29.35	0.49	0.14
@ MCHC (Mean Corpuscular Hb Concentration)	g/dL	21.50	21.45	0.24	0.26
@ Platelet Count	1000/ $\mu$ L	100.00	113.92	-1.05	-0.27
@ Hematocrit	%	27.40	27.81	-0.45	-0.53
@ RDW CV%	%	22.10	21.45	0.65	0.12

- **gend @** : Acceptable
- : 2.0 < z score < 3.0 - Warning
- X** : z score  $\geq$  3 - Unacceptable
- #** : Not Evaluated
- ⊙ : Delayed Result Entry
- \* : Not considered for evaluation.

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	0
EQAS Score Haematology - CBC	100.00 %

*Dr Puneet Kumar Nigam*

**Dr Puneet Kumar Nigam**  
 PT coordinator & Technical Manager, MHL EQAS  
 Unit No. 409-416.  
 Commercial Building - 1A  
 Kohinoor Mall, Kiroi Road, Kurla (W),  
 Mumbai - 400070