



Metropolis Healthcare Ltd., EQAS

REDCLIFFE LIFETECH PRIVATE LIMITED, AGRA (1093)

Outlier And Analyte Summary Report

Result Details For Cycle No 230412 and Sample No 02

Report Date : 13/02/2024

Instrument : Roche cobas c 311

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ HbA1c	%	5.00	5.22	-0.94	0.34

Legend @ : Acceptable	Total Parameters	1
○ : 2.0 < z score < 3.0 - Warning	Not Evaluated Parameters	0
X : z score ≥ 3 - Unacceptable	Evaluated Parameters	1
# : Not Evaluated	Outlier Parameters (X)	0
⊙ : Delayed Result Entry	EQAS Score HbA1c	100.00 %
* : Not considered for evaluation.		

Dr Puncet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A
Kohinoor Mall, Kirod Road, Kurla (W),
Mumbai - 400070



Metropolis Healthcare Ltd., EQAS

REDCLIFFE LIFETECH PRIVATE LIMITED, AGRA (1093)

Outlier And Analyte Summary Report

Result Details For Cycle No 230412 and Sample No 01

Report Date : 13/02/2024

Instrument : Roche cobas c 311

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ HbA1c	%	5.00	5.06	-0.12	0.34

Legend @ : Acceptable
 : 2.0 < z score < 3.0 - Warning
 X : z score ≥ 3 - Unacceptable
 # : Not Evaluated
 ⊙ : Delayed Result Entry
 * : Not considered for evaluation.

Total Parameters	1
Not Evaluated Parameters	0
Evaluated Parameters	1
Outlier Parameters (X)	0
EQAS Score HbA1c	100.00 %

Satisfactory

Dr Puneet Kumar Nigam

Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A
 Kohinoor Mall, Kirod Road, Kurla (W),
 Mumbai - 400070