



ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO. : 240502A/N-117	Page : 1 of 1			
CALIBRATED ON. : 04 / 05 / 2024	Ref. Std. : CG-11			
CALIBRATION DUE. : 03 / 05 / 2025	C.P. No. : ACS/CP/07			
SRF No. & Dt. OF RECPT : 240502A/N Dt. 02 05 2024	Phy. Inspection : O.K.			
ULR No. & ISSUE DATE : CC342724000000494F Dt. 07.05.2024	Cal. At : Site / Lab			
CALIBRATED FOR : M/s Redcliffe Lifetech Pvt. Ltd. DC II 06_ II Floor Shanti. Madhuvan Plaza, Agra - 282002				
ITEM CALIBRATED : Digital Temperature Indicator Of Refrigerator				
Make : -	Range : (-50 to 110) °C			
L. Count : 0.1°C	ID No. : RCL/BIO/D Temp-01			
Location : Clinical Biochemistry	S. No. : -			
ENVIRONMENT CONDITION: Temp. { 30°C ± 10°C } R.H. {30% to 80%}				
STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Universal Calibrator	ACS/UC/01 / 300561	Anshaankan	AIPL/041123/1-1	07/11/2024
RESULTS : ET Discipline				
* U. C. INSTT. READS (°C)	STANDARD READING (°C)	ERROR (°C)		
2.1	2	0.1		
4.4	4	0.4		
6.5	6	0.5		
8.6	8	0.6		
(average of readings)				

REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is $\pm 1.00^\circ\text{C}$ at coverage Factor $k=1.96$.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.



Prepared by

... End of Calibration Certificate ...

