



ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO. : 240502A/N-101	Page : 1 of 1
CALIBRATED ON. : 02 / 05 / 2024	Ref. Std. : ISO : 8655:6
CALIBRATION DUE. : 01 / 05 / 2025	C.P. No. : ACS/CP/08
SRF No. & Dt. OF RECPT : 240502A/N Dt. 02.05.2024	Phy. Inspection : O.K
ULR No. & ISSUE DATE : CC342724000000478F Dt. 07.05.2024	Cal. At : Site / Lab

CALIBRATED FOR : M/s Redcliffe Lifetech Pvt. Ltd. DC II 06_ II Floor Shanti. Madhuvan Plaza, Agra - 282002

ITEM CALIBRATED : Micro Pipette	
Make : Thermo Scientific	Range : (2 to 20) µl
L. Count : -	ID No. : RCL/HEM/PIP-01
Location : Haematology	S. No. : TW09716

ENVIRONMENT CONDITION: Temp. { 25°C ± 4°C } R.H. { 30% to 75% }

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Weighing Balance	ACS/DB/02/D449927064	BlueBox	BBT/691/JAN/24	04/01/2025

RESULTS : Mechanical Discipline				
Nominal Capacity (µl)	Graduation (µl)	Nominal Value (µl)	Volume at 20 °C (µl)	Error (µl)
		10	9.83	0.17
20	-	20	19.71	0.29

REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is $\pm 0.4 \mu\text{l}$ at coverage factor $k=1.96$.

*U.C.

: Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- This calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.



Prepared by

...End of Calibration Certificate...

