



# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



## Calibration Certificate

<b>CERTIFICATE NO.</b> : 240502A/N-102	<b>Page</b> : 1 of 1
<b>CALIBRATED ON.</b> : 02 / 05 / 2024	<b>Ref Std</b> : ISO 6655 6
<b>CALIBRATION DUE.</b> : 01 / 05 / 2025	<b>C.P. No</b> : ACS/CP/08
<b>SRF No. &amp; Dt. OF RECPT</b> : 240502A/N Dt. 02 05 2024	<b>Phy. Inspection</b> : O K
<b>ULR No. &amp; ISSUE DATE</b> : CC342724000000479F Dt. 07.05.2024	<b>Cal. At</b> : Site / Lab

<b>CALIBRATED FOR</b> : M/s Redcliffe Lifetech Pvt. Ltd DC II 06 II Floor Shanti. Madhuvan Plaza, Agra - 282002
---

<b>ITEM CALIBRATED</b> : Micro Pipette	
<b>Make</b> : Thermo Scientific	<b>Range</b> : (10 to 100) µl
<b>L. Count</b> : -	<b>ID No</b> : RCL/HEM/PIP-02
<b>Location</b> : Haematology	<b>S. No.</b> : TW13783

<b>ENVIRONMENT CONDITION:</b> Temp. ( 25°C ± 4°C) R.H. (30% to 75%)
---

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Weighing Balance	ACS/DB/02/D449927064	BlueBox	BBT/691/JAN/24	04/01/2025

RESULTS : Mechanical Discipline				
Nominal Capacity (µl)	Graduation (µl)	Nominal Value (µl)	Volume at 20 °C (µl)	Error (µl)
		10	9.78	0.22
100	-	50	49.61	0.39
		100	99.53	0.47

### REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is ± 0.4 µl at coverage factor k=1.96.

### \*U.C.

#### Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use
- Conformity statement not provided since not required



Prepared by

...End of Calibration Certificate...

