



ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

C-23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO.	: 240502A/N-103	Page	: 1 of 1	
CALIBRATED ON.	: 02 / 05 / 2024	Ref. Std	: ISO : 8655:6	
CALIBRATION DUE.	: 01 / 05 / 2025	C.P. No	: ACS/CP/08	
SRF No. & Dt. OF RECPT	: 240502A/N Dt. 02.05.2024	Phy. Inspection	: O.K	
ULR No. & ISSUE DATE	: CC342724000000480F Dt. 07.05.2024	Cal. At	: Site / Lab	
CALIBRATED FOR	: M/s Redcliffe Lifetech Pvt Ltd DC II 06, II Floor Shanti Madhuvan Plaza, Agra - 282002			
ITEM CALIBRATED	: Micro Pipette			
	Make	Thermo Scientific	Range	: (100 to 1000) µl
	L. Count	-	ID No	: RCL/HEM/PIP-03
	Location	Haematology	S. No.	: TW13656
ENVIRONMENT CONDITION:	Temp { 25°C ± 4°C } R.H. {30% to 75%}			
STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No / S No	TRACEABILITY	CERTIFICATE No.	DUE DATE
Weighing Balance	ACS/DB/02/D449927064	BlueBox	BBT/691/JAN/24	04/01/2025
RESULTS :	Mechanical Discipline			
Nominal Capacity (µl)	Graduation (µl)	Nominal Value (µl)	Volume at 20 °C (µl)	Error (µl)
		100	99.78	0.22
1000	-	500	500.19	(-) 0.19
		1000	1000.23	(-) 0.23

REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is ± 2.2 µl at coverage factor k=1.96.

*U.C.

: Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.



Prepared by

... End of Calibration Certificate ...

