



# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



## Calibration Certificate

CERTIFICATE NO. : 240502A/N-104		Page : 1 of 1		
CALIBRATED ON. : 02 / 05 / 2024		Ref. Std : ISO : 8655 B		
CALIBRATION DUE. : 01 / 05 / 2025		C.P. No : ACS/CP/08		
SRF No. & Dt. OF RECPT : 240502A/N Dt. 02.05.2024		Phy. Inspection : O.K.		
ULR No. & ISSUE DATE : CC342724000000481F Dt. 07.05.2024		Cal At : Site / Lab		
CALIBRATED FOR : M/s Redcliffe Lifetech Pvt. Ltd DC II 06, II Floor Shanti. Madhuvan Plaza, Agra - 282002				
ITEM CALIBRATED : Micro Pipette				
Make : Thermo Scientific		Range : (2 to 20) µl		
L Count : -		ID No. : RCL/BIO/PIP-04		
Location : Clinical Biochemistry		S. No. : TW08213		
ENVIRONMENT CONDITION: Temp. { 25°C ± 4°C } R.H. { 30% to 75% }				
STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No / S No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Weighing Balance	ACS/DB/02/D449927064	BlueBox	BBT/691/JAN/24	04/01/2025
RESULTS : Mechanical Discipline				
Nominal Capacity (µl)	Graduation (µl)	Nominal Value (µl)	Volume at 20 °C (µl)	Error (µl)
		10	9.63	0.37
20	-	20	19.83	0.17

### REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is ± 0.4 µl at coverage factor k=1.96.

\*U.C.

: Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmity statement not provided since not required.



Prepared by

....End of Calibration Certificate....

