




# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



## Calibration Certificate

<b>CERTIFICATE NO.</b> : 240502A/N-105	<b>Page</b> : 1 of 1
<b>CALIBRATED ON.</b> : 02 / 05 / 2024	<b>Ref. Std.</b> : ISO : 8655:6
<b>CALIBRATION DUE.</b> : 01 / 05 / 2025	<b>C.P. No.</b> : ACS/CP/08
<b>SRF No.&amp; Dt.OF RECPT</b> : 240502A/N Dt. 02.05.2024	<b>Phy. Inspection</b> : O.K.
<b>ULR No. &amp; ISSUE DATE</b> : CC342724000000482F Dt. 07.05.2024	<b>Cal. At</b> : Site / Lab

**CALIBRATED FOR** : M/s Redcliffe Lifetech Pvt. Ltd.  
DC II 06\_ II Floor Shanti. Madhuvan Plaza, .  
Agra - 282002

<b>ITEM CALIBRATED</b> : Micro Pipette	
Make : Thermo Scientific	Range : (10 to 100) µl
L. Count : -	ID No. : RCL/BIO/PIP-05
Location : Clinical Biochemistry	S. No. : TH48653

**ENVIRONMENT CONDITION:** Temp. { 25°C ± 4°C } R.H. {30% to 75%}

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Weighing Balance	ACS/DB/02/D449927064	BlueBox	BBT/691/JAN/24	04/01/2025

RESULTS : Mechanical Discipline				
Nominal Capacity (µl)	Graduation (µl)	Nominal Value (µl)	Volume at 20 °C (µl)	Error (µl)
		10	9.59	0.41
100	-	50	49.64	0.36
		100	99.51	0.49

**REMARKS** : The Expanded uncertainty of measurement at 95% confidence level is ± 0.4 µl at coverage factor k=1.96.

- \*U.C.** : Under Calibration.
- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
  - The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
  - This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
  - The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
  - Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
  - Conformity statement not provided since not required.



Prepared by

....End of Calibration Certificate....

