



ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO.	: 240502A/N-107	Page	: 1 of 1	
CALIBRATED ON.	: 02 / 05 / 2024	Ref. Std.	: ISO : 8655 6	
CALIBRATION DUE.	: 01 / 05 / 2025	C.P. No.	: ACS/CP/08	
SRF No. & Dt. OF RECPT	: 240502A/N Dt. 02.05.2024	Phy. Inspection	: O.K.	
ULR No. & ISSUE DATE	: CC34272400000484F Dt. 07.05.2024	Cal. At	: Site / Lab	
CALIBRATED FOR	: M/s Redcliffe Lifetech Pvt. Ltd. DC II 06_ II Floor Shanti. Madhuvan Plaza, - Agra - 282002			
ITEM CALIBRATED	: Micro Pipette			
	Make : Erba	Range	: (10 to 100) µl	
	L. Count : -	ID No.	: RCL/BIO/PIP-07	
	Location : Clinical Biochemistry	S. No.	: YE213AS0159694	
ENVIRONMENT CONDITION:	Temp. { 25°C ± 4°C } R.H. {30% to 75%}			
STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Weighing Balance	ACS/DB/02/D449927064	BlueBox	BBT/691/JAN/24	04/01/2025
RESULTS :				Mechanical Discipline
Nominal Capacity (µl)	Graduation (µl)	Nominal Value (µl)	Volume at 20 °C (µl)	Error (µl)
		10	9.48	0.52
100	-	50	49.78	0.22
		100	99.61	0.39

REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is ± 0.4 µl at coverage factor k=1.96.

*U.C.

: Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.



Prepared by

... End of Calibration Certificate...

