



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - APRIL 2024



Lab Name **ANUSHKA DIAGNOSTICS** Lab No **19709**
Constituent Group **Chemistry II** Date of Result Entered : **19/04/2024**
PT item **Lyophilized human serum based** Date of Report Published : **08/05/2024**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	AV	Participants		Your Value	Z Score	u*
						CV	SDPA			
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	284	205.24	8.55	17.54	164 mg/dL	-2.35	2.08
2	UREA II	Others (Urease Berthelot / Ned Dye) II	Any Analyser (Automation / Semi Automation)	844	37.77	12.71	4.80	37.7 mg/dL	-0.01	0.33
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Any Analyser (Automation / Semi Automation)	218	3.74	11.72	0.44	4.47 mg/dL	1.67	0.06
4	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	257	5.32	9.01	0.48	5.53 g/dL	0.44	0.06
5	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation)	259	3.19	10.03	0.32	3.16 g/dL	-0.09	0.04
6	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	253	7.02	10.51	0.74	6.36 mg/dL	-0.89	0.09
7	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	276	106.46	10.03	10.68	112 mg/dL	0.52	1.29
8	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	241	332.01	8.77	29.12	326 mg/dL	-0.21	3.75

u* - Method of Uncertainty

Z-Score	Interpretation
$ z \leq 2.0$	Acceptable
$2.0 < z < 3.0$	Warning Signal
$ z \geq 3.0$	Unacceptable (action Signal)

LAB ADDRESS :

ANUSHKA DIAGNOSTICS
ANUSHKA LIFECARE PVT LTD, 115/102, STATION ROAD,
BOLPUR
WEST BENGAL 731204

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

******* End of Report *******