



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - MARCH 2024



PC-1024

Lab Name **ANUSHKA DIAGNOSTICS** Lab No **19709**
Constituent Group **Chemistry II** Date of Result Entered : **18/03/2024**
PT item **Lyophilized human serum based** Date of Report Published : **06/04/2024**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	AV	Participants		Your Value	Z Score	u*
						CV	SDPA			
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	282	397.15	8.45	33.54	349 mg/dL	-1.44	3.99
2	UREA II	Others (Urease Berthelot / Ned Dye) II	Any Analyser (Automation / Semi Automation)	856	64.58	11.79	7.62	64.7 mg/dL	0.02	0.52
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Any Analyser (Automation / Semi Automation)	206	6.95	11.77	0.82	7.17 mg/dL	0.27	0.11
4	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	230	4.11	8.64	0.36	5.46 g/dL	3.75	0.05
5	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation)	239	2.55	11.99	0.31	2.48 g/dL	-0.23	0.04
6	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	224	4.22	11.01	0.46	3.76 mg/dL	-0.99	0.06
7	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	265	86.20	11.21	9.66	92.1 mg/dL	0.61	1.19
8	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	221	74.39	15.46	11.50	71.9 mg/dL	-0.22	1.55

u* - Method of Uncertainty

Z-Score	Interpretation
$ z \leq 2.0$	Acceptable
$2.0 < z < 3.0$	Warning Signal
$ z \geq 3.0$	Unacceptable (action Signal)

LAB ADDRESS :

ANUSHKA DIAGNOSTICS
ANUSHKA LIFE CARE PVT LTD, 115/102,STATION ROAD,
BOLPUR
WEST BENGAL731204

Coordinator Contact Details:
Email:clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

******* End of Report *******