



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - JANUARY 2024



PC-1024

Lab Name **ANUSHKA DIAGNOSTICS** Lab No **19709**
Constituent Group **Chemistry II** Date of Result Entered : **21/01/2024**
PT item **Lyophilized human serum based** Date of Report Published : **06/02/2024**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	AV	Participants		Your Value	Z Score	u*
						CV	SDPA			
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	257	152.86	8.28	12.65	149 mg/dL	-0.31	1.58
2	UREA II	Others (Urease Berthelot / Ned Dye) II	Any Analyser (Automation / Semi Automation)	745	47.32	11.05	5.23	56.1 mg/dL	1.68	0.38
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Any Analyser (Automation / Semi Automation)	206	1.73	11.25	0.19	2.15 mg/dL	2.16	0.03
4	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	216	5.08	9.00	0.46	5.39 g/dL	0.68	0.06
5	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation)	236	3.03	10.48	0.32	3 g/dL	-0.09	0.04
6	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	223	8.22	9.41	0.77	8.73 mg/dL	0.66	0.10
7	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	220	103.08	10.04	10.35	111 mg/dL	0.77	1.40
8	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	240	148.09	12.09	17.90	143 mg/dL	-0.28	2.31

u* - Method of Uncertainty

Z-Score	Interpretation
$ z \leq 2.0$	Acceptable
$2.0 < z < 3.0$	Warning Signal
$ z \geq 3.0$	Unacceptable (action Signal)

LAB ADDRESS :

ANUSHKA DIAGNOSTICS
ANUSHKA LIFE CARE PVT LTD, 115/102, STATION ROAD
BOLPUR
WEST BENGAL731204

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

******* End of Report *******