



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - DECEMBER 2023



PC-1024

Lab Name **ANUSHKA DIAGNOSTICS** Lab No **19709**  
 Constituent Group **Chemistry II** Date of Result Entered : **19/12/2023**  
 PT item **Lyophilized human serum based** Date of Report Published : **02/01/2024**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	AV	Participants		Your Value	Z Score	u*
						CV	SDPA			
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation )	286	104.19	8.88	9.25	89.1 mg/dL	-1.63	1.09
2	UREA II	Others ( Urease Berthelot / Ned Dye ) II	Any Analyser (Automation / Semi Automation )	849	105.59	13.68	14.44	127 mg/dL	1.48	0.99
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Any Analyser (Automation / Semi Automation )	239	1.90	12.08	0.23	2.2 mg/dL	1.31	0.03
4	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation )	215	5.10	6.76	0.34	5.57 g/dL	1.36	0.05
5	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation )	261	3.11	10.34	0.32	3.19 g/dL	0.25	0.04
6	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation )	253	5.88	11.26	0.66	6.33 mg/dL	0.68	0.08
7	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation )	262	105.40	8.86	9.33	114 mg/dL	0.92	1.15
8	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation )	237	145.32	10.39	15.10	128 mg/dL	-1.15	1.96

u\* - Method of Uncertainty

Z-Score	Interpretation
$ z  \leq 2.0$	Acceptable
$2.0 <  z  < 3.0$	Warning Signal
$ z  \geq 3.0$	Unacceptable (action Signal)

## LAB ADDRESS :

**ANUSHKA DIAGNOSTICS**  
**ANUSHKA LIFE CARE PVT LTD, 115/102,STATION ROAD,**  
**BOLPUR**  
**WEST BENGAL731204**

Coordinator Contact Details:  
 Email:clinqc@cmcvellore.ac.in  
 Contact Number: 0416-2283102

*Pamela Christudoss*

Dr. Pamela Christudoss  
 CMC EQAS Coordinator  
 Christian Medical College, Vellore

**Homogeneity and Stability of the sample is passed.**

**Data in CMC EQAS reports is confidential**

**CMC EQAS does not sub contract any components**

**\*\*\*\*\* End of Report \*\*\*\*\***