



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - NOVEMBER 2023



PC-1024

Lab Name **ANUSHKA DIAGNOSTICS** Lab No **19709**  
 Constituent Group **Chemistry II** Date of Result Entered : **16/11/2023**  
 PT item **Lyophilized human serum based** Date of Report Published : **05/12/2023**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	AV	Participants		Your Value	Z Score	u*
						CV	SDPA			
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation )	301	247.49	8.06	19.96	177 mg/dL	-3.53	2.30
2	UREA II	Others ( Urease Berthelot / Ned Dye ) II	Any Analyser (Automation / Semi Automation )	917	66.61	10.47	6.98	82.4 mg/dL	2.26	0.46
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Any Analyser (Automation / Semi Automation )	243	4.70	11.31	0.53	4.98 mg/dL	0.53	0.07
4	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation )	269	4.65	10.22	0.48	4.95 g/dL	0.63	0.06
5	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation )	267	2.78	12.58	0.35	2.63 g/dL	-0.43	0.04
6	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation )	192	8.11	7.08	0.57	8.7 mg/dL	1.03	0.08
7	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation )	279	96.27	9.55	9.20	102 mg/dL	0.62	1.10
8	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation )	249	83.07	14.78	12.28	85.5 mg/dL	0.20	1.56

u\* - Method of Uncertainty

Z-Score	Interpretation
$ z  \leq 2.0$	Acceptable
$2.0 <  z  < 3.0$	Warning Signal
$ z  \geq 3.0$	Unacceptable (action Signal)

## LAB ADDRESS :

**ANUSHKA DIAGNOSTICS**  
**ANUSHKA LIFE CARE PVT LTD, 115/102, STATION ROAD**  
**BOLPUR**  
**WEST BENGAL731204**

Coordinator Contact Details:  
 Email:clinqc@cmcvellore.ac.in  
 Contact Number: 0416-2283102

*Pamela Christudoss*  
 Dr. Pamela Christudoss  
 CMC EQAS Coordinator  
 Christian Medical College, Vellore

**Homogeneity and Stability of the sample is passed.**

**Data in CMC EQAS reports is confidential**

**CMC EQAS does not sub contract any components**

**\*\*\*\*\* End of Report \*\*\*\*\***