



# TMSCC

TESTING MACHINE SERVICE AND CALIBRATION CENTRE  
Precision is Our Destination.....

An ISO 9001 : 2015 Certified company  
Calibration of Various Instruments & Testing Machines...



CC - 3125

Form No.: TMSCC/R/23

### CALIBRATION CERTIFICATE

CALIBRATION CERTIFICATE NO.: 2024/02/778 Page 1 of 1

ISSUE DATE: 16-Feb-2024

ULR No.: CC31252400002631F

1.0 ISSUED TO: M/s: Anushka Diagnostics  
119/148, Chandidas Road, Trisulapatty,  
P.O.: Bolpur, Birbhum - 731204

1.1 Service Request Form No.: SRF/2024/02/12/03

1.2 Service Request Date: 12-02-2024

1.3 Location: At Lab

1.4 Description identification of item to be calibrated:

i Name:	Digital Thermo Hygrometer	ii Make:	Aptechdeals
iii Model / Type No.:	HTC-1	iv S/L No.:	-
v I.D.No.:	AD/DTHM/01	vi Job Code No.:	2024/02/778
vii Range:	(0 to 70)°C(10 to 99)%RH	viii Resolution:	0.1°C/1%RH
ix Accuracy:	As Per DKD-R5-7, DKD R5 - 1	x End User:	-

1.5 Full / Partial Calibration: Full Calibration

1.6 Applicable specification of item to be calibrated: Accuracy / permissible limit : Not Specified.

1.7 Date of receipt of item : 12-Feb-2024 1.8 Date of calibration : 12-Feb-2024

1.9 Calibration due on : 11-Feb-2025 2.0 Frequency of calibration once in : 12 Months

2.1 Environmental condition during calibration :  
Temperature : 25.2°C  
Humidity : 52% RH

2.2 Basis of calibration : SOP/10/02, SOP/10/04

2.3 Traceability : Standards used for calibration are traceable to National Standards through NABL accredited Laboratory.

Name of Instrument	Sl. No. / Id no.	Certificate No.	Lab Certificate No.	Calibrated on	Due on
Temp. & Humidity Probe with Indicator	(Sl. No.: No.- 2022C01005/20571921)	ATL/T/090323/001	CC-2590	06-03-2023	06-03-2024
Digital Temperature Indicator with Sensor(RTD)	(Indicator Sl. No.- 18K588073) (Id No.- TMSCC/RTD/01)	TSC/23-24/13010-4	CC - 2231	09-11-2023	08-11-2024

### CALIBRATION RESULTS

Sl. No.	Ref. Bath set in °C	Observed Reading at DUC in °C (Avg. of five readings)	Observed Reading at Ref. Std. in °C (Avg. of five readings)	Error in °C	Uncertainty in ± °C	Acceptance Criteria
1	10.0	10.0	9.986	-0.014	0.18	Pass
2	20.0	20.0	19.971	-0.029	0.18	Pass
3	30.0	30.0	29.957	-0.043	0.31	Pass
4	50.0	50.0	49.928	-0.072	0.31	Pass
5	70.0	70.0	69.899	-0.101	0.31	Pass

Maximum Permissible Error : ± 0.3% of rdg

Sl. No.	Ref. Source set in %	Observed Reading at DUC in % (Avg. of five readings)	Observed Reading at Ref. Std. in % (Avg. of five readings)	Error in %	Measurement Uncertainty in ± %	Acceptance Criteria
1	20	20	20.3	0.3	1.3	Pass
2	25	25	25.3	0.3	1.3	Pass
3	50	50	50.7	0.7	1.3	Pass
4	90	90	91.2	1.2	1.3	Pass

Maximum Permissible Error : ± 3% of rdg

Measurement Uncertainty at 95% Confidence Level where Coverage Factor k = 2

REMARKS : The DUC has been calibrated over its range. The readings observed are tabulated above. The reference standard is traceable to National standard.

DUC: Device Under Calibration.

Physical Status of the DUC : OK

Calibrated By:  
  
R. Ghosh  
(Calibration Engineer)

Checked By:  
  
C. Ghosh  
(Technical Director)



Approved By:  
  
P. K. Modak  
Sr. Calibration Engineer

END OF CERTIFICATE