

CIN No: U85195DL2006PTC152508
GSTIN: 07AAECP0016C1Z5

DRUG L. No. : 20B/DL-ROH-128185
DRUG L. No. : 21B/DL-ROH-128186

PSJ BIOCHEM DIAGNOSTIC PVT. LTD.

(AN ISO 13485:2012 Certified Company)


Your Diagnostic Partner

Regd.Office: Plot No. 1109, First Floor, Bagga Link Road, Rithala, Delhi-110085(INDIA)
Phone: 011-27054224, 9810098224 E-Mail: psjbiochem@rediffmail.com

CALIBRATION CERTIFICATE			
Customer Name	<i>Welldynamic Diagnostic Pvt ltd</i>		
Department	Department of Laboratory Medicine - Hematology		
Address	139, Pocket 4 Rohini Sector-22, New delhi-110085		
		Date of Calibration	23.05.2024
Instrument	<i>Mindray BC-2800</i>	Next Calibration Date	22.05.2025
Serial No	RP-0B105726	Calibrator Lot Number	1X0301
		Calibrator Expiry date	05.09.2024

	<u>Parameter</u>	<u>Target</u>	<u>Range</u>	<u>New Factor</u>	<u>Old Factor</u>
Calibration Values	WBC x 10 ⁹ /L	7.3	6.3 – 8.3	102.5	90.0
	RBC x10 ¹² /L	4.52	4.22 - 4.82	91.8	100.00
	HGB g/L	13.8	132- 144	103.3	95.00
	MCV fl	87.0	81.0 - 93.0	100.6	90.00
	PLT x 10 ⁹ /L	213	168 - 258	78.5	90.00

This is to certify that the Mindray Bc-2800 Hematology Analyzer Sr.No: RP-0B105726 system, installed at Welldynamic Diagnostic Pvt Ltd ROHINI , DELHI has been calibrated using the standard reference material D-Check-D Calibrator. Traceability of SC CAL PLUS Calibrator 1X0301 material from the manufacturer is attached along with this certificate.



Date:23.05.2024

Pappu Kumar Service Engineer
PSJ Biochem Diagnostic Pvt Ltd