



Metropolis Healthcare Ltd., EQAS

KRSNAA DIAGNOSTICS,LTD,MOGA (1054)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 230310 and Sample No 01

Report Date : 19/12/2023

Analyte	Instrument	Result Value	Standard Unit	Z-Score
X Bilirubin Total	Diasys 200 Pro	0.89	mg/dl	6.51
Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : Diasys 200 Pro

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
# Urea	mg/dl	22.80	20.87		0.09
@ Creatinine	mg/dl	0.81	0.83	0.27	-0.58
X Bilirubin Total	mg/dl	0.89	0.30	6.51	3.71
@ Total Protein	g/dL	5.04	5.76	-1.49	-1.51
@ Albumin	g/dL	3.73	3.59	1.17	1.12
@ Uric Acid	mg/dl	3.77	3.74	-0.46	-0.34
@ Calcium	mg/dl	7.17	7.31	0.04	0.20
@ Cholesterol	mg/dl	130.00	130.32	-0.08	0.02
@ Triglycerides	mg/dl	121.00	118.41	-0.18	0.30
@ HDL	mg/dl	38.00	33.30	1.63	1.67
@ LDL	mg/dl	59.26	73.11	-1.75	-1.29
@ ALT	U/L	8.50	13.62	-1.83	-0.92
@ AST	U/L	27.00	26.43	0.27	0.27
@ ALP	U/L	40.34	42.72	-0.59	-0.38

Instrument : ACCULYTE-3P

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Chloride	mmol/L	80.94	81.59	0.17	-0.20
@ Sodium	mmol/L	100.55	106.54	-1.15	-0.68
@ Potassium	mmol/L	3.39	3.48	-0.39	-0.23

Legend @ : Acceptable

! : 2.0 < z score < 3.0 - Warning


X : z score ≥ 3 - Unacceptable

: Not Evaluated

Ⓞ : Delayed Result Entry

* : Not considered for evaluation.

Total Parameters	17
Not Evaluated Parameters	1
Evaluated Parameters	16
Outlier Parameters (X)	1
EQAS Score Biochemistry	93.75 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS

Unit No. 409-416.

Commercial Building - 1A

Kohinoor Mall, Kirol Road, Kurla (W),

Mumbai - 400070