



Metropolis Healthcare Ltd., EQAS

KRSNAA DIAGNOSTICS,LTD,MOGA (1054)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 230310 and Sample No 02

Report Date : 19/12/2023

Analyte	Instrument	Result Value	Standard Unit	Z-Score
! Creatinine	Diasys 200 Pro	0.32	mg/dl	-2.61
Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : Diasys 200 Pro

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Glucose	mg/dl	81.60	78.47	1.10	0.89
@ Urea	mg/dl	20.60	20.58	0.17	0.09
! Creatinine	mg/dl	0.32	0.58	-2.61	-0.58
@ Bilirubin Total	mg/dl	2.94	2.33	0.91	3.71
@ Bilirubin Direct	mg/dl	0.35	0.30	0.53	0.53
@ Total Protein	g/dL	5.01	5.79	-1.53	-1.51
@ Albumin	g/dL	3.88	3.68	1.32	1.12
@ Uric Acid	mg/dl	3.81	3.80	-0.49	-0.34
@ Calcium	mg/dl	7.56	7.48	0.59	0.20
@ Cholesterol	mg/dl	136.00	136.84	0.11	0.02
@ Triglycerides	mg/dl	118.00	112.92	0.77	0.30
@ HDL	mg/dl	25.00	16.83	1.71	1.67
@ LDL	mg/dl	74.47	73.56	-0.83	-1.29
@ ALP	U/L	42.82	42.19	-0.16	-0.38

Instrument : ACCULYTE-3P

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Chloride	mmol/L	83.42	84.96	0.15	-0.20
@ Sodium	mmol/L	102.65	109.02	-1.10	-0.68
@ Potassium	mmol/L	3.44	3.56	-0.49	-0.23

Legend @ : Acceptable

! : 2.0 < z score < 3.0 - Warning


X : z score ≥ 3 - Unacceptable

: Not Evaluated

Ⓢ : Delayed Result Entry

* : Not considered for evaluation.

Total Parameters	17
Not Evaluated Parameters	0
Evaluated Parameters	17
Outlier Parameters (X)	0
EQAS Score Biochemistry	100.00 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS

Unit No. 409-416.

Commercial Building - 1A

Kohinoor Mall, Kirol Road, Kurla (W),

Mumbai - 400070