

Lab : 1054 / KRSNAA DIAGNOSTICS,LTD,MOGA

Cycle + Sample No : 23031001

Analyte Scheme :

Biochemistry

PT Result Received Date	: 08/12/2023	Name Of The Test	: Bilirubin Total
Result	: 0.89	Z Score	: 6.51
		Upload Date	: 15/01/2024

Previous trends/Unacceptable results :

last 6-12 month : No Trends in last 6 months
Last 3 months : No Trends in last 3 months
Data Entry Check : Yes
Investigation : No Any finding
Conclusion : It is a Random error
Was patient data affected? : No
Classification Of Problem : Technical
Immediate Action : Rerun the sample
Corrective Action : Rerun the EQAS sample
Preventive Action : On Daily basis Review & Monitor the IQC
Remark : After Rerun Rerun Value is 0.38 mg/dl, And Z score is 1.61 after rerun the sample Z score is within Range

PT Result Received Date	: 08/12/2023	Name Of The Test	: Bilirubin Total
Result	: 0.89	Z Score	: 6.51
		Upload Date	: 04/01/2024

Previous trends/Unacceptable results :

last 6-12 month : no trend in last 6 months
Last 3 months : no trend in last 3 months
Data Entry Check : yes correct data entry
Investigation : rerun test total bilirubin
Conclusion :
Was patient data affected? : no
Classification Of Problem : Technical
Immediate Action : check calibration and qc
Corrective Action : rerun and check z score
Preventive Action : review date before send
Remark : rerun the sample and z score are within acceptable range