

Lab : 1054 / KRSNAA DIAGNOSTICS,LTD,MOGA

Cycle + Sample No : 23031001

Analyte Scheme : Clinical Path - Urine Routine Chemistry

PT Result Received Date : 08/12/2023 Name Of The Test : Ketone (urine)
Result : 1+ Z Score : -- Upload Date : 15/01/2024

Previous trends/Unacceptable results :

last 6-12 month : No Trends
Last 3 months : No Trends
Data Entry Check : Yes All ok
Investigation : Random Error
Conclusion : Training arrange to taff for Handling of PT Sample
Was patient data affected? : No
Classification Of Problem : Technical
Immediate Action : Split Testing Done
Corrective Action : Split Testing Done
Preventive Action : Trainings arrange on Handling the PT Sample
Remark : Split testing Done. Split test is satisfactory

PT Result Received Date : 08/12/2023 Name Of The Test : Nitrite (urine)
Result : Absent Z Score : -- Upload Date : 15/01/2024

Previous trends/Unacceptable results :

last 6-12 month : No Trends
Last 3 months : No Trends
Data Entry Check : Yes All ok
Investigation : Random Error
Conclusion : Training arrange to taff for Handling of PT Sample
Was patient data affected? : No
Classification Of Problem : Technical
Immediate Action : Split Testing Done
Corrective Action : Split Testing Done
Preventive Action : Trainings arrange on Handling the PT Sample
Remark : Split testing Done. Split test is satisfactory