



Metropolis Healthcare Ltd., EQAS

KRSNAA DIAGNOSTICS,LTD,MOGA (1054)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 230310 and Sample No 03

Report Date : 19/12/2023

Analyte	Instrument	Result Value	Standard Unit	Z-Score
* Total Leucocytes (WBC) count	Horiba Yumizen H550	1.85	/c.mm	*
Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : Horiba Yumizen H550

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Erythrocyte (RBC) Count	mill/cu.mm	1.04	1.02	0.01	-0.02
* Total Leucocytes (WBC) count	/c.mm	1.85	1835.53	-6.40 *	0.00
@ Haemoglobin	g/dL	2.70	2.68	0.00	0.05
@ MCV (Mean Corpuscular Volume)	fL	123.50	125.66	-0.21	-0.25
@ MCH (Mean Corpuscular Hb)	pg	26.10	26.65	0.03	0.08
@ MCHC (Mean Corpuscular Hb Concentration)	g/dL	21.10	20.97	0.25	0.29
@ Platelet Count	1000/ µL	75.00	87.19	-0.46	-0.19
@ Hematocrit	%	12.80	12.66	-0.21	-0.22
@ RDW CV%	%	17.90	24.07	-1.51	-1.16

Legend @ : Acceptable

! : 2.0 < z score < 3.0 - Warning


X : z score ≥3 - Unacceptable

: Not Evaluated

⌚ : Delayed Result Entry

* : Not considered for evaluation.

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	1
EQAS Score Haematology - CBC	88.89 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS

Unit No. 409-416.

Commercial Building - 1A

Kohinoor Mall, Kirol Road, Kurla (W),

Mumbai - 400070